

Service Repair Sheet 69453
Contact Name Paul Denyer
Company/ Hospital Name Crawley Hospital
Department EME Department
Position EME Services Technician
Direct Phone 01293600300
General Phone 01737768511
Opera Account 00001300
Email paul.denyer@nhs.net
Order Number 106349260
Date Received 18/Jun/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38884	2510000	H0002232	Microstim DB3	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	2540010		
	Qty: 1	2520000		
	Qty: 1	2510000		

The Microstim DB3 has been repaired and the patient lead has been replaced.
