

Service Repair Sheet 69474

Contact Name	Paul James
Company/ Hospital Name	Coventry University Hospital
Department	EBME Department
Position	Electronics Engineer
Direct Phone	02476968386
General Phone	02476964000
Opera Account	00001288
Email	paul.james@uhcw.nhs.uk
Order Number	999064032
Date Received	20/Jul/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38951	1410000	GE71743941	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
