

Service Repair Sheet SRS69468

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Stinder Bassi

Queen Elizabeth University Hospi

Medical Physics Department

Specialist Clinical Technologist

01414523030

01412011100

00001853

stinder.bassi2@nhs.scot

13/Jul/2026

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 08/Jul/2026 Emily Hanson

08/Jul/2026 Emily Hanson

VM-2500-S C02/SpO2 monitor S/N 60940148 that has an issue with the real time clock. The device will not store the correct time when entered.

13/Jul/2026 Robert Connor

Received 1 x VM-2500-S s/n 60940148, with silicone cover.

Ready For quote

CCN 24.7.26

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38916	Capnograph	4410520	60940148	N

4440000 x 1 @ 0.5hr = £49.35

S/N, SES, SRN

UPS x 1 @ £12.