

Service Repair Sheet 69473

Contact Name	David Edmondson
Company/ Hospital Name	Royal Preston Hospital
Department	Medical Engineering
Position	Technician
Direct Phone	01772524598
General Phone	01772716565
Opera Account	00004260
Email	david.edmondson@lthtr.nhs.uk
Order Number	TP38873
Date Received	20/Jul/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38952	0310030	VMEF01	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 0380000 Annual service - Tom Thumb

Notes:

Tom Thumb TT480 VMEF01 has been stripped and cleaned then reassembled with new O rings.
The two valves have been checked and reset before QA.

Engineer Notes:

22/Jul/2026 Philip CrossleyRoutine service.

Repair	Ref	S/N	Equipment Type	
SRN38954	0310030	VMEF09	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 0380000 Annual service - Tom Thumb

Notes:

Tom Thumb TT480 VMEF09 has been stripped and cleaned then reassembled with new O rings.
The two valves have been checked and reset before QA.

Engineer Notes:

22/Jul/2026 Philip CrossleyRoutine service.

Repair	Ref	S/N	Equipment Type	
SRN38955	0310200	B1551977	Blender	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 0380020 Blender Service - Return to base.
Qty: 1 0332000 Blender - Annual service kit.

Notes:

Blender B1551977 has had an annual service with new gas filters and the calibration has been checked.

Engineer Notes:

22/Jul/2026 Philip CrossleyRoutine annual service and calibration test.

Service Repair Sheet 69473
Contact Name David Edmondson
Company/ Hospital Name Royal Preston Hospital
Department Medical Engineering
Position Technician
Direct Phone 01772524598
General Phone 01772716565
Opera Account 00004260
Email david.edmondson@lthtr.nhs.uk
Order Number TP38873
Date Received 20/Jul/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38956	0310200	B1551980	Blender	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380020	Blender Service - Return to base.
Qty: 1	0332000	Blender - Annual service kit.

Notes:

Blender B1551980 has had an annual service with new gas filters and the calibration has been checked.

Engineer Notes:

22/Jul/2026 Philip CrossleyRoutine annual service and calibration test.
