

Service Repair Sheet 69477

Contact Name	Dennis Vessey
Company/ Hospital Name	Dennis Vessey and Co
Direct Phone	01652 688 904
General Phone	01652 688 904
Opera Account	CID21569
Email	dennis.vessey@btconnect.com
Date Received	22/Jul/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN38958	0110215	261899	Oxygen Sensor Vandagraph	Replaced Time :0 Hour(s)

Sensor has very low output of less than 1mV. Failed or depleted within warranty. To be replaced free of charge.
