

Service Repair Sheet 69479

Contact Name	Robert Barr
Company/ Hospital Name	Cardiotec
Direct Phone	07980656316
General Phone	07980656316/02895908890
Opera Account	CID17049
Email	robert@cardiotec.co.uk
Order Number	PO-2613
Date Received	27/Jul/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38966	1410000	PR03372A17	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
