

Service Repair Sheet SRS69484

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Joseph Sojan
Kings College Hospital (use this o
Medical Engineering and Physics
Medical Engineering Technician
020 3299 3538
02032997878
00002990
joseph.sojan@nhs.net

04/Aug/2026
Robert Connor
Viamed
Quote

VIAMEDclean

Goods In Only
Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: RC
Date: 4/8/26

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 29/Jul/2026 Emily Hanson

29/Jul/2026 Emily Hanson

I have a faulty MaxO2 ME oxygen monitor. The sound alarm of the unit is not working (The device is not on mute)

SN FJ31899020

04/Aug/2026 Robert Connor

Received 1 x MAXO2ME s/n FJ31899020, with 4 x AA battery.

Ready For quote

CG Neen 10.8.26

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38979	Oxygen Monitor	0111263	FJ31899020	<u>N</u>

0140120 @ £99 per hr 0.5 = £49.50

S/N, SRS, SRN

0131711 x 1 @ £23.50

SRS, SRN

UPS x 1 £12