

69453

Service Repair Sheet SRS69496

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Paul Denyer

Crawley Hospital

EME Department

EME Services Technician

01293600300

01737768511

00001300

paul.denyer@nhs.net

11/Aug/2026

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only
Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒
Signed: _____
Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 06/Aug/2026 Aqib Majeed

06/Aug/2026 Aqib Majeed

Customer has a DB3 Nerve stimulator SN: H0002232, no longer works, the battery connections have voltage on them, output light doesn't flash regardless which switch is pressed or what strength control it is.

11/Aug/2026 Robert Connor

Received 1 x microstim DB3 s/n H0002232.

Changed to warranty.

Ready For quote

Repair Complete Signed

C Connor 12.8.26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38996	Microstim DB3	2510000	H0002232	Yes (r)

2540000

~~2510000~~ x 1 FOC

S/N, SRS, SRN

UPS x 1 FOC.