

Your ref:	SRS69494
Our ref:	
Tel:	01384-456111 extn. 1261
E-Mail:	dgft.medicalengineering@nhs.net
Date	05/08/2026 13:06

Dear Sirs

RE: REPAIR OF EQUIPMENT - QUOTATION

Please provide a quotation for the repair of the equipment below. Please email to the above email address as soon as possible.

When the quotation has been received and accepted I will then forward an order number to you. Please note the invoice address and the return address on the Purchase Order.


Our Job No: 276491 (THIS IS NOT AN ORDER NUMBER)
Company VIAMED LTD
Equipment / Model APGAR TIMER - 0310100
Serial No: 00106
Asset No: EBME29408
Fault TURNS OFF AND ON BY ITSELF. ALSO WILL NOT RESET
Enc APGAR TIMER

HAS ALL PATIENT DATA BEEN REMOVED? Yes

Please return the equipment to the address shown below.

Thank you for your help in this matter. Please do not hesitate to contact me if there are any queries.

Yours faithfully



Matt Hill (Technician)

Medical Engineering Technician

Medical Engineering Department
Delivery Point 1
Russells Hall Hospital
Dudley
West Midlands
DY1 2HQ