

Service Repair Sheet 69489

Contact Name	David Edmondson
Company/ Hospital Name	Royal Preston Hospital
Department	Medical Engineering
Position	Technician
Direct Phone	01772524598
General Phone	01772716565
Opera Account	00004260
Email	david.edmondson@lthtr.nhs.uk
Order Number	TP38873
Date Received	10/Aug/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38991	0310030	VMD123	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
--------	---------

Tom Thumb TT480 VMD123 has been stripped and cleaned then reassembled with new O rings.

The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN38992	0310030	VMD124	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
--------	---------

Tom Thumb TT480 VMD124 has been stripped and cleaned then reassembled with new O rings.

The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN38993	0310200	B1551736	Blender	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380020
Qty: 1	0332000

Blender B1551736 has had an annual service with new gas filters and the calibration has been checked.

Repair	Ref	S/N	Equipment Type	
SRN38994	0310200	B1551985	Blender	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380020
Qty: 1	0332000

Blender B1551985 has had an annual service with new gas filters and the calibration has been checked.
