

Service Repair Sheet 69470  
Contact Name CM Hor  
Company/ Hospital Name Emstec Services  
Direct Phone 6046581791  
General Phone 6046581791  
Opera Account 00013104  
Email cmhor@emstecmedical.com  
Date Received 22/Jul/2026  
Booked in By Robert Connor

---

| Repair         | Ref     | S/N        | Equipment Type   |                                        |
|----------------|---------|------------|------------------|----------------------------------------|
| SRN38959       | 1410000 | PR02041A11 | Foetal Simulator | Complete - Repaired<br>Time :0 Hour(s) |
| Parts Replaced |         |            |                  |                                        |
|                |         | Qty: 1     | 1480000          |                                        |
|                |         | Qty: 1     | 1430309          |                                        |

V1000 has had transducer interface cushion replaced and calibration has been checked.

---