

Service Repair Sheet 69501

Contact Name	Katy Romanowski
Company/ Hospital Name	Tenacore LLC
Position	Director of Operations
Direct Phone	8002972241
General Phone	8002972241
Opera Account	00007996
Email	Katy.Romanowski@prescottsmed.com
Order Number	SRS69501/ PO10652
Date Received	17/Aug/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38997	1410000	PR02976A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
