

File:/meter.sam



Part Number	9901-15
Description	Pressure Gauge 0 - 50 cm H2O
Picture File	meter.jpg
TechScan	-
Text File	meter.sam
Title	Pressure Gauge 0 - 50 cm H2O
Material	-N/A
Surface Finish	-N/A
Process Finish	-N/A
Is a Part Of	Unit
Supplier	WIKA Instruments
Supplier P/N	9758709
Lead Time	4-6 Weeks

WIKA INSTRUMENTS LIMITED



STATION APPROACH, COULSDON, SURREY CR5 2UD
Telephone: 0181 - 668 0924 Fax: 0181 - 660 5956

Customers Ref.: DL/EG/8889
WIKA Ref.: 203633
Customer: VIAMED LTD

Certificate of Calibration

WIKA Model No.: 611.10.63

Size and Style: 63mm back

Calibrated Range: -5/+60 cm wg

Serial Number: 203633

Tag Number: N/A

Accuracy of Determination: $\pm 1.6\%$ FSD

Test Equipment: RUSKA AIR BENCH

DKD Number: 4733

Accuracy of Test Equipment: $\pm 0.009\%$ FS 800mbar

Testing Temp.: 20 ° C

Method of Test: Direct Comparison

Test Medium: Air

True Units cm wg	Pressure	Gauge Readings			
		Rising Pressure		Falling Pressure	
		Reading	Error [%]	Reading	Error [%]
-5,000		-5,650	-1,00	-5,740	-1,14
0,000		0,000	0,00	0,000	0,00
10,000		10,120	0,18	10,060	0,09
30,000		30,320	0,49	30,070	0,11
60,000		60,850	1,31	60,810	1,25

All equipment utilised in the preparation of this Certificate is directly traceable to PTB primary standards under licence DKD-K-3701

Tested by:

D.P. Winkler

Signed

P. Weller

Date:

07.06.1997

Quality Assurance

WIKA INSTRUMENTS LIMITED



STATION APPROACH, COULSDON, SURREY CR5 2UD
Telephone: 0181 - 668 0924 Fax: 0181 - 660 5956

Customers Ref.: DL/EG/8889
WIKA Ref.: 203633
Customer: VIAMED LTD

Certificate of Calibration

WIKA Model No.: 611.10.63
Size and Style: 63mm back
Calibrated Range: -5/+60 cm wg
Serial Number: 203633
Tag Number: N/A
Accuracy of Determination: $\pm 1.6\%$ FSD
Test Equipment: RUSKA AIR BENCH DKD Number: 4733
Accuracy of Test Equipment: $\pm 0.009\%$ FS 800mbar Testing Temp.: 20 °C
Method of Test: Direct Comparison Test Medium: Air

True Units cm wg	Pressure	Gauge Readings			
		Rising Pressure		Falling Pressure	
		Reading	Error [%]	Reading	Error [%]
-5,000		-5,650	-1,00	-5,740	-1,14
0,000		0,000	0,00	0,000	0,00
10,000		10,120	0,18	10,060	0,09
30,000		30,320	0,49	30,070	0,11
60,000		60,850	1,31	60,810	1,25

All equipment utilised in the preparation of this Certificate is directly traceable to PTB primary standards under licence DKD-K-3701

Tested by :

D.P. Wicks

Signed

P. Weller

Date:

07.06.1997

Quality Assurance