



VIA MED

# Bluepoint Pulse Oximetry Sensors 4

Reusable and disposable pulse oximetry sensors for continuous and spot-check monitoring of functional arterial oxygen saturation and pulse rate.












Class IIb

Assesment Route Annex II  
NBOG MD1302

Carried out by Derek Lamb  
27 / 10 / 17



# VIA MED

-  Stock References Review
-  Supplier Review
-  Sales Review
-  Countries Review
-  Returns Review
-  Design Changes Review
-  User Instructions Review
-  Labels Review
-  Document Updates Review
-  Internal Issues Review
-  Clinical / FDA Incidents online search

# Stock References Review

Stock Reference	Description
0014590	SpO2 Silicone Finger Sensor - Adult.
0014591	SpO2 Silicone Finger Sensor - Paediatric
0014592	SpO2 Wrap Sensor- Paediatric
0014595	SpO2 Silicone Finger Sensor - Adult.
0014596	SpO2 Silicone Finger Sensor - Paediatric
0014601	SpO2 Finger Sensor - Adult.
0014602	SpO2 Finger Sensor - Adult.
0014603	SpO2 Finger Sensor - Adult.
0014604	SpO2 Finger Sensor - Adult.
0014605	SpO2 Finger Sensor - Adult.
0014606	SpO2 Finger Sensor - Adult.
0014607	SpO2 Finger Sensor - Adult.
0014608	SpO2 Finger Sensor - Adult.
0014609	SpO2 Finger Sensor - Adult.
0014610	SpO2 Finger Sensor - Adult.
0014611	SpO2 Finger Sensor - Adult.
0014612	SpO2 Finger Sensor - Adult.
0014613	SpO2 Finger Sensor - Adult.
0014614	SpO2 Finger Sensor - Adult.
0014615	SpO2 Finger Sensor - Adult.
0014617	SpO2 Finger Sensor - Adult.
0014618	SpO2 Finger Sensor - Adult.
0014619	SpO2 Finger Sensor - Adult.
0014620	SpO2 Finger Sensor - Adult.
0014621	SpO2 Finger Sensor - Adult.
0014623	SpO2 Finger Sensor - Adult.
0014625	SpO2 Finger Sensor - Adult.
0014649	SpO2 Finger Sensor - Adult.
0014650	SpO2 Finger Sensor - Adult.
0014701	SpO2 Silicone Finger Sensor - Adult.
0014702	SpO2 Silicone Finger Sensor - Adult.
0014703	SpO2 Silicone Finger Sensor - Adult.
0014704	SpO2 Silicone Finger Sensor - Adult.
0014705	SpO2 Silicone Finger Sensor - Adult.
0014706	SpO2 Silicone Finger Sensor - Adult.
0014707	SpO2 Silicone Finger Sensor - Adult.
0014708	SpO2 Silicone Finger Sensor - Adult.
0014709	SpO2 Silicone Finger Sensor - Adult.
0014710	SpO2 Silicone Finger Sensor - Adult.
0014711	SpO2 Silicone Finger Sensor - Adult.
0014712	SpO2 Silicone Finger Sensor - Adult.
0014713	SpO2 Silicone Finger Sensor - Adult.

Stock Reference	Description
0014714	SpO2 Silicone Finger Sensor - Adult.
0014715	SpO2 Silicone Finger Sensor - Adult.
0014717	SpO2 Silicone Finger Sensor - Adult.
0014718	SpO2 Silicone Finger Sensor - Adult.
0014719	SpO2 Silicone Finger Sensor - Adult.
0014720	SpO2 Silicone Finger Sensor - Adult.
0014721	SpO2 Silicone Finger Sensor - Adult.
0014722	SpO2 Silicone Finger Sensor - Adult.
0014723	SpO2 Silicone Finger Sensor - Adult.
0014724	SpO2 Silicone Finger Sensor - Adult.
0014725	SpO2 Silicone Finger Sensor - Adult.
0014740	SpO2 Silicone Finger Sensor - Adult.
0014750	SpO2 Silicone Finger Sensor - Adult.
0014751	SpO2 Silicone Finger Sensor - Paediatric
0014755	SpO2 Silicone Finger Sensor - Adult.
0014756	SpO2 Silicone Finger Sensor - Paediatric
0014761	SpO2 Silicone Finger Sensor - Paediatric
0014762	SpO2 Silicone Finger Sensor - Paediatric
0014763	SpO2 Silicone Finger Sensor - Paediatric
0014764	SpO2 Silicone Finger Sensor - Paediatric
0014765	SpO2 Silicone Finger Sensor - Paediatric
0014766	SpO2 Silicone Finger Sensor - Paediatric
0014767	SpO2 Silicone Finger Sensor - Paediatric
0014768	SpO2 Silicone Finger Sensor - Paediatric
0014769	SpO2 Silicone Finger Sensor - Paediatric
0014770	SpO2 Silicone Finger Sensor - Paediatric
0014771	SpO2 Silicone Finger Sensor - Paediatric
0014772	SpO2 Silicone Finger Sensor - Paediatric
0014773	SpO2 Silicone Finger Sensor - Paediatric
0014774	SpO2 Silicone Finger Sensor - Paediatric
0014775	SpO2 Silicone Finger Sensor - Paediatric
0014777	SpO2 Silicone Finger Sensor - Paediatric
0014778	SpO2 Silicone Finger Sensor - Paediatric
0014779	SpO2 Silicone Finger Sensor - Paediatric
0014780	SpO2 Silicone Finger Sensor - Paediatric
0014781	SpO2 Silicone Finger Sensor - Paediatric
0014782	SpO2 Silicone Finger Sensor - Paediatric
0014785	SpO2 Silicone Finger Sensor - Paediatric
0014790	SpO2 Silicone Finger Sensor - Paediatric
0014801	Pulse Oximetry Ear Sensor
0014802	Pulse Oximetry Ear Sensor
0014803	Pulse Oximetry Ear Sensor
0014804	Pulse Oximetry Ear Sensor
0014805	Pulse Oximetry Ear Sensor
0014806	Pulse Oximetry Ear Sensor

Stock Reference	Description
0014807	Pulse Oximetry Ear Sensor
0014808	Pulse Oximetry Ear Sensor
0014816	Pulse Oximetry Ear Sensor
0014820	Pulse Oximetry Wrap Sensor
0014821	Pulse Oximetry Wrap Sensor
0014822	Pulse Oximetry Wrap Sensor
0014823	Pulse Oximetry Wrap Sensor
0014824	Pulse Oximetry Wrap Sensor
0014825	Pulse Oximetry Ear Sensor
0014826	Pulse Oximetry Wrap Sensor
0014827	Pulse Oximetry Wrap Sensor
0014828	Pulse Oximetry Wrap Sensor
0014829	Pulse Oximetry Wrap Sensor
0014830	Pulse Oximetry Wrap Sensor
0014831	Pulse Oximetry Wrap Sensor
0014832	Pulse Oximetry Wrap Sensor
0014833	Pulse Oximetry Wrap Sensor
0014834	Pulse Oximetry Wrap Sensor
0014835	Pulse Oximetry Wrap Sensor
0014836	Pulse Oximetry Wrap Sensor
0014837	Pulse Oximetry Wrap Sensor
0014838	Pulse Oximetry Wrap Sensor
0014840	Pulse Oximetry Wrap Sensor
0014841	Pulse Oximetry Wrap Sensor
0014842	Pulse Oximetry Wrap Sensor
0014843	Pulse Oximetry Wrap Sensor
0014844	Pulse Oximetry Wrap Sensor
0014845	Pulse Oximetry Multi-Site Y Sensor
0014848	Pulse Oximetry Wrap Sensor
0014895	Pulse Oximetry Extension Cable 1.2 m
0014896	Pulse Oximetry Extension Cable 2.4 m
0014901	Disposable SpO2 Sensors - Adult.
0014902	Disposable SpO2 Sensors - Paediatric.
0014903	Disposable SpO2 Sensors - Infant.
0014904	Disposable SpO2 Sensors - Neonatal.
0014905	Disposable SpO2 Sensors - Adult.
0014906	Disposable SpO2 Sensors - Paediatric.
0014909	Disposable SpO2 Sensors - Adult.
0014910	Disposable SpO2 Sensors - Paediatric.
0014911	Disposable SpO2 Sensors - Infant.
0014912	Disposable SpO2 Sensors - Neonatal.
0014913	Disposable SpO2 Sensors - Adult.
0014914	Disposable SpO2 Sensors - Paediatric.
0014915	Disposable SpO2 Sensors - Infant.
0014916	Disposable SpO2 Sensors - Neonatal.

Stock Reference	Description
0014917	Disposable SpO2 Sensors - Adult.
0014918	Disposable SpO2 Sensors - Paediatric.
0014921	Disposable SpO2 Sensors - Adult.
0014922	Disposable SpO2 Sensors - Paediatric.
0014923	Disposable SpO2 Sensors - Infant.
0014924	Disposable SpO2 Sensors - Neonatal.
0014925	Disposable SpO2 Sensors - Adult.
0014926	Disposable SpO2 Sensors - Paediatric.
0014927	Disposable SpO2 Sensors - Infant.
0014928	Disposable SpO2 Sensors - Neonatal.
0014929	Disposable SpO2 Sensors - Adult.
0014930	Disposable SpO2 Sensors - Paediatric.
0014933	Disposable SpO2 Sensors - Adult.
0014934	Disposable SpO2 Sensors - Paediatric.
0014935	Disposable SpO2 Sensors - Infant.
0014936	Disposable SpO2 Sensors - Neonatal.
0014937	Disposable SpO2 Sensors - Adult.
0014938	Disposable SpO2 Sensors - Paediatric.
0014939	Disposable SpO2 Sensors - Infant.
0014940	Disposable SpO2 Sensors - Neonatal.
0014941	Disposable SpO2 Sensors - Adult.
0014942	Disposable SpO2 Sensors - Paediatric.
0014945	Disposable SpO2 Sensors - Adult.
0014946	Disposable SpO2 Sensors - Paediatric.
0014947	Disposable SpO2 Sensors - Infant.
0014948	Disposable SpO2 Sensors - Neonatal.
0014949	Disposable SpO2 Sensors - Adult.
0014950	Disposable SpO2 Sensors - Paediatric.
0014951	Disposable SpO2 Sensors - Infant.
0014952	Disposable SpO2 Sensors - Neonatal.
0014953	Disposable SpO2 Sensors - Adult.
0014954	Disposable SpO2 Sensors - Paediatric.
0014957	Disposable SpO2 Sensors - Adult.
0014958	Disposable SpO2 Sensors - Paediatric.
0014959	Disposable SpO2 Sensors - Infant.
0014960	Disposable SpO2 Sensors - Neonatal.
0014961	Disposable SpO2 Sensors - Adult.
0014962	Disposable SpO2 Sensors - Paediatric.
0014963	Disposable SpO2 Sensors - Infant.
0014964	Disposable SpO2 Sensors - Neonatal.
0014965	Disposable SpO2 Sensors - Adult.
0014966	Disposable SpO2 Sensors - Paediatric.
0014969	Disposable SpO2 Sensors - Adult.
0014970	Disposable SpO2 Sensors - Paediatric.
0014971	Disposable SpO2 Sensors - Infant.

Stock Reference	Description
0014972	Disposable SpO2 Sensors - Neonatal.
0014973	Disposable SpO2 Sensors - Adult.
0014974	Disposable SpO2 Sensors - Paediatric.
0014975	Disposable SpO2 Sensors - Infant.
0014976	Disposable SpO2 Sensors - Neonatal.
0014977	Disposable SpO2 Sensors - Adult.
0014978	Disposable SpO2 Sensors - Paediatric.
0014981	Disposable SpO2 Sensors - Adult.
0014982	Disposable SpO2 Sensors - Paediatric.
0014983	Disposable SpO2 Sensors - Infant.
0014984	Disposable SpO2 Sensors - Neonatal.
0014985	Disposable SpO2 Sensors - Adult.
0014986	Disposable SpO2 Sensors - Paediatric.
0014987	Disposable SpO2 Sensors - Infant.
0014988	Disposable SpO2 Sensors - Neonatal.
0014989	Disposable SpO2 Sensors - Adult.
0014990	Disposable SpO2 Sensors - Paediatric.
0014993	Disposable SpO2 Sensors - Adult.
0014994	Disposable SpO2 Sensors - Paediatric.
0014995	Disposable SpO2 Sensors - Infant.
0014996	Disposable SpO2 Sensors - Neonatal.
0019001	Nellcor DEC-4 compatible SpO2 extension
0019002	Nellcor DEC-8 compatible SpO2 extension

Comments on Stock references review:

Stock reference upto date for now,  
 unsure if references will changed when we finally get Bluepoints name on the labels

# Supplier Review

Stock Ref.	Description	Supplier A/C	Supplier P/N	Supplier Name	Rating
0014740	SpO2 Silicone Finger Sen	00009299		Bluepoint Medical	A
0014790	SpO2 Silicone Finger Sen	00009299		Bluepoint Medical	A
0014901	Disposable SpO2 Sensors	00009299	1020131122	Bluepoint Medical	A
0014902	Disposable SpO2 Sensors	00009299	1020131125	Bluepoint Medical	A
0014903	Disposable SpO2 Sensors	00009299	1020131128	Bluepoint Medical	A
0014904	Disposable SpO2 Sensors	00009299	1020131130	Bluepoint Medical	A
0014905	Disposable SpO2 Sensors	00009299	1020131123	Bluepoint Medical	A
0014906	Disposable SpO2 Sensors	00009299	1020131126	Bluepoint Medical	A
0014909	Disposable SpO2 Sensors	00009299	1020131124	Bluepoint Medical	A
0014910	Disposable SpO2 Sensors	00009299	1020131127	Bluepoint Medical	A
0014911	Disposable SpO2 Sensors	00009299	1020131129	Bluepoint Medical	A
0014912	Disposable SpO2 Sensors	00009299	1020131131	Bluepoint Medical	A
0014925	Disposable SpO2 Sensors	00009299	1020131142	Bluepoint Medical	A
0014926	Disposable SpO2 Sensors	00009299	1020131145	Bluepoint Medical	A
0014927	Disposable SpO2 Sensors	00009299	1020131148	Bluepoint Medical	A
0014928	Disposable SpO2 Sensors	00009299	1020131150	Bluepoint Medical	A
0014957	Disposable SpO2 Sensors	00009299	1020131164	Bluepoint Medical	A
0014958	Disposable SpO2 Sensors	00009299	1020131167	Bluepoint Medical	A
0014959	Disposable SpO2 Sensors	00009299	1020131069	Bluepoint Medical	A
0014960	Disposable SpO2 Sensors	00009299	1020131071	Bluepoint Medical	A
0014590	SpO2 Silicone Finger Sen	00009300	302132142	Bluepoint Medical	A
0014591	SpO2 Silicone Finger Sen	00009300	3020132143	Bluepoint Medical	A
0014592	SpO2 Wrap Sensor- Paedia	00009300	5020132008	Bluepoint Medical	A
0014601	SpO2 Finger Sensor - Adu	00009300	1020132179	Bluepoint Medical	A
0014602	SpO2 Finger Sensor - Adu	00009300	1020132180	Bluepoint Medical	A
0014603	SpO2 Finger Sensor - Adu	00009300	1020132181	Bluepoint Medical	A
0014604	SpO2 Finger Sensor - Adu	00009300	1020132182	Bluepoint Medical	A
0014605	SpO2 Finger Sensor - Adu	00009300	1020132183	Bluepoint Medical	A
0014606	SpO2 Finger Sensor - Adu	00009300	1020132184	Bluepoint Medical	A
0014607	SpO2 Finger Sensor - Adu	00009300	1020132185	Bluepoint Medical	A
0014608	SpO2 Finger Sensor - Adu	00009300	1020132186	Bluepoint Medical	A
0014609	SpO2 Finger Sensor - Adu	00009300	1020132187	Bluepoint Medical	A
0014610	SpO2 Finger Sensor - Adu	00009300	1020132188	Bluepoint Medical	A
0014611	SpO2 Finger Sensor - Adu	00009300	1020132189	Bluepoint Medical	A
0014612	SpO2 Finger Sensor - Adu	00009300	1020132190	Bluepoint Medical	A
0014617	SpO2 Finger Sensor - Adu	00009300	1020132195	Bluepoint Medical	A
0014618	SpO2 Finger Sensor - Adu	00009300	1020132196	Bluepoint Medical	A
0014625	SpO2 Finger Sensor - Adu	00009300	4020132005	Bluepoint Medical	A
0014650	SpO2 Finger Sensor - Adu	00009300	1020132200	Bluepoint Medical	A
0014701	SpO2 Silicone Finger Sen	00009300	1020132223	Bluepoint Medical	A
0014702	SpO2 Silicone Finger Sen	00009300	1020132224	Bluepoint Medical	A
0014703	SpO2 Silicone Finger Sen	00009300	1020132225	Bluepoint Medical	A

Stock Ref.	Description	Supplier A/C	Supplier P/N	Supplier Name	Rating
0014704	SpO2 Silicone Finger Sen	00009300	1020132226	Bluepoint Medical	A
0014705	SpO2 Silicone Finger Sen	00009300	1020132227	Bluepoint Medical	A
0014706	SpO2 Silicone Finger Sen	00009300	1020132228	Bluepoint Medical	A
0014707	SpO2 Silicone Finger Sen	00009300	1020132229	Bluepoint Medical	A
0014708	SpO2 Silicone Finger Sen	00009300	1020132230	Bluepoint Medical	A
0014709	SpO2 Silicone Finger Sen	00009300	1020132231	Bluepoint Medical	A
0014710	SpO2 Silicone Finger Sen	00009300	1020132232	Bluepoint Medical	A
0014711	SpO2 Silicone Finger Sen	00009300	1020132233	Bluepoint Medical	A
0014712	SpO2 Silicone Finger Sen	00009300	1020132234	Bluepoint Medical	A
0014713	SpO2 Silicone Finger Sen	00009300	1020132235	Bluepoint Medical	A
0014714	SpO2 Silicone Finger Sen	00009300	1020132236	Bluepoint Medical	A
0014715	SpO2 Silicone Finger Sen	00009300	1020132237	Bluepoint Medical	A
0014721	SpO2 Silicone Finger Sen	00009300	1020132243	Bluepoint Medical	A
0014725	SpO2 Silicone Finger Sen	00009300	4020132001	Bluepoint Medical	A
0014750	SpO2 Silicone Finger Sen	00009300	1020132245	Bluepoint Medical	A
0014751	SpO2 Silicone Finger Sen	00009300	1020132302	Bluepoint Medical	A
0014755	SpO2 Silicone Finger Sen	00009300	102132301	Bluepoint Medical	A
0014756	SpO2 Silicone Finger Sen	00009300	2020132102	Bluepoint Medical	A
0014761	SpO2 Silicone Finger Sen	00009300	2020132060	Bluepoint Medical	A
0014762	SpO2 Silicone Finger Sen	00009300	2020132061	Bluepoint Medical	A
0014763	SpO2 Silicone Finger Sen	00009300	2020132082	Bluepoint Medical	A
0014764	SpO2 Silicone Finger Sen	00009300	2020132083	Bluepoint Medical	A
0014766	SpO2 Silicone Finger Sen	00009300	2020132085	Bluepoint Medical	A
0014769	SpO2 Silicone Finger Sen	00009300	2020132088	Bluepoint Medical	A
0014770	SpO2 Silicone Finger Sen	00009300	2020132089	Bluepoint Medical	A
0014771	SpO2 Silicone Finger Sen	00009300	2020132090	Bluepoint Medical	A
0014774	SpO2 Silicone Finger Sen	00009300	2020132093	Bluepoint Medical	A
0014775	SpO2 Silicone Finger Sen	00009300		Bluepoint Medical	A
0014778	SpO2 Silicone Finger Sen	00009300	2020132097	Bluepoint Medical	A
0014785	SpO2 Silicone Finger Sen	00009300		Bluepoint Medical	A
0014790	SpO2 Silicone Finger Sen	00009300		Bluepoint Medical	A
0014801	Pulse Oximetry Ear Senso	00009300	1020132255	Bluepoint Medical	A
0014802	Pulse Oximetry Ear Senso	00009300	1020132256	Bluepoint Medical	A
0014804	Pulse Oximetry Ear Senso	00009300	102132258	Bluepoint Medical	A
0014805	Pulse Oximetry Ear Senso	00009300	1020132259	Bluepoint Medical	A
0014806	Pulse Oximetry Ear Senso	00009300	1020132251	Bluepoint Medical	A
0014807	Pulse Oximetry Ear Senso	00009300	1020132261	Bluepoint Medical	A
0014816	Pulse Oximetry Ear Senso	00009300	6020132001	Bluepoint Medical	A
0014820	Pulse Oximetry Wrap Sens	00009300		Bluepoint Medical	A
0014822	Pulse Oximetry Wrap Sens	00009300	2020132114	Bluepoint Medical	A
0014823	Pulse Oximetry Wrap Sens	00009300	202132115	Bluepoint Medical	A
0014824	Pulse Oximetry Wrap Sens	00009300	2020132123	Bluepoint Medical	A
0014825	Pulse Oximetry Ear Senso	00009300	1020132263	Bluepoint Medical	A
0014827	Pulse Oximetry Wrap Sens	00009300	2020132107	Bluepoint Medical	A
0014828	Pulse Oximetry Wrap Sens	00009300	2020132108	Bluepoint Medical	A

Stock Ref.	Description	Supplier A/C	Supplier P/N	Supplier Name	Rating
0014829	Pulse Oximetry Wrap Sens	00009300	2020132121	Bluepoint Medical	A
0014830	Pulse Oximetry Wrap Sens	00009300	2020132122	Bluepoint Medical	A
0014831	Pulse Oximetry Wrap Sens	00009300	0014031	Bluepoint Medical	A
0014835	Pulse Oximetry Wrap Sens	00009300	2020132008	Bluepoint Medical	A
0014848	Pulse Oximetry Wrap Sens	00009300		Bluepoint Medical	A
0014895	Pulse Oximetry Extension	00009300	1020132288	Bluepoint Medical	A
0014896	Pulse Oximetry Extension	00009300	1020132296	Bluepoint Medical	A
0014901	Disposable SpO2 Sensors	00009300	1020131122	Bluepoint Medical	A
0014902	Disposable SpO2 Sensors	00009300	1020131125	Bluepoint Medical	A
0014903	Disposable SpO2 Sensors	00009300	1020131128	Bluepoint Medical	A
0014904	Disposable SpO2 Sensors	00009300	1020131130	Bluepoint Medical	A
0014905	Disposable SpO2 Sensors	00009300	1020131123	Bluepoint Medical	A
0014906	Disposable SpO2 Sensors	00009300	1020131126	Bluepoint Medical	A
0014909	Disposable SpO2 Sensors	00009300	1020131124	Bluepoint Medical	A
0014910	Disposable SpO2 Sensors	00009300	1020131127	Bluepoint Medical	A
0014911	Disposable SpO2 Sensors	00009300	1020131129	Bluepoint Medical	A
0014912	Disposable SpO2 Sensors	00009300	1020131131	Bluepoint Medical	A
0014921	Disposable SpO2 Sensors	00009300	1020131134	Bluepoint Medical	A
0014922	Disposable SpO2 Sensors	00009300	1020131137	Bluepoint Medical	A
0014923	Disposable SpO2 Sensors	00009300	1020131139	Bluepoint Medical	A
0014924	Disposable SpO2 Sensors	00009300	1020131141	Bluepoint Medical	A
0014940	Disposable SpO2 Sensors	00009300	1020131160	Bluepoint Medical	A
0014957	Disposable SpO2 Sensors	00009300	1020131164	Bluepoint Medical	A
0014958	Disposable SpO2 Sensors	00009300	1020131167	Bluepoint Medical	A
0014959	Disposable SpO2 Sensors	00009300	1020131069	Bluepoint Medical	A
0014960	Disposable SpO2 Sensors	00009300	1020131071	Bluepoint Medical	A
0014964	Disposable SpO2 Sensors	00009300		Bluepoint Medical	A
0014981	Disposable SpO2 Sensors	00009300		Bluepoint Medical	A
0014982	Disposable SpO2 Sensors	00009300		Bluepoint Medical	A
0014983	Disposable SpO2 Sensors	00009300		Bluepoint Medical	A
0014984	Disposable SpO2 Sensors	00009300		Bluepoint Medical	A
0014990	Disposable SpO2 Sensors	00009300	1020131206	Bluepoint Medical	A
0014993	Disposable SpO2 Sensors	00009300	1020131204	Bluepoint Medical	A
0014994	Disposable SpO2 Sensors	00009300	1020131207	Bluepoint Medical	A
0014995	Disposable SpO2 Sensors	00009300	1020131209	Bluepoint Medical	A
0014996	Disposable SpO2 Sensors	00009300	1020131211	Bluepoint Medical	A
0019001	Nellcor DEC-4 compatible	00009300		Bluepoint Medical	A
0019002	Nellcor DEC-8 compatible	00009300		Bluepoint Medical	A
0031238	Connector kit Philips/H	00012399	KT-087	Unimed	B
0031256	Connector kit Nihon Koh	00012399	KT-084	Unimed	B

Comments on Suppliers:

Suppliers up to date

# Sales Information

Stock Reference	Description	2011	2012	2013	2014	2015	2016	2017
0014590	SpO2 Silicone Finger Sen						2	
0014591	SpO2 Silicone Finger Sen		1				4	
0014592	SpO2 Wrap Sensor- Paedia			1	11	6	2	5
0014595	SpO2 Silicone Finger Sensor - Adult.							
0014596	SpO2 Silicone Finger Sensor - Paediatric							
0014601	SpO2 Finger Sensor - Adu	105	6	20	25	4	31	6
0014602	SpO2 Finger Sensor - Adu	9	0	2	226	245	1	
0014603	SpO2 Finger Sensor - Adu	3						0
0014604	SpO2 Finger Sensor - Adu			11	10		6	4
0014605	SpO2 Finger Sensor - Adu	-2	5	4				
0014606	SpO2 Finger Sensor - Adu	163	54	50	42	42	8	2
0014607	SpO2 Finger Sensor - Adu	110	21	7	6		0	
0014608	SpO2 Finger Sensor - Adu	38	22	16	24	12	12	8
0014609	SpO2 Finger Sensor - Adu	36	5	5	12	6		
0014610	SpO2 Finger Sensor - Adu	14	8	6	6			1
0014611	SpO2 Finger Sensor - Adu	11	3	7	1	28	2	3
0014612	SpO2 Finger Sensor - Adu	70	14	11	18	5	11	10
0014613	SpO2 Finger Sensor - Adult.		3					
0014614	SpO2 Finger Sensor - Adult.							
0014615	SpO2 Finger Sensor - Adult.							
0014617	SpO2 Finger Sensor - Adu			3				
0014618	SpO2 Finger Sensor - Adu			5	15	5	4	3
0014619	SpO2 Finger Sensor - Adult.							
0014620	SpO2 Finger Sensor - Adult.							
0014621	SpO2 Finger Sensor - Adult.	15						
0014623	SpO2 Finger Sensor - Adult.			8				
0014625	SpO2 Finger Sensor - Adu	83	28	32	66	15	8	10
0014649	SpO2 Finger Sensor - Adult.			45	5	1		
0014650	SpO2 Finger Sensor - Adu	255	228	202	165	189	205	76
0014701	SpO2 Silicone Finger Sen	64	17	2	14	1	2	
0014702	SpO2 Silicone Finger Sen	16	1	1	20	5		
0014703	SpO2 Silicone Finger Sen	8	18		1			0
0014704	SpO2 Silicone Finger Sen	134	77	76	94	40	51	85
0014705	SpO2 Silicone Finger Sen		2		1			
0014706	SpO2 Silicone Finger Sen	71	13	4			0	
0014707	SpO2 Silicone Finger Sen			15			0	
0014708	SpO2 Silicone Finger Sen		1				4	
0014709	SpO2 Silicone Finger Sen	2				1		
0014710	SpO2 Silicone Finger Sen		2					
0014711	SpO2 Silicone Finger Sen				2	8	5	
0014712	SpO2 Silicone Finger Sen			10	1	4		
0014713	SpO2 Silicone Finger Sen				2			

Stock Reference	Description	2011	2012	2013	2014	2015	2016	2017
0014714	SpO2 Silicone Finger Sen							
0014715	SpO2 Silicone Finger Sen	86	12	37	23	24	2	11
0014717	SpO2 Silicone Finger Sensor - Adult.							
0014718	SpO2 Silicone Finger Sensor - Adult.							
0014719	SpO2 Silicone Finger Sensor - Adult.							
0014720	SpO2 Silicone Finger Sensor - Adult.							
0014721	SpO2 Silicone Finger Sen							
0014722	SpO2 Silicone Finger Sensor - Adult.							
0014723	SpO2 Silicone Finger Sensor - Adult.	31	30	23	45	25	26	39
0014724	SpO2 Silicone Finger Sensor - Adult.		11	8	15	28	13	4
0014725	SpO2 Silicone Finger Sen	157	3	8	6	5	5	1
0014740	SpO2 Silicone Finger Sen							
0014750	SpO2 Silicone Finger Sen	540	409	654	503	305	378	425
0014751	SpO2 Silicone Finger Sen	88	47	51	43	66	56	15
0014755	SpO2 Silicone Finger Sen				2	3		
0014756	SpO2 Silicone Finger Sen						6	
0014761	SpO2 Silicone Finger Sen	16		22	11			
0014762	SpO2 Silicone Finger Sen	81	9	25	9	18	2	2
0014763	SpO2 Silicone Finger Sen		1					
0014764	SpO2 Silicone Finger Sen	5		3	3			
0014765	SpO2 Silicone Finger Sensor - Paediatric							
0014766	SpO2 Silicone Finger Sen	1	5	2				
0014767	SpO2 Silicone Finger Sensor - Paediatric							
0014768	SpO2 Silicone Finger Sensor - Paediatric							
0014769	SpO2 Silicone Finger Sen	1						
0014770	SpO2 Silicone Finger Sen		1					
0014771	SpO2 Silicone Finger Sen	6					2	
0014772	SpO2 Silicone Finger Sensor - Paediatric							
0014773	SpO2 Silicone Finger Sensor - Paediatric							
0014774	SpO2 Silicone Finger Sen			0				
0014775	SpO2 Silicone Finger Sen			5				1
0014777	SpO2 Silicone Finger Sensor - Paediatric							
0014778	SpO2 Silicone Finger Sen							
0014779	SpO2 Silicone Finger Sensor - Paediatric							
0014780	SpO2 Silicone Finger Sensor - Paediatric							
0014781	SpO2 Silicone Finger Sensor - Paediatric							
0014782	SpO2 Silicone Finger Sensor - Paediatric							
0014785	SpO2 Silicone Finger Sen	67	7	1	1	3		
0014790	SpO2 Silicone Finger Sen							
0014801	Pulse Oximetry Ear Senso	7	12	8	16	26	13	
0014802	Pulse Oximetry Ear Senso	11	9	28	71	54	2	2
0014803	Pulse Oximetry Ear Sensor							
0014804	Pulse Oximetry Ear Senso	18	20	25	6		0	
0014805	Pulse Oximetry Ear Senso						0	
0014806	Pulse Oximetry Ear Senso							

Stock Reference	Description	2011	2012	2013	2014	2015	2016	2017
0014807	Pulse Oximetry Ear Senso		6	3	5	5		6
0014808	Pulse Oximetry Ear Sensor							
0014816	Pulse Oximetry Ear Senso		18	17	16	12	1	3
0014820	Pulse Oximetry Wrap Sens							
0014821	Pulse Oximetry Wrap Sensor		4	15			1	
0014822	Pulse Oximetry Wrap Sens	3	5			0		
0014823	Pulse Oximetry Wrap Sens		4					
0014824	Pulse Oximetry Wrap Sens	7	7	6	10	1		
0014825	Pulse Oximetry Ear Senso	55	34	70	82	104	132	94
0014826	Pulse Oximetry Wrap Sensor							
0014827	Pulse Oximetry Wrap Sens							
0014828	Pulse Oximetry Wrap Sens	6	3					
0014829	Pulse Oximetry Wrap Sens			1		1		
0014830	Pulse Oximetry Wrap Sens	1						
0014831	Pulse Oximetry Wrap Sens					13	22	
0014832	Pulse Oximetry Wrap Sensor							
0014833	Pulse Oximetry Wrap Sensor							
0014834	Pulse Oximetry Wrap Sensor							
0014835	Pulse Oximetry Wrap Sens	39	36	60	26	18	35	79
0014836	Pulse Oximetry Wrap Sensor							
0014837	Pulse Oximetry Wrap Sensor							
0014838	Pulse Oximetry Wrap Sensor							
0014840	Pulse Oximetry Wrap Sensor							
0014841	Pulse Oximetry Wrap Sensor							
0014842	Pulse Oximetry Wrap Sensor							
0014843	Pulse Oximetry Wrap Sensor							
0014844	Pulse Oximetry Wrap Sensor							
0014845	Pulse Oximetry Multi-Site Y Sensor							
0014848	Pulse Oximetry Wrap Sens		22					
0014895	Pulse Oximetry Extension	25	26	35	36	18	101	184
0014896	Pulse Oximetry Extension	4	3	3				2
0014901	Disposable SpO2 Sensors	131	22	15	10	3	4	
0014902	Disposable SpO2 Sensors	10	15	15	13	32	47	45
0014903	Disposable SpO2 Sensors	24			3		6	
0014904	Disposable SpO2 Sensors	69	18	15		20	130	64
0014905	Disposable SpO2 Sensors	58	49	42	45	24	60	59
0014906	Disposable SpO2 Sensors	46	49	5	15			
0014909	Disposable SpO2 Sensors	37		318	499	658	98	22
0014910	Disposable SpO2 Sensors	16	5	13	35		2	
0014911	Disposable SpO2 Sensors	68	30	232	84	8	16	13
0014912	Disposable SpO2 Sensors	158	206	66	98	47	154	27
0014913	Disposable SpO2 Sensors - Adult.							
0014914	Disposable SpO2 Sensors - Paediatric.							
0014915	Disposable SpO2 Sensors - Infant.							
0014916	Disposable SpO2 Sensors - Neonatal.							

Stock Reference	Description	2011	2012	2013	2014	2015	2016	2017
0014917	Disposable SpO2 Sensors - Adult.							
0014918	Disposable SpO2 Sensors - Paediatric.							
0014921	Disposable SpO2 Sensors				1			
0014922	Disposable SpO2 Sensors				1			
0014923	Disposable SpO2 Sensors							
0014924	Disposable SpO2 Sensors	1	2	3	2	2	1	
0014925	Disposable SpO2 Sensors		6					
0014926	Disposable SpO2 Sensors							
0014927	Disposable SpO2 Sensors							
0014928	Disposable SpO2 Sensors							
0014929	Disposable SpO2 Sensors - Adult.							
0014930	Disposable SpO2 Sensors - Paediatric.							
0014933	Disposable SpO2 Sensors - Adult.							
0014934	Disposable SpO2 Sensors - Paediatric.							
0014935	Disposable SpO2 Sensors - Infant.							
0014936	Disposable SpO2 Sensors - Neonatal.							
0014937	Disposable SpO2 Sensors - Adult.							
0014938	Disposable SpO2 Sensors - Paediatric.							
0014939	Disposable SpO2 Sensors - Infant.							
0014940	Disposable SpO2 Sensors		5					
0014941	Disposable SpO2 Sensors - Adult.							
0014942	Disposable SpO2 Sensors - Paediatric.							
0014945	Disposable SpO2 Sensors - Adult.							
0014946	Disposable SpO2 Sensors - Paediatric.							
0014947	Disposable SpO2 Sensors - Infant.							
0014948	Disposable SpO2 Sensors - Neonatal.							
0014949	Disposable SpO2 Sensors - Adult.							
0014950	Disposable SpO2 Sensors - Paediatric.							
0014951	Disposable SpO2 Sensors - Infant.							
0014952	Disposable SpO2 Sensors - Neonatal.							
0014953	Disposable SpO2 Sensors - Adult.							
0014954	Disposable SpO2 Sensors - Paediatric.							
0014957	Disposable SpO2 Sensors	8	4	12	51	11	8	8
0014958	Disposable SpO2 Sensors	18	18	24	55	27	16	29
0014959	Disposable SpO2 Sensors	17	28	45	90	48	40	30
0014960	Disposable SpO2 Sensors	5	24	25	29	13	22	8
0014961	Disposable SpO2 Sensors - Adult.							
0014962	Disposable SpO2 Sensors - Paediatric.							
0014963	Disposable SpO2 Sensors - Infant.							
0014964	Disposable SpO2 Sensors	13	3	2	4	2	2	1
0014965	Disposable SpO2 Sensors - Adult.							
0014966	Disposable SpO2 Sensors - Paediatric.							
0014969	Disposable SpO2 Sensors - Adult.							
0014970	Disposable SpO2 Sensors - Paediatric.							
0014971	Disposable SpO2 Sensors - Infant.							

Stock Reference	Description	2011	2012	2013	2014	2015	2016	2017
0014972	Disposable SpO2 Sensors - Neonatal.	1						
0014973	Disposable SpO2 Sensors - Adult.							
0014974	Disposable SpO2 Sensors - Paediatric.							
0014975	Disposable SpO2 Sensors - Infant.							
0014976	Disposable SpO2 Sensors - Neonatal.							
0014977	Disposable SpO2 Sensors - Adult.							
0014978	Disposable SpO2 Sensors - Paediatric.							
0014981	Disposable SpO2 Sensors					5	7	13
0014982	Disposable SpO2 Sensors					11	34	25
0014983	Disposable SpO2 Sensors					30	58	75
0014984	Disposable SpO2 Sensors					9		
0014985	Disposable SpO2 Sensors - Adult.							
0014986	Disposable SpO2 Sensors - Paediatric.							
0014987	Disposable SpO2 Sensors - Infant.							
0014988	Disposable SpO2 Sensors - Neonatal.							
0014989	Disposable SpO2 Sensors - Adult.	1		7	11	2	2	2
0014990	Disposable SpO2 Sensors	3		1		5	34	23
0014993	Disposable SpO2 Sensors	2			7	44	12	8
0014994	Disposable SpO2 Sensors	3		10	17	30	51	53
0014995	Disposable SpO2 Sensors	6	5	4	46	5	5	4
0014996	Disposable SpO2 Sensors	27	28	22	4	8	26	16
0019001	Nellcor DEC-4 compatible	50	18	6	5	8	5	10
0019002	Nellcor DEC-8 compatible	3	1	2	5			2

Comments on Sales Information:

consistent sales in reasonable numbers

# Countries Review

Country	2011	2012	2013	2014	2015	2016	2017
AU Austria	[X]	[X]	[X]	[X]	[X]	[X]	[X]
B Belgium	[X]	[X]	[X]	[X]	[X]	[X]	[X]
BA Kingdom of Bahrain			[X]			[X]	[X]
C China			[X]	[X]			
CAN Canada			[X]				
CH Chile		[X]					
CI Canary Islands	[X]						
CRO Croatia	[X]						
CY Cyprus	[X]		[X]				
CZ Czech Republic	[X]	[X]	[X]	[X]	[X]	[X]	[X]
DE Denmark	[X]	[X]	[X]	[X]	[X]	[X]	[X]
EG Egypt	[X]						
ES Estonia	[X]		[X]	[X]	[X]		[X]
F France	[X]	[X]	[X]	[X]	[X]	[X]	[X]
FI Finland	[X]	[X]	[X]	[X]	[X]		
G Germany	[X]	[X]	[X]	[X]			[X]
GR Greece	[X]	[X]	[X]	[X]	[X]	[X]	[X]
HK Hong Kong	[X]	[X]	[X]		[X]	[X]	
I Iran	[X]	[X]	[X]				
IN India	[X]	[X]					
IND Indonesia			[X]				[X]
IQ IRAQ			[X]	[X]			
IRE Ireland	[X]	[X]	[X]	[X]	[X]	[X]	[X]
IS Israel		[X]		[X]	[X]		
JO Jordan		[X]				[X]	
KE Kenya		[X]			[X]		
KU Kuwait	[X]			[X]	[X]	[X]	
LEB Lebanon		[X]					
LI Lithuania						[X]	
MA Malaysia	[X]	[X]	[X]	[X]	[X]	[X]	
MAU Mauritius			[X]				
MEX Mexico				[X]		[X]	
NE Netherlands		[X]			[X]	[X]	
NG Nigera							
NO Norway	[X]	[X]	[X]	[X]	[X]	[X]	[X]
NZ New Zealand			[X]		[X]		
OM Oman		[X]	[X]	[X]	[X]	[X]	[X]
P Poland	[X]	[X]	[X]	[X]	[X]	[X]	[X]
PE Peru						[X]	[X]
PO Portugal	[X]	[X]	[X]	[X]	[X]	[X]	[X]
RO Romania	[X]	[X]	[X]		[X]	[X]	[X]
RU Russia			[X]		[X]		

Country	2011	2012	2013	2014	2015	2016	2017
SA South Africa	[X]	[X]					
SAU Saudi Arabia	[X]	[X]	[X]	[X]	[X]	[X]	
SI Singapore		[X]					
SL Slovenia			[X]	[X]	[X]		
SP Spain	[X]	[X]	[X]	[X]	[X]	[X]	[X]
SRI Sri Lanka			[X]				
SW Sweden	[X]	[X]	[X]	[X]	[X]	[X]	[X]
SWI Switzerland	[X]		[X]	[X]	[X]	[X]	[X]
SY Syria	[X]						
T Turkey	[X]	[X]	[X]	[X]	[X]		[X]
TH Thailand	[X]	[X]	[X]	[X]	[X]		
UAE U A E	[X]	[X]	[X]	[X]	[X]	[X]	[X]
UG UGANDA			[X]				
UK United Kingdom	[X]	[X]	[X]	[X]	[X]	[X]	[X]
VI Vietnam	[X]	[X]	[X]	[X]	[X]		

Comments on Sales to Countries:

selling to multiple countries

Comments on Risks with Sales to Countries:

no risks with countries

# Returns and Q.A. Fails Review

Stock Reference	Fault	2011	2012	2013	2014	2015	2016	2017
0014592	Unchecked - Returned To Stock					1		
0014601	Faulty Probe	1						
0014601	Intermittent	1						
0014601	Low SpO2 Readings	1						
0014601	New			1			1	
0014601	Unchecked - Returned To Stock							
0014602	Needs Evaluating				30	51		
0014602	Faulty Probe				9	2		
0014602	Unchecked - Returned To Stock		6					5
0014603	Needs Evaluating							4
0014605	Unchecked - Returned To Stock	3						
0014606	Connector Damaged	1						
0014606	Damaged Finger Clip	1						
0014606	Faulty Electronics	1						
0014606	Faulty Probe	2						
0014606	Intermittent	1	1					
0014606	New			1				
0014606	Unchecked - Returned To Stock				1	2		
0014607	Needs Evaluating						1	
0014607	Faulty Probe	1	2					
0014607	No Fault Return To Stock				2			
0014607	Unchecked - Returned To Stock	4	6	2				
0014608	Break At The Connector				1			
0014608	Faulty Probe		2					
0014608	Intermittant	1						
0014608	Intermittent		2					
0014608	Investigate	1						
0014608	New				1			
0014608	No Fault Found	1	2					
0014608	Photo-diode Intermittent In Finger Clip					1		
0014609	Unchecked - Returned To Stock	2						
0014610	Unchecked - Returned To Stock			1				
0014611	Unchecked - Returned To Stock				1			
0014612	Unchecked - Returned To Stock			2				1
0014625	Cable Replaced Connectors Rewired						1	
0014625	Intermittent	1	1					
0014649	Faulty Probe			3	3	1		
0014649	Sensor Fault			3	2			
0014650	Needs Evaluating							
0014650	Break At The Connector				1			
0014650	Chip Shorts To Emitter			1				
0014650	Chip Shorts To Photodiode			4				

Stock Reference	Fault	2011	2012	2013	2014	2015	2016	2017
0014650	Emitter Intermittent At The Connector			2				
0014650	Faulty Probe	3	19	13	5	30	5	3
0014650	Intermittant					1		
0014650	Intermittent	2						
0014650	New				1	1	1	
0014650	No Fault Found	1	2					
0014650	Photodiode Intermittent At The Connector			2				
0014650	Sensor Fault		4	7	12	17	2	1
0014701	Intermittent	1	1					
0014701	IR Emitter Failed					1		
0014701	New	1						
0014701	No Fault Found		1					
0014701	Photo-diode Intermittent In Finger Clip					1		
0014701	Red Emitter Faulty.		1					
0014701	Unchecked - Returned To Stock					2		
0014702	New							
0014703	Needs Evaluating							2
0014703	Faulty Electronics							
0014703	Faulty Probe	1	1					
0014703	Silicone Moulding Split							
0014703	Unchecked - Returned To Stock	1						
0014704	Break At The Connector		4					
0014704	Damaged Finger Clip							
0014704	Faulty Electronics	1						
0014704	Faulty Probe		7					
0014704	Intermittant							
0014704	Intermittent	1	2					
0014704	New		4					
0014704	No Fault Found	1						
0014704	No Fault Return To Stock							
0014704	Red Emitter Faulty.		1					
0014704	Unit Dead		1					
0014706	Connector Damaged	2						
0014706	New		2					
0014706	No Fault Return To Stock						2	
0014706	Not Working On Monitor	1						
0014707	Needs Evaluating						1	
0014708	New							
0014709	No Fault Found					1		
0014715	Damaged Finger Clip							
0014715	Faulty Electronics							
0014715	Intermittent	2						
0014715	Not Working On Monitor							
0014750	Needs Evaluating	3		1		1		1
0014750	Break At Strain Relief							1

Stock Reference	Fault	2011	2012	2013	2014	2015	2016	2017
0014750	Break In Sensor Circuit					1		
0014750	Connector Damaged		1					
0014750	Damaged Finger Clip	1						
0014750	Declined Quote		2					
0014750	Faulty Electronics							
0014750	Faulty Probe	1	4	8	1	6	1	3
0014750	Intermittant			4				
0014750	Intermittent			1			1	1
0014750	New	5	1	6	8	14	12	7
0014750	No Fault Found	19	12	4	5	3		1
0014750	No Fault Return To Stock			1			3	
0014750	Not Working On Monitor							
0014750	Photodiode Intermittent At The Connector			2				
0014750	Sensor Fault		1	1		6	1	2
0014750	Silicone Moulding Split				1			
0014750	Unable To Repair							
0014750	Unchecked - Returned To Stock	20	1	23	1			
0014751	Needs Evaluating	1			1			
0014751	Faulty Probe		1					
0014751	New	1	1			2		
0014751	No Fault Found					1		
0014751	Silicone Moulding Split					2		
0014751	Unchecked - Returned To Stock							1
0014762	Unchecked - Returned To Stock				10			
0014766	New			1				
0014774	Unchecked - Returned To Stock			3				
0014801	New					2		
0014802	New						5	
0014802	No Fault Found						6	
0014804	Faulty Electronics	1						
0014804	No Fault Return To Stock						2	
0014804	Unchecked - Returned To Stock		2					
0014805	Needs Evaluating						1	
0014822	Unchecked - Returned To Stock					6		
0014825	Needs Evaluating	1						
0014825	Faulty Electronics			1				
0014825	New			1	1	2	6	4
0014825	No Fault Found	2	1	3	1			
0014825	Sensor Fault			1	1	2	1	1
0014825	Unable To Repair		1					
0014835	Needs Evaluating	1						
0014835	Intermittent			1				
0014835	New							
0014835	No Fault Found			1				
0014835	Unchecked - Returned To Stock			1				

Stock Reference	Fault	2011	2012	2013	2014	2015	2016	2017
0014895	Needs Evaluating	1						
0014895	New		1	1				
0014895	No Fault Found						1	
0014904	Needs Evaluating	2						
0014905	Unchecked - Returned To Stock	9						
0014909	No Fault Return To Stock							
0014910	No Fault Return To Stock							
0014911	No Fault Return To Stock	1						
0014911	Unchecked - Returned To Stock	10						
0014912	Unchecked - Returned To Stock		2					
0014924	Unchecked - Returned To Stock							
0014960	Unchecked - Returned To Stock		3					
0014994	Needs Evaluating							1
0014995	Unchecked - Returned To Stock		2					

Comments on Returns:

Comments on Risks with Returns and Potential Re-work:

# Design Changes Review

## Showing Documents Filed in Y 14 Design Changes

Comments on Design Changes:

No changes,  
but in process of moving to bluepoint labelled as the manufacturer.

Comments on Risks with Design Changes:

no notifications required

# User Instructions Review

## Showing Documents Filed in F 5 User Instructions

Document ID	Description	Date Added/Updated
16821	4000 Series SpO2 Sensors Reusable Finger Sensors User Instructio...	25/04/16
14707	4000 Series SpO2 Sensors Disposable Sensors User Instructions	29/01/15
14477	Pulse Oximeter Sensors BPM Instructions for Use / User Manual ...	24/10/14

Comments on User Instructions:

IFU updated,  
but in process of moving to bluepoint labelled as the manufacturer.

Comments on Risks User Instructions:

no risks with IFU

# Labels Review

## Showing Documents Filed in F 7 / F 8 Labels

Document ID	Description	Date Added/Updated
14961	Pulse Oximetry Sensors 4000 Series Labels	29/04/15
11735	Pulse Oximetry Extension Cable 0014896 Dum Bell Label	27/03/12
11734	Pulse Oximetry Extension Cable 0014895 Dum Bell Label	27/03/12
6694	Pulse Oximetry Sensors 4000 Series Labels probe label	13/11/09
6693	Pulse Oximetry Sensors 4000 Series Labels pkg label	13/11/09

Comments on Labels:

Labels I believe have been changed but I dont seem to have latest versions in my files

Comments on Risks Labels:

no risks with labels

# Documentation Updates / Changes

Document ID	Description	Date Added/Updated
23481	Bluepoint Pulse Oximetry Sensors 4000 Series Risk analysis repor...	27/10/17
18306	OBL Bluepoint Medical Contract OEM QA contract	21/12/16
18301	Bluepoint Pulse Oximetry Sensors 4000 Series Introduction to the...	19/12/16
18239	Bluepoint Pulse Oximetry Sensors 4000 Series Quality Assurance a...	13/12/16
18137	Bluepoint Pulse Oximetry Sensors 4000 Series User Feedback Lost ...	29/11/16
18037	Bluepoint Pulse Oximetry Sensors 4000 Series Details of Clinical...	15/11/16
18035	Bluepoint Pulse Oximetry Sensors 4000 Series Software verificati...	15/11/16
18033	Bluepoint Pulse Oximetry Sensors 4000 Series Detailed test resul...	15/11/16
18031	Bluepoint Pulse Oximetry Sensors 4000 Series Risk management doc...	15/11/16
18029	Bluepoint Pulse Oximetry Sensors 4000 Series Manufacturing proce...	15/11/16
18027	Bluepoint Pulse Oximetry Sensors 4000 Series Design input specif...	15/11/16
18025	Bluepoint Pulse Oximetry Sensors 4000 Series Solutions adopted t...	15/11/16
18023	Bluepoint Pulse Oximetry Sensors 4000 Series Classification rati...	15/11/16
18021	Bluepoint Pulse Oximetry Sensors 4000 Series medicinal substance...	15/11/16
18019	Bluepoint Pulse Oximetry Sensors 4000 Series Evidence of Usabili...	15/11/16
17858	Bluepoint Capnograph Location of responsibilities. documentatio...	09/11/16
17844	Bluepoint Pulse Oximetry Sensors 4000 Series EMC report IEC 6060...	04/11/16
17838	Bluepoint Capnograph Location of responsibilities.	03/11/16
17631	Bluepoint Pulse Oximetry Sensors 4000 Series Description of inte...	28/09/16
17628	Bluepoint Capnograph Quality Assurance and Agreements Appendix ...	28/09/16
17626	Bluepoint Details of significant subcontractors	28/09/16
17625	Bluepoint Location of responsibilities.	28/09/16
16821	4000 Series SpO2 Sensors Reusable Finger Sensors User Instructio...	25/04/16
16796	Pulse Oximetry Sensors 4000 Post Market Surveillance	20/04/16

## Comments on Document Changes:

added extra documentation from bluepoint,  
but we changing to be manufacturer officially by bluepoint using their CE

## Comments on Risks with Document Changes:

no risk with documentation

# Internal Issues Review

Number of Issues reviewed: 517

Issue ID	Subject
81382	Office Meeting Opera Partnumber / Prices Updates 0014623 HP Sensor
103552	Office Meeting Production 0014723
69433	Office Meeting Purchase Back Orders Review 0014904 lead times
77801	Office Meeting Stock Queries 0014985-0014988 Superseeded?
92062	Office Meeting Purchase Back Orders Review 20 x 0012161 is now in picking
91773	Office Meeting Document Control Amendments needed 4000 Series Labels (Czech Republic)
96777	Office Meeting Price Lists UK 4000 series sensors to remove from price lists
97766	Office Meeting Sales Account 00012906 Stop Status Changed to Stopped
97682	Office Meeting Sales Account 00012906 Stop Status Changed to Stopped
101279	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
99243	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
97191	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
95190	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
93293	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
91397	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
89766	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
87575	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
85878	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
83976	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
82132	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
79737	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
77317	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
75474	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)
73408	Office Meeting VIAMED Stock Meeting Acorn 0014904 (522)

Issue ID	Subject
106219	Office Meeting VIAMED Stock Meeting Acorn Distribution ac 12906 Stock levels for part number 0014904 (522)
103496	Office Meeting VIAMED Stock Meeting Acorn Distribution ac 12906 Stock levels for part number 0014904 (522)
78168	Office Meeting Audit Goods Inwards and Product Identity Audit 09 Product Identity - Stock in a incorrect location.
78167	Office Meeting Audit Goods Inwards and Product Identity Audit 09 Product Identity - Stock in a incorrect location.
78166	Office Meeting Audit Goods Inwards and Product Identity Audit 09 Product Identity - Stock in a incorrect location.
81887	Office Meeting Sales Back Orders Review - By Customer Backorder 00000333 0014996 ORD82901
83349	Office Meeting Sales Back Orders Review - By Customer Backorder 00000333 0014996 ORD83324
83689	Office Meeting Sales Back Orders Review - By Customer Backorder 00001038 0014750 ORD83404
89430	Office Meeting Sales Back Orders Review - By Customer Backorder 00001661 0014606 ORD84957
101669	Office Meeting Sales Back Orders Review - By Customer Backorder 00001810 0014750 ORD88040
97131	Office Meeting Sales Back Orders Review - By Customer Backorder 00002948 0014618 ORD86919
106018	Office Meeting Sales Back Orders Review - By Customer Backorder 00003007 0014750 ORD88668
106019	Office Meeting Sales Back Orders Review - By Customer Backorder 00003007 0014750 ORD88893
72195	Office Meeting Sales Back Orders Review - By Customer Backorder 00003007 0014756 ORD80491
72768	Office Meeting Sales Back Orders Review - By Customer Backorder 00003007 0014895 ORD80491
87590	Office Meeting Sales Back Orders Review - By Customer Backorder 00003310 0014989 ORD84510
95078	Office Meeting Sales Back Orders Review - By Customer Backorder 00003310 0014989 ORD86382
99964	Office Meeting Sales Back Orders Review - By Customer Backorder 00004640 0014650 ORD87652
90481	Office Meeting Sales Back Orders Review - By Customer Backorder 00004729 0014994 ORD85209
72511	Office Meeting Sales Back Orders Review - By Customer Backorder 00005030 0014608 ORD80644
73474	Office Meeting Sales Back Orders Review - By Customer Backorder 00005030 0014608 ORD80961
80265	Office Meeting Sales Back Orders Review - By Customer Backorder 00005110 0014802 ORD82558
75293	Office Meeting Sales Back Orders Review - By Customer Backorder 00005430 0014612 ORD81389
105696	Office Meeting Sales Back Orders Review - By Customer Backorder 00005430 0014612 ORD88878
83115	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014601 ORD83265
96581	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014603 ORD86790

Issue ID	Subject
95901	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014604 ORD86608
84147	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014904 ORD83523
72353	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014905 ORD80592
74166	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014905 ORD81123
85271	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014909 ORD83890
90564	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014909 ORD85235
91490	Office Meeting Sales Back Orders Review - By Customer Backorder 00006277 0014909 ORD85503
68257	Office Meeting Sales Back Orders Review - By Customer Backorder 00006297 0014590 ORD79055
87589	Office Meeting Sales Back Orders Review - By Customer Backorder 00006343 0014750 ORD84507
85574	Office Meeting Sales Back Orders Review - By Customer Backorder 00006343 0014751 ORD83971
102913	Office Meeting Sales Back Orders Review - By Customer Backorder 00006343 0014995 ORD88326
91697	Office Meeting Sales Back Orders Review - By Customer Backorder 00006350 0014993 ORD85546
96465	Office Meeting Sales Back Orders Review - By Customer Backorder 00006350 0014993 ORD86758
98525	Office Meeting Sales Back Orders Review - By Customer Backorder 00006350 0014993 ORD87237
99649	Office Meeting Sales Back Orders Review - By Customer Backorder 00006439 0014750 ORD87572
98536	Office Meeting Sales Back Orders Review - By Customer Backorder 00006441 0014750 ORD87228
88519	Office Meeting Sales Back Orders Review - By Customer Backorder 00006673 0019002 ORD84743
71878	Office Meeting Sales Back Orders Review - By Customer Backorder 00007125 0014750 ORD80383
94943	Office Meeting Sales Back Orders Review - By Customer Backorder 00007224 0014750 ORD86345
103379	Office Meeting Sales Back Orders Review - By Customer Backorder 00007246 0012101 ORD88417
98468	Office Meeting Sales Back Orders Review - By Customer Backorder 00007246 0014750 ORD87220
98968	Office Meeting Sales Back Orders Review - By Customer Backorder 00007246 0014750 ORD87372
75485	Office Meeting Sales Back Orders Review - By Customer Backorder 00007246 0014751 ORD81447
72568	Office Meeting Sales Back Orders Review - By Customer Backorder 00007246 0014825 ORD80658
85489	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014801 ORD83948
71170	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014903 ORD80179

Issue ID	Subject
74793	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014903 ORD81267
78610	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014903 ORD82225
71171	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014905 ORD80179
75217	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014905 ORD81371
89856	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014905 ORD85040
91578	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014905 ORD85482
96082	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014905 ORD86613
98800	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014905 ORD87330
100361	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014905 ORD87734
104640	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014905 ORD88382
95915	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014911 ORD86613
72305	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014912 ORD80538
93754	Office Meeting Sales Back Orders Review - By Customer Backorder 00007356 0014912 ORD86085
76517	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014601 ORD81774
76978	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014701 ORD81774
74881	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014704 ORD81306
89328	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014704 ORD84923
82063	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014708 ORD82958
74362	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014711 ORD81192
79060	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014715 ORD82335
97642	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014715 ORD87025
98213	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014715 ORD87167
70423	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014724 ORD79898
75587	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014724 ORD81413
78447	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014724 ORD82176
83867	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014724 ORD83472

Issue ID	Subject
88508	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014724 ORD84744
89550	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014724 ORD84976
89959	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014775 ORD85077
78550	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014801 ORD82201
76979	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014802 ORD81842
90767	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014802 ORD85285
90827	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014802 ORD85296
77657	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014957 ORD81974
87073	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014957 ORD84362
100367	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014957 ORD87746
71740	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD80361
80442	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD82606
87220	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD84387
90970	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD85328
92991	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD85840
95138	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD86397
98612	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD87258
99385	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD87514
102724	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD88248
106381	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014958 ORD88976
85043	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014959 ORD83819
89651	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014959 ORD84987
94175	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014959 ORD86171
101008	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014959 ORD87892
82304	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014981 ORD83046
95369	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014981 ORD86464

Issue ID	Subject
106459	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014981 ORD88958
70425	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014982 ORD79903
71850	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014982 ORD80361
91312	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014982 ORD85442
93484	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014982 ORD85840
97143	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014982 ORD86932
102015	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014982 ORD88088
105531	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014982 ORD88824
106651	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014982 ORD89043
74148	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD81119
84622	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD83670
87384	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD84387
88569	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD84745
89857	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD85043
98799	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD87328
99505	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD87514
100577	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD87779
101290	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD87926
101358	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD87973
102818	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0014983 ORD88248
78402	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0110017 ORD82172
78403	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0110022 ORD82172
78405	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0110040 ORD82172
78406	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 0110057 ORD82172
78404	Office Meeting Sales Back Orders Review - By Customer Backorder 00007374 2530004 ORD82073
75597	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014650 ORD81473

Issue ID	Subject
77945	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014650 ORD81820
87633	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014650 ORD84503
94400	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014650 ORD86038
99361	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014650 ORD87429
103440	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014650 ORD88430
76973	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014750 ORD81822
82078	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014750 ORD82949
85669	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014750 ORD83974
81166	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014825 ORD82703
84634	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014825 ORD83686
93426	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014825 ORD85913
91120	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014835 ORD85229
81149	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014994 ORD82704
91411	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014994 ORD85471
95445	Office Meeting Sales Back Orders Review - By Customer Backorder 00007515 0014994 ORD86396
79128	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014704 ORD82342
86793	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014704 ORD84232
88645	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014704 ORD84779
98382	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014704 ORD87203
100360	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014704 ORD87732
79823	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014723 ORD82342
81976	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014723 ORD82916
83350	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014723 ORD83326
90305	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014723 ORD85170
94506	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014723 ORD86252
97161	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014723 ORD86877

Issue ID	Subject
102997	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014723 ORD88332
74662	Office Meeting Sales Back Orders Review - By Customer Backorder 00007530 0014902 ORD81247
96781	Office Meeting Sales Back Orders Review - By Customer Backorder 00007539 0014723 ORD86832
71439	Office Meeting Sales Back Orders Review - By Customer Backorder 00007542 0014650 ORD80254
76284	Office Meeting Sales Back Orders Review - By Customer Backorder 00007542 0014825 ORD81705
98637	Office Meeting Sales Back Orders Review - By Customer Backorder 00007542 0014825 ORD87276
78598	Office Meeting Sales Back Orders Review - By Customer Backorder 00007592 0012160 ORD82215
86142	Office Meeting Sales Back Orders Review - By Customer Backorder 00007594 0014592 ORD84141
102491	Office Meeting Sales Back Orders Review - By Customer Backorder 00007594 0014751 ORD88122
88148	Office Meeting Sales Back Orders Review - By Customer Backorder 00007594 0014835 ORD84651
90149	Office Meeting Sales Back Orders Review - By Customer Backorder 00007594 0014835 ORD85113
102032	Office Meeting Sales Back Orders Review - By Customer Backorder 00007594 0014835 ORD88122
82620	Office Meeting Sales Back Orders Review - By Customer Backorder 00007688 0014596 ORD83093
85572	Office Meeting Sales Back Orders Review - By Customer Backorder 00007688 0014751 ORD83093
74308	Office Meeting Sales Back Orders Review - By Customer Backorder 00007688 0014996 ORD81150
80876	Office Meeting Sales Back Orders Review - By Customer Backorder 00007688 0014996 ORD82674
82674	Office Meeting Sales Back Orders Review - By Customer Backorder 00007688 0014996 ORD83093
89425	Office Meeting Sales Back Orders Review - By Customer Backorder 00007688 0014996 ORD84947
99014	Office Meeting Sales Back Orders Review - By Customer Backorder 00007786 0014750 ORD87397
77762	Office Meeting Sales Back Orders Review - By Customer Backorder 00007885 0014912 ORD81993
90475	Office Meeting Sales Back Orders Review - By Customer Backorder 00007995 0014750 ORD85173
90906	Office Meeting Sales Back Orders Review - By Customer Backorder 00008043 0014750 ORD85310
93131	Office Meeting Sales Back Orders Review - By Customer Backorder 00008044 0014816 ORD85915
72293	Office Meeting Sales Back Orders Review - By Customer Backorder 00009044 0014750 ORD80380
95284	Office Meeting Sales Back Orders Review - By Customer Backorder 00009044 0014750 ORD86442
99077	Office Meeting Sales Back Orders Review - By Customer Backorder 00009130 0014835 ORD87390

Issue ID	Subject
102810	Office Meeting Sales Back Orders Review - By Customer Backorder 00011982 0014994 ORD88301
81888	Office Meeting Sales Back Orders Review - By Customer Backorder 00011982 0014996 ORD82902
93753	Office Meeting Sales Back Orders Review - By Customer Backorder 00012605 0014601 ORD86073
85821	Office Meeting Sales Back Orders Review - By Customer Backorder 00012605 0014625 ORD84074
91604	Office Meeting Sales Back Orders Review - By Customer Backorder 00012605 0014625 ORD85443
99639	Office Meeting Sales Back Orders Review - By Customer Backorder 00012605 0014625 ORD87567
106654	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014751 ORD89038
70427	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD79888
74269	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD81159
75489	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD81455
77951	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD82041
77990	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD82064
84810	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD83721
87602	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD84517
88814	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD84837
89326	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD84922
99437	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD87518
106463	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014990 ORD88995
90480	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014994 ORD85208
90907	Office Meeting Sales Back Orders Review - By Customer Backorder 00012672 0014994 ORD85311
72814	Office Meeting Sales Back Orders Review - By Customer Backorder 00012748 0014750 ORD80758
79876	Office Meeting Sales Back Orders Review - By Customer Backorder 00012906 0014904 ORD82470
85050	Office Meeting Sales Back Orders Review - By Customer Backorder 00012906 0014904 ORD83406
90828	Office Meeting Sales Back Orders Review - By Customer Backorder 00012906 0014904 ORD85297
81913	Office Meeting Sales Back Orders Review - By Customer Backorder 00012907 0014996 ORD82919
69929	Office Meeting Sales Back Orders Review - By Customer Backorder 00012923 0014750 ORD79668

Issue ID	Subject
77969	Office Meeting Sales Back Orders Review - By Customer Backorder 00012923 0014750 ORD81910
69665	Office Meeting Sales Back Orders Review - By Customer Backorder 00012923 4410520 ORD79668
82400	Office Meeting Sales Back Orders Review - By Customer Backorder 00012969 0014750 ORD83065
92630	Office Meeting Sales Back Orders Review - By Customer Backorder 00013011 0014835 ORD85771
104944	Office Meeting Sales Back Orders Review - By Customer Backorder 00013041 0014835 ORD88704
69492	Office Meeting Mailshots UK BCI compatible disposables
69490	Office Meeting Mailshots UK BCI compatible disposables
69488	Office Meeting Mailshots UK BCI compatible disposables
81840	Office Meeting Sales Order Issues Bulk packing 0014904 disposable sensors
81839	Office Meeting Sales Order Issues Bulk packing 0014904 disposable sensors
90977	Office Meeting Telesales Call Log #2149 - Acorn cancelling back order
70803	Office Meeting Other Certificates of origin: 0012160...
85000	Office Meeting Telesales Change of shipping date for back order 83406
84999	Office Meeting Telesales Change of shipping date for back order 83406
84998	Office Meeting Telesales Change of shipping date for back order 83406
98189	Office Meeting Sales Viamed Medical Export Cortessis - Greek Ambulance Tender
98188	Office Meeting Sales Viamed Medical Export Cortessis - Greek Ambulance Tender
98187	Office Meeting Sales Viamed Medical Export Cortessis - Greek Ambulance Tender
103723	Office Meeting Sales Viamed Medical Export Cortessis Order (Viamed VM-2160)
70940	Office Meeting Stock Queries Declarations of Origin Required
70939	Office Meeting Stock Queries Declarations of Origin Required
74834	Office Meeting VIAMED General Issue Demo stock to move
92549	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87993	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87992	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87991	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors

Issue ID	Subject
87990	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87989	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87988	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87987	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87986	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87985	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87984	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87983	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87982	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87981	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
87980	Office Meeting Any other business Discontinued - Autoclavable pulse oximetry sensors
74259	Office Meeting Stock Processing Disposable probes
81818	Office Meeting Document Control Amendments needed Document Update Request (6321)
84461	Office Meeting GHX Web Pricing GHX Stock Group Master
83989	Office Meeting Cross Reference Database Tables Updates Maxtec SpO2 sensor p/n R134P09-001 is 0014701
96890	Office Meeting Sales Viamed Medical Export Oxylink and 0114723 stock
96889	Office Meeting Sales Viamed Medical Export Oxylink and 0114723 stock
96888	Office Meeting Sales Viamed Medical Export Oxylink and 0114723 stock
96887	Office Meeting Sales Viamed Medical Export Oxylink and 0114723 stock
94343	Office Meeting Any Other Business Parts for SBRI SpO2 trial
70852	Office Meeting Purchase Back Orders Review POR10638 0014904 Have Been Delivered or Cancelled in Opera
70853	Office Meeting Purchase Back Orders Review POR10655 0014904 Have Been Delivered or Cancelled in Opera
70544	Office Meeting Back Order Report POR10721 0014825 //
70694	Office Meeting Back Order Report POR10721 0014825 31/05/2016
70545	Office Meeting Back Order Report POR10721 0014982 //
70695	Office Meeting Back Order Report POR10721 0014982 31/05/2016

Issue ID	Subject
71158	Office Meeting Back Order Report POR10737 0014650 //
71209	Office Meeting Back Order Report POR10737 0014650 17/06/2016
71200	Office Meeting Back Order Report POR10743 0014902 //
71336	Office Meeting Back Order Report POR10743 0014902 17/06/2016
72435	Office Meeting Back Order Report POR10781 0014756 //
72439	Office Meeting Back Order Report POR10781 0014756 10/06/2016
74811	Office Meeting Back Order Report POR10832 0014711 //
74672	Office Meeting Back Order Report POR10832 0014711 //
75079	Office Meeting Back Order Report POR10832 0014711 31/08/2016
74801	Office Meeting Back Order Report POR10832 0014711 31/08/2016
74812	Office Meeting Back Order Report POR10832 0014905 //
74673	Office Meeting Back Order Report POR10832 0014905 //
75080	Office Meeting Back Order Report POR10832 0014905 31/08/2016
74802	Office Meeting Back Order Report POR10832 0014905 31/08/2016
79123	Office Meeting Purchase Back Orders Review POR10832 0014905 Have Been Delivered or Cancelled in Opera
74813	Office Meeting Back Order Report POR10832 0014912 //
74674	Office Meeting Back Order Report POR10832 0014912 //
75081	Office Meeting Back Order Report POR10832 0014912 31/08/2016
74803	Office Meeting Back Order Report POR10832 0014912 31/08/2016
74814	Office Meeting Back Order Report POR10832 0014990 //
74675	Office Meeting Back Order Report POR10832 0014990 //
75082	Office Meeting Back Order Report POR10832 0014990 31/08/2016
74804	Office Meeting Back Order Report POR10832 0014990 31/08/2016
75146	Office Meeting Back Order Report POR10845 0014825 //
75416	Office Meeting Back Order Report POR10845 0014825 06/09/2016
75147	Office Meeting Back Order Report POR10845 0014895 //

Issue ID	Subject
75417	Office Meeting Back Order Report POR10845 0014895 06/09/2016
77756	Office Meeting Back Order Report POR10901 0014912 //
77754	Office Meeting Back Order Report POR10901 0014912 //
78039	Office Meeting Back Order Report POR10901 0014912 05/10/2016
78678	Office Meeting Back Order Report POR10917 0014904 //
78070	Office Meeting Back Order Report POR10917 0014904 //
78051	Office Meeting Back Order Report POR10917 0014904 //
80068	Office Meeting Back Order Report POR10927 0014825 //
79911	Office Meeting Back Order Report POR10927 0014825 //
78474	Office Meeting Back Order Report POR10927 0014825 //
78854	Office Meeting Back Order Report POR10934 0014825 //
78849	Office Meeting Back Order Report POR10934 0014825 //
79046	Office Meeting Back Order Report POR10934 0014825 24/10/2016
78856	Office Meeting Back Order Report POR10935 0014750 //
78848	Office Meeting Back Order Report POR10935 0014750 //
79045	Office Meeting Back Order Report POR10935 0014750 24/10/2016
79914	Office Meeting Back Order Report POR10958 0014902 //
79372	Office Meeting Back Order Report POR10958 0014902 //
79367	Office Meeting Back Order Report POR10958 0014902 //
80043	Office Meeting Back Order Report POR10958 0014902 25/10/2016
80645	Office Meeting Back Order Report POR10974 0014895 //
80726	Office Meeting Back Order Report POR10974 0014895 04/11/2016
80646	Office Meeting Back Order Report POR10974 0014909 //
80727	Office Meeting Back Order Report POR10974 0014909 30/11/2016
80647	Office Meeting Back Order Report POR10974 0014993 //
80728	Office Meeting Back Order Report POR10974 0014993 30/11/2016

Issue ID	Subject
80648	Office Meeting Back Order Report POR10974 0014994 //
80729	Office Meeting Back Order Report POR10974 0014994 30/11/2016
82785	Office Meeting Back Order Report POR11004 0014904 //
81892	Office Meeting Back Order Report POR11004 0014904 //
81845	Office Meeting Back Order Report POR11004 0014904 //
83254	Office Meeting Back Order Report POR11027 0014596 //
83245	Office Meeting Back Order Report POR11027 0014596 //
82802	Office Meeting Back Order Report POR11027 0014596 //
83365	Office Meeting Back Order Report POR11027 0014596 12/12/2016
83255	Office Meeting Back Order Report POR11027 0014703 //
83246	Office Meeting Back Order Report POR11027 0014703 //
82803	Office Meeting Back Order Report POR11027 0014703 //
83366	Office Meeting Back Order Report POR11027 0014703 12/12/2016
83256	Office Meeting Back Order Report POR11027 0014751 //
83247	Office Meeting Back Order Report POR11027 0014751 //
82804	Office Meeting Back Order Report POR11027 0014751 //
83367	Office Meeting Back Order Report POR11027 0014751 10/11/2016
84985	Office Meeting Back Order Report POR11027 0014751 12/12/2016
83257	Office Meeting Back Order Report POR11027 0014835 //
83248	Office Meeting Back Order Report POR11027 0014835 //
82805	Office Meeting Back Order Report POR11027 0014835 //
83368	Office Meeting Back Order Report POR11027 0014835 12/12/2016
83258	Office Meeting Back Order Report POR11027 0014996 //
83249	Office Meeting Back Order Report POR11027 0014996 //
82806	Office Meeting Back Order Report POR11027 0014996 //
83369	Office Meeting Back Order Report POR11027 0014996 12/12/2016

Issue ID	Subject
83845	Office Meeting Back Order Report POR11052 0014650 //
84720	Office Meeting Back Order Report POR11052 0014650 06/01/2017
83846	Office Meeting Back Order Report POR11052 0014704 //
84721	Office Meeting Back Order Report POR11052 0014704 27/01/2017
83847	Office Meeting Back Order Report POR11052 0014750 //
84722	Office Meeting Back Order Report POR11052 0014750 06/01/2017
86829	Office Meeting Back Order Report POR11109 0014704 //
87065	Office Meeting Back Order Report POR11109 0014704 10/03/2017
86830	Office Meeting Back Order Report POR11109 0014835 //
87066	Office Meeting Back Order Report POR11109 0014835 10/03/2017
88833	Office Meeting Back Order Report POR11130 0014601 //
87781	Office Meeting Back Order Report POR11130 0014601 //
87773	Office Meeting Back Order Report POR11130 0014601 //
87425	Office Meeting Back Order Report POR11130 0014601 //
89674	Office Meeting Back Order Report POR11130 0014601 24/03/2017
88834	Office Meeting Back Order Report POR11130 0014905 //
87782	Office Meeting Back Order Report POR11130 0014905 //
87774	Office Meeting Back Order Report POR11130 0014905 //
87426	Office Meeting Back Order Report POR11130 0014905 //
89675	Office Meeting Back Order Report POR11130 0014905 13/04/2017
88835	Office Meeting Back Order Report POR11130 0014912 //
87783	Office Meeting Back Order Report POR11130 0014912 //
87775	Office Meeting Back Order Report POR11130 0014912 //
87427	Office Meeting Back Order Report POR11130 0014912 //
89676	Office Meeting Back Order Report POR11130 0014912 24/03/2017
88836	Office Meeting Back Order Report POR11130 0014960 //

Issue ID	Subject
87784	Office Meeting Back Order Report POR11130 0014960 //
87776	Office Meeting Back Order Report POR11130 0014960 //
87428	Office Meeting Back Order Report POR11130 0014960 //
89677	Office Meeting Back Order Report POR11130 0014960 13/04/2017
88838	Office Meeting Back Order Report POR11130 0014983 //
88837	Office Meeting Back Order Report POR11130 0014983 //
87785	Office Meeting Back Order Report POR11130 0014983 //
87777	Office Meeting Back Order Report POR11130 0014983 //
87429	Office Meeting Back Order Report POR11130 0014983 //
89679	Office Meeting Back Order Report POR11130 0014983 13/04/2017
89678	Office Meeting Back Order Report POR11130 0014983 13/04/2017
93228	Office Meeting Purchase Back Orders Review POR11130 0014983 Have Been Delivered or Cancelled in Opera
88839	Office Meeting Back Order Report POR11130 0014990 //
87786	Office Meeting Back Order Report POR11130 0014990 //
87778	Office Meeting Back Order Report POR11130 0014990 //
87430	Office Meeting Back Order Report POR11130 0014990 //
89680	Office Meeting Back Order Report POR11130 0014990 13/04/2017
91046	Office Meeting Back Order Report POR11189 0014835 //
90678	Office Meeting Back Order Report POR11189 0014835 //
91519	Office Meeting Back Order Report POR11215 0014905 //
91761	Office Meeting Back Order Report POR11215 0014905 12/05/2017
91520	Office Meeting Back Order Report POR11215 0014909 //
91762	Office Meeting Back Order Report POR11215 0014909 12/05/2017
91521	Office Meeting Back Order Report POR11215 0014982 //
91763	Office Meeting Back Order Report POR11215 0014982 12/05/2017
91522	Office Meeting Back Order Report POR11215 0014990 //

Issue ID	Subject
91764	Office Meeting Back Order Report POR11215 0014990 12/05/2017
91523	Office Meeting Back Order Report POR11215 0014994 //
91765	Office Meeting Back Order Report POR11215 0014994 12/05/2017
91524	Office Meeting Back Order Report POR11215 0014996 //
91766	Office Meeting Back Order Report POR11215 0014996 12/05/2017
97739	Office Meeting Back Order Report POR11321 0014904 //
97220	Office Meeting Back Order Report POR11321 0014904 //
97740	Office Meeting Back Order Report POR11321 0014912 //
97221	Office Meeting Back Order Report POR11321 0014912 //
97734	Office Meeting Back Order Report POR11322 0014650 //
97215	Office Meeting Back Order Report POR11322 0014650 //
97735	Office Meeting Back Order Report POR11322 0014703 //
97216	Office Meeting Back Order Report POR11322 0014703 //
99379	Office Meeting Back Order Report POR11322 0014703 30/08/2017
97736	Office Meeting Back Order Report POR11322 0014704 //
97217	Office Meeting Back Order Report POR11322 0014704 //
99380	Office Meeting Back Order Report POR11322 0014704 30/08/2017
97737	Office Meeting Back Order Report POR11322 0014750 //
97218	Office Meeting Back Order Report POR11322 0014750 //
99381	Office Meeting Back Order Report POR11322 0014750 30/08/2017
97738	Office Meeting Back Order Report POR11322 0014825 //
97219	Office Meeting Back Order Report POR11322 0014825 //
100700	Office Meeting Back Order Report POR11342 0014750 //
99288	Office Meeting Back Order Report POR11342 0014750 //
98639	Office Meeting Back Order Report POR11342 0014750 //
100701	Office Meeting Back Order Report POR11342 0014895 //

Issue ID	Subject
99289	Office Meeting Back Order Report POR11342 0014895 //
98640	Office Meeting Back Order Report POR11342 0014895 //
101300	Office Meeting Back Order Report POR11389 0014983 //
101245	Office Meeting Back Order Report POR11389 0014983 //
100697	Office Meeting Back Order Report POR11389 0014983 //
104659	Office Meeting Back Order Report POR11432 0014835 //
102897	Office Meeting Back Order Report POR11432 0014835 //
104871	Office Meeting Back Order Report POR11432 0014835 07/11/2017
104870	Office Meeting Back Order Report POR11444 0014750 //
104658	Office Meeting Back Order Report POR11444 0014750 //
103704	Office Meeting Back Order Report POR11444 0014750 //
104294	Office Meeting Back Order Report POR11451 0014895 //
104055	Office Meeting Back Order Report POR11451 0014895 //
105535	Office Meeting Back Order Report POR11475 0014825 //
84795	Office Meeting VIAMED Stock Meeting Production Statistics (551)
80702	Office Meeting VIAMED Stock Meeting Production Statistics (551)
106710	Office Meeting Sale or Returns Request Scanning Item to Sale or return
106709	Office Meeting Sale or Returns Request Scanning Item to Sale or return
103215	Office Meeting Sale or Returns Request Scanning Item to Sale or return
96392	Office Meeting Sale or Returns Request Scanning Item to Sale or return
84633	Office Meeting Sale or Returns Request Scanning Item to Sale or return
80567	Office Meeting Sale or Returns Request Scanning Item to Sale or return
78892	Office Meeting Sale or Returns Request Scanning Item to Sale or return
78396	Office Meeting Sale or Returns Request Scanning Item to Sale or return
74840	Office Meeting Sale or Returns Request Scanning Item to Sale or return
74399	Office Meeting Sale or Returns Request Scanning Item to Sale or return

Issue ID	Subject
70813	Office Meeting Sale or Returns Request Scanning Item to Sale or return
70151	Office Meeting Sale or Returns Request Scanning Item to Sale or return
70150	Office Meeting Sale or Returns Request Scanning Item to Sale or return
58501	Office Meeting Sale or Returns Returned Items SOR457
70814	Office Meeting Sale or Returns Returned Items SOR688
71961	Office Meeting Sale or Returns Returned Items SOR709
82082	Office Meeting Sale or Returns Returned Items SOR733
78566	Office Meeting Sale or Returns Returned Items SOR734
79896	Office Meeting Sale or Returns Returned Items SOR747
84009	Office Meeting Sale or Returns Returned Items SOR748
87705	Office Meeting Sale or Returns Returned Items SOR758
88507	Office Meeting Sale or Returns Returned Items SOR768
93816	Office Meeting Sale or Returns Returned Items SOR769
92639	Office Meeting Sale or Returns Returned Items SOR778
57855	Office Meeting VIAMED General Issue Sale or return giving
96779	Office Meeting Any other business Sensormedics SpO2 sensors are discontinued
96778	Office Meeting Any other business Sensormedics SpO2 sensors are discontinued
103719	Office Meeting Sale or Returns SMC Northern Ireland exhibition stock - Urgent
87719	Office Meeting Sales Viamed Medical Export Smiths Medical - BCI compatible Paediatric Finger Sensor
87718	Office Meeting Sales Viamed Medical Export Smiths Medical - BCI compatible Paediatric Finger Sensor
98567	Office Meeting Sale or Returns SOR - Gulf & World
92571	Office Meeting Sale or Returns SOR778 - James Paget Hospital sold to AGA Equipment
81486	Office Meeting Cross Reference Database Tables Updates SpO2 Cross Reference
86641	Office Meeting Non Conformance Issues Stock Picked out of s/n order
78584	Office Meeting Stock Queries Stock request
86810	Office Meeting Sales Viamed Medical Export Stock required for a meeting, will be returned as new - Deadline 19th January

Issue ID	Subject
71139	Office Meeting Stock Processing Stock to move, count etc
82072	Office Meeting ISO Issues Symbols on the 4000 Series Packaging
78965	Office Meeting Stock Processing things to do
84462	Office Meeting Marketing Job Logger VM-2160 Leaflet - Remove Microfoam
74837	Office Meeting Marketing Job Logger VM-2160 on SOR to SMC Ltd
101024	Office Meeting Sales Viamed Medical UK VM-2500-S sales lead - Suffolk CH
102931	Office Meeting Stock Processing WYA Consignment Stock (348)
68832	Office Meeting Stock Processing WYA Consignment Stock (348)
87450	Office Meeting Stock Processing WYA Consignment Stock (349)
85438	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85437	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85436	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85435	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85434	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85433	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85432	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85431	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85430	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85429	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85428	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85427	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85426	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
85425	Office Meeting Sales Viamed Medical UK Yorkshire Ambulance - Major changes to consignment stock
102025	Office Meeting Missing Stock or Adjustments Yorkshire Ambulance consignment stock update
102024	Office Meeting Missing Stock or Adjustments Yorkshire Ambulance consignment stock update

Comments on Issues:

FEb 17 86641 Non Conformance Issues Non Conformance Issues Non Conformance Issues  
picked out of order.

Sep 2016 Audits Calander Audit Goods Inwards and Product Identity Audit Goods Inwards and Product  
Identity  
0014774 on the wrong shelf. These have been recently moved and location was not updated.

Mar 2017 BS EN ISO 13485:2003 Document Control Amendments needed

Comments on Risks with Issues:

No end user risks identified

---

VIAMED Stock Meeting - Opera Partnumber / Prices Updates

**0014623 HP Sensor**

Issue ID #81382

Date Created 12/Oct/2016

Issued To: Steve Nixon

Completed Status: 17 Oct 2016

---

Notes:

Added by Catrin Hird sent to Steve Nixon

Do you know if there is a reason why we didn't include 0014623 on the price list, if not I shall update.

---

15 Oct 2016 Steve Nixon

No, but may have been missed as I believe this is one we produce from another sensor.

---

17 Oct 2016 Catrin Hird

Thanks

---

Issues:81382

---

Production - Production

**0014723**

Issue ID #103552

Date Created 22/Sep/2017

Issued To: Steve Nixon

Completed Status: Still Outstanding

---

Notes:

Added by Jonathan Connor sent to Steve Nixon

We have no connectors left to make 0014723. Can we purchase these from BPM or should we get more connectors to produce more.

---

Issues:103552

---

VIAMED Stock Meeting - Purchase Back Orders Review

**0014904 lead times**

Issue ID #69433

Date Created 04/Apr/2016

Issued To: Steve Nixon

Completed Status: 18 Nov 2016

---

Notes:

Added by Steve Hardaker sent to Steve Nixon

I have just been advised that part number 0014904 Disposable SpO2 sensors (Nellcor non-Oximax) is expected to ship from Bluepoint 10th May, whereas the previous estimate was during the week of 4th April.

I'm not sure whether this was a mistake or if they have been put back, but this is for the customer that wanted 25 boxes/month. I am concerned that another month on top may lose the order and future commitment.

Can you see if there is anything you can do to speed this up? Thanks.

---

05 Apr 2016 Steve Nixon

I will look into this today.

Looks like existing sales low:

2015 20 boxes

2016 26 boxes so far.

Are we ordering sufficient stock for the new customer? I assume it's Acorn?

---

05 Apr 2016 Steve Nixon

Updated to leave BPM on the 29th April.

---

05 Apr 2016 Steve Hardaker Thanks. I have added it to the sales predictions and discussed it with JC about maintaining 2-3 months stock, he has ordered the following:

25 - POR10638

50 - POR10655

I'm not sure if there is some way of setting up a kind of standing order reminder for him or if he has a system to log them?

---

05 Apr 2016 Steve Hardaker

Sorry, yes, it is Acorn Distributions.

---

18 Nov 2016 Steve Hardaker We now have stock coming in regularly.

---

Issues:69433

---

VIAMED Stock Meeting - Stock Queries

**0014985-0014988 Superseded?**

Issue ID #77801

Date Created 25/Aug/2016

Issued To: Steve Nixon

Completed Status: 30 Nov 2016

---

Notes:

Added by Catrin Hird sent to Steve Nixon

If these sensors are superseded should they be removed from the price list?

Also the UK price list has 2 different dates on, one at the beginning one at the end.

30 Nov 2016 Catrin HirdRyan informed that 0014985-88 have been discontinued/superseded. So have removed from VM-2160 price list.

Removed from Disposables prices list, Capnograph and VM-2160 price list and added to Z.

Also removed from Pulse Oximetry Accessories leaflet and added to Z.

---

Issues:77801

---

VIAMED Stock Meeting - Purchase Back Orders Review

**20 x 0012161 is now in picking**

Issue ID #92062

Date Created 03/Apr/2017

Issued To: Jonathan Connor

Completed Status: 11 Apr 2017

---

Notes:

Added by Steve Nixon sent to Jonathan Connor

Looks like 20 x 0012161 is now in picking, why was there a delay? Due to not receiving from Bluepoint or do we/did we convert?

---

04 Apr 2017 Jonathan Connor

we are waiting on 0014835 from BPM, customer does not want part shipment, appeared yellow so cathy picked, but didn't have enough of the cables, stock is linked to the order to keep it saved for the customer

---

Issues:92062

---

BS EN ISO 13485:2003 - Document Control Amendments needed  
**4000 Series Labels (Czech Republic)**

Issue ID #91773

Date Created 28/Mar/2017

Issued To: Steve Nixon

Completed Status: Still Outstanding

Linked Files

25 Apr 2017 Updated notes from Ralf Detweiler

25 Apr 2017 SF4011VM\_Label\_Box\_Top

25 Apr 2017 Humidity symbol

25 Apr 2017 Temperature range symbol

28 Apr 2017 ASQA 4000 Series Stock

---

**Notes:**

Added by Ryan Swaine sent to Steve Nixon  
Email from Eva @ ASQA, dated 27/03/2017

Hi Ryan

I am coming back to my below e-mail: we have received the official Inspection protocol from SUKL.  
As I mentioned below, they picked for their inspection your product, which was sitting in our warehouse since June 2010... 4000 Series Finger Sensor, item No 0014605.

the protocol is listing following problems:

1. missing name of manufacturer on product packaging
2. humidity sign on product packaging is not according to harmonised standards  
Meaning of non-harmonised sign was not explained in IFU.
3. "temperature underneath the roof" sign on product packaging is not according to harmonised standards  
Meaning of non-harmonised sign was not explained in IFU.
4. Instructions for use are showing different Environmental Requirements than product packaging:  
operating temperature: 10-40C  
storage and transportation: 0-50C  
relative humidity: 0-75% non-condensing  
Attached is the product packaging picture - SUKL is correct, limits are not the same..

I need to send our comment to SUKL within 30 days. Will you please let me know if you had any reply from your MD - as in your below e-mail?

The best solution how to solve the situation quickly might be to put a new sticker on the product packaging, which will include Viamed Ltd. name in the manufacturer's address and will have harmonised signs for temperature and humidity, showing the same limits as listed in IFU.

Then I would inform SUKL that the product packaging was changed according to Medical Devices Law requirements and I would send them picture of the new product packaging.

Please let me know your recommendation as soon as possible.

Many thanks and best regards

Eva

---

25 Apr 2017 Steve Nixon

Ryan, I have the updated label artwork from Ralf. I still need for him to change the upper humidity value from 95% to 90%. I will ask him to reprint the labels so that we have exactly the same labels, do you know the serial numbers of the sensors that we need to update? Or will just a copy of the updated artwork suffice?

---

25 Apr 2017 Steve Nixon

Updated artwork and notes attached. Still need the edit as stated above.

---

28 Apr 2017 Ryan Swaine ASQA require the new labels for their stock. I have attached an Excel document with the part numbers and serial number of their stock, please can we print the labels and post them to them?

---

Issues:91773

---

VIAMED Sales And Marketing - Price Lists UK  
**4000 series sensors to remove from price lists**  
Issue ID #96777  
Date Created 15/Jun/2017  
Issued To: Catrin Hird  
Completed Status: 04 Jul 2017

---

Notes:

Added by Steve Hardaker sent to Catrin Hird

The following Sensormedics compatible 4000 series sensors are all showing as yellow in the Intrastats stock pages, we have never had stock and SN advises that they are no longer viable, so can be considered discontinued.

0014620  
0014720  
0014780  
0014843

I have removed them from the GHX master stock group, please can you remove them from the relevant 4000 series price lists and have them updated in L drive thanks.

---

03 Jul 2017 Catrin Hird

Duplicate issue, linked issue from SN.

---

04 Jul 2017 Steve Hardaker

---

Issues:96777

---

VIAMED Sales And Marketing - Sales

**Account 00012906 Stop Status Changed to Stopped**

Issue ID #97766

Date Created 29/Jun/2017

Issued To: Helen Lamb

Completed Status: 30 Jun 2017

---

**Notes:**

Added by Steve Hardaker sent to Helen Lamb

I think you need to chase this one up as a priority, I'm not even sure how he ended up on a credit account, did he ever complete and return a credit account application form? I sent it many times but he was always happy to pay on his VISA card against a proforma.

Just some background info:

He is no longer buying from us as he was selling products on into Saudi Arabia and Ryan's distributor Jamjoom ended up going for the same orders, so we had to stop supplying to Acorn.

It looks like the outstanding amount is from his last order, IN149728, which was for 42 boxes of 0014904. It was originally 84 boxes but we didn't have the stock and ended up agreeing to cancel the outstanding 42 boxes due to the distribution rights issues detailed above. See issue 87575 for details.

He's not the best at answering emails so I suggest you try calling and he will probably just pay by card, phone numbers are in Goldmine and CRM.

---

30 Jun 2017 Helen Lamb

they did fill a form in and have been on credit account since last year. Will keep an eye on it.

---

30 Jun 2017 Steve Hardaker

---

Issues:97766

---

Sales - Sales

**Account 00012906 Stop Status Changed to Stopped**

Issue ID #97682

Date Created 29/Jun/2017

Issued To: Steve Hardaker

Completed Status: 29 Jun 2017

---

**Notes:**

Account 00012906 Acorn Distributions Stop Status Changed to Stopped

---

29 Jun 2017 Steve Hardaker

Created Related Issue #97766

Added by Steve Hardaker sent to Helen Lamb

I think you need to chase this one up as a priority, I'm not even sure how he ended up on a credit account, did he ever complete and return a credit account application form? I sent it many times but he was always happy to pay on his VISA card against a proforma.

Just some background info:

He is no longer buying from us as he was selling products on into Saudi Arabia and Ryan's distributor Jamjoom ended up going for the same orders, so we had to stop supplying to Acorn.

It looks like the outstanding amount is from his last order, IN149728, which was for 42 boxes of 0014904. It was originally 84 boxes but we didn't have the stock and ended up agreeing to cancel the outstanding 42 boxes due to the distribution rights issues detailed above. See issue 87575 for details.

He's not the best at answering emails so I suggest you try calling and he will probably just pay by card, phone numbers are in Goldmine and CRM.

---

Issues:97682

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #101279

Date Created 22/Aug/2017

Issued To: Jonathan Connor

Completed Status: 31 Aug 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

29 Aug 2017 Jonathan Connor

Not required at the moment

---

Issues:101279

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #99243

Date Created 24/Jul/2017

Issued To: Jonathan Connor

Completed Status: 31 Aug 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

29 Aug 2017 Jonathan Connor

Not required at the moment

---

Issues:99243

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #97191

Date Created 22/Jun/2017

Issued To: Jonathan Connor

Completed Status: 06 Jul 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

29 Jun 2017 Jonathan Connor

Not required at the moment

---

Issues:97191

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acorn 0014904 (522)**

Issue ID #95190

Date Created 22/May/2017

Issued To: Jonathan Connor

Completed Status: 01 Jun 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

30 May 2017 Jonathan Connor

Not required

---

Issues:95190

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #93293

Date Created 24/Apr/2017

Issued To: Jonathan Connor

Completed Status: 28 Apr 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

27 Apr 2017 Jonathan Connor

Not required, to discuss with RS when Jamjoom order

---

Issues:93293

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #91397

Date Created 22/Mar/2017

Issued To: Jonathan Connor

Completed Status: 23 Mar 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

22 Mar 2017 Jonathan Connor  
is 87575 open

---

Issues:91397

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acorn 0014904 (522)**

Issue ID #89766

Date Created 22/Feb/2017

Issued To: Jonathan Connor

Completed Status: 02 Mar 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

23 Feb 2017 Jonathan Connor

Other issue open, sent to SH to discuss with the customer

---

Issues:89766

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acorn 0014904 (522)**

Issue ID #87575

Date Created 23/Jan/2017

Issued To: Jonathan Connor

Completed Status: 30 Mar 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

14 Feb 2017 Jonathan Connor

Next Action Changed From Jonathan Connor To Steve Hardaker any news on acorn?

---

15 Feb 2017 Steve Hardaker

Emailed him yesterday, will update when he replies or will call later if he doesn't.

---

09 Mar 2017 Steve Hardaker

Next Action Changed From Steve Hardaker To Jonathan Connor This has been up in the air as it turns out that Acorn's distributor was going up against Jamjoom in a Saudi private hospital.

I discussed it with SN and Ryan and they agreed that if Acorn gets this order and Jamjoom dont contact us to make an objection, this order can go through.

After that though, Ryan will be trying to secure the business via Jamjoom so this may be that last of the orders for Acorn if he succeeds.

You might then need to stock these for Jamjoom, or they may try to sell the 0014912 plaster tape versions instead, you will have to see what develops.

In the short-term, Acorn has today placed an order for 84 boxes of 0014904, we have 49 showing in stock and I have agreed to part ship 42.

You will need at least enough to cover this order and whatever demand we have excluding Acorn (Nosbet bought 20 boxes last year).

---

14 Mar 2017 Steve Hardaker

Elan has just rung to say to cancel his back order for 42 boxes of 0014904 on back order (ORD85297). He has received 42 boxes already against this order, I get the impression that he is concerned that if Jamjoom make a fuss with the hospital or his Saudi contact, he could get stuck with these.

We will still need some stock but are now at the stage where it depends whether Jamjoom want these or will stick with 0014912, so you are best speaking to Ryan.

---

23 Mar 2017 Jonathan Connor

discussed with rs

---

Issues:87575

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acorn 0014904 (522)**

Issue ID #85878

Date Created 22/Dec/2016

Issued To: Jonathan Connor

Completed Status: 18 Jan 2017

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

12 Jan 2017 Jonathan Connor  
stock ok

---

Issues:85878

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acorn 0014904 (522)**

Issue ID #83976

Date Created 22/Nov/2016

Issued To: Jonathan Connor

Completed Status: 19 Dec 2016

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

08 Dec 2016 Jonathan Connor

ok for stock

---

Issues:83976

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acorn 0014904 (522)**

Issue ID #82132

Date Created 24/Oct/2016

Issued To: Jonathan Connor

Completed Status: 27 Oct 2016

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

25 Oct 2016 Jonathan Connor  
stock ok

---

Issues:82132

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #79737

Date Created 22/Sep/2016

Issued To: Jonathan Connor

Completed Status: 05 Oct 2016

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

04 Oct 2016 Jonathan Connor

On order

---

Issues:79737

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #77317

Date Created 22/Aug/2016

Issued To: Jonathan Connor

Completed Status: 03 Sep 2016

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

30 Aug 2016 Jonathan Connor  
Complete

---

Issues:77317

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #75474

Date Created 22/Jul/2016

Issued To: Jonathan Connor

Completed Status: 30 Jul 2016

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

27 Jul 2016 Jonathan Connor

Done

---

Issues:75474

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acom 0014904 (522)**

Issue ID #73408

Date Created 22/Jun/2016

Issued To: Jonathan Connor

Completed Status: 14 Jul 2016

---

Notes:

System Generated

Task To be Completed

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

11 Jul 2016 Jonathan Connor  
enough stock

---

Issues:73408

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acorn Distribution ac 12906 Stock levels for part number 0014904 (522)**

Issue ID #106219

Date Created 23/Oct/2017

Issued To: Jonathan Connor

Completed Status: Still Outstanding

---

Notes:

System Generated

Task To be Completed

Acorn Distribution ac 12906

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

Issues:106219

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Acorn Distribution ac 12906 Stock levels for part number 0014904 (522)**

Issue ID #103496

Date Created 22/Sep/2017

Issued To: Jonathan Connor

Completed Status: 22 Sep 2017

---

Notes:

System Generated

Task To be Completed

Acorn Distribution ac 12906

Check stock 0014904. 25 per month for acorn required. See if ordered and place new PO

---

22 Sep 2017 Jonathan Connor

Not required at the moment

---

Issues:103496

---

Audits Calander - Audit Goods Inwards and Product Identity  
**Audit 09 Product Identity - Stock in a incorrect location.**

Issue ID #78168

Date Created 01/Sep/2016

Issued To: Jonathan Connor

Completed Status: 02 Sep 2016

---

Notes:

Copy Issue Main Issue #78166

Added by Helen Lamb sent to Robert Connor

0110017 on the wrong shelf - I know these are in the process of being changed. But please try to keep the location up to date as you go along.

0014774 on the wrong shelf. These have been recently moved and location was not updated.

These have now both been corrected.

Please make sure that you double check you have updated the locations, as soon as possible, as you work and move things around.

---

01 Sep 2016 Jonathan Connor

Thanks. No problem

---

Issues:78168

---

Audits Calander - Audit Goods Inwards and Product Identity  
**Audit 09 Product Identity - Stock in a incorrect location.**

Issue ID #78167

Date Created 01/Sep/2016

Issued To: Callum Watson

Completed Status: 08 Sep 2016

---

Notes:

Copy Issue Main Issue #78166

Added by Helen Lamb sent to Robert Connor

0110017 on the wrong shelf - I know these are in the process of being changed. But please try to keep the location up to date as you go along.

0014774 on the wrong shelf. These have been recently moved and location was not updated.

These have now both been corrected.

Please make sure that you double check you have updated the locations, as soon as possible, as you work and move things around.

---

02 Sep 2016 Callum Watson

---

Issues:78167

---

Audits Calander - Audit Goods Inwards and Product Identity  
**Audit 09 Product Identity - Stock in a incorrect location.**

Issue ID #78166

Date Created 01/Sep/2016

Issued To: Robert Connor

Completed Status: 02 Sep 2016

---

Notes:

Added by Helen Lamb sent to Robert Connor

0110017 on the wrong shelf - I know these are in the process of being changed. But please try to keep the location up to date as you go along.

0014774 on the wrong shelf. These have been recently moved and location was not updated.

These have now both been corrected.

Please make sure that you double check you have updated the locations, as soon as possible, as you work and move things around.

---

01 Sep 2016 Robert Connor

Noted, stock is now scanned to the correct shelf.

---

Issues:78166

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00000333 0014996 ORD82901**

Issue ID #81887

Date Created 19/Oct/2016

Issued To: General

Completed Status: 20 Oct 2016

Stock Ref: 0014996

Opera A/c: 00000333

Order Number: ORD82901

---

Notes:

19 Oct 2016 07:41:56 AM Catherine Spence  
Shipping Wed 19 Oct 2016

19 Oct 2016 08:10:14 AM Catherine Spence  
See Attached Notes  
to re-label

---

Issues:81887

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00000333 0014996 ORD83324**

Issue ID #83349

Date Created 11/Nov/2016

Issued To: General

Completed Status: 15 Dec 2016

Stock Ref: 0014996

Opera A/c: 00000333

Order Number: ORD83324

---

Notes:

11 Nov 2016 08:08:23 AM Catherine Spence  
Back Order - Supplier

14 Nov 2016 12:17:59 PM Katie Evans  
Advised of 4-6 week lead time and offered infant version as per SJH

---

Issues:83349

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00001038 0014750 ORD83404**

Issue ID #83689

Date Created 16/Nov/2016

Issued To: General

Completed Status: 06 Dec 2016

Stock Ref: 0014750

Opera A/c: 00001038

Order Number: ORD83404

---

Notes:

16 Nov 2016 12:58:09 PM Katie Evans

See Attached Notes

Moved to priority 8 as I suspect it is no longer a BMI hospital (Website says Phoenix hospitals and BMI have no record of it) EH emailing to clarify invoice address and to advise of lead time

28 Nov 2016 10:14:35 AM Emily Hanson

chased

28 Nov 2016 02:24:09 PM Emily Hanson

okay to go- aware of lead time

---

Issues:83689

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00001661 0014606 ORD84957**

Issue ID #89430

Date Created 16/Feb/2017

Issued To: General

Completed Status: 17 Feb 2017

Stock Ref: 0014606

Opera A/c: 00001661

Order Number: ORD84957

---

Notes:

16 Feb 2017 09:29:10 AM Catherine Spence  
Shipping Thu 16 Feb 2017

---

Issues:89430

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00001810 0014750 ORD88040**

Issue ID #101669

Date Created 29/Aug/2017

Issued To: General

Completed Status: 30 Aug 2017

Stock Ref: 0014750

Opera A/c: 00001810

Order Number: ORD88040

---

Notes:

29 Aug 2017 10:02:16 AM Emily Hanson  
okay to fo

---

Issues:101669

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00002948 0014618 ORD86919**

Issue ID #97131

Date Created 21/Jun/2017

Issued To: General

Completed Status: 21 Jun 2017

Stock Ref: 0014618

Opera A/c: 00002948

Order Number: ORD86919

---

Notes:

21 Jun 2017 07:59:27 AM Catherine Spence  
Shipping Wed 21 Jun 2017

---

Issues:97131

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00003007 0014750 ORD88668**

Issue ID #106018

Date Created 19/Oct/2017

Issued To: General

Completed Status: 19 Oct 2017

Stock Ref: 0014750

Opera A/c: 00003007

Order Number: ORD88668

---

Notes:

19 Oct 2017 07:56:34 AM Jonathan Connor

See Attached Notes

Not a duplicate order, customer wants one new unit and one new unit to solve repair

---

Issues:106018

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00003007 0014750 ORD88893**

Issue ID #106019

Date Created 19/Oct/2017

Issued To: General

Completed Status: 19 Oct 2017

Stock Ref: 0014750

Opera A/c: 00003007

Order Number: ORD88893

---

Notes:

19 Oct 2017 07:56:48 AM Jonathan Connor

See Attached Notes

Not a duplicate order, customer wants one new unit and one new unit to solve repair

---

Issues:106019

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00003007 0014756 ORD80491**

Issue ID #72195

Date Created 31/May/2016

Issued To: General

Completed Status: 20 Jun 2016

Stock Ref: 0014756

Opera A/c: 00003007

Order Number: ORD80491

---

Notes:

31 May 2016 Catherine Spence  
Back Order - Supplier

03 Jun 2016 Sarah Hawker  
customer aware of lead time sarahh 27/05/2016

---

Issues:72195

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00003007 0014895 ORD80491**

Issue ID #72768

Date Created 10/Jun/2016

Issued To: General

Completed Status: 20 Jun 2016

Stock Ref: 0014895

Opera A/c: 00003007

Order Number: ORD80491

---

Notes:

10 Jun 2016 Catherine Spence  
Shipping Fri 10 Jun 2016

---

Issues:72768

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00003310 0014989 ORD84510**

Issue ID #87590

Date Created 23/Jan/2017

Issued To: General

Completed Status: 23 Jan 2017

Stock Ref: 0014989

Opera A/c: 00003310

Order Number: ORD84510

---

Notes:

23 Jan 2017 07:46:20 AM Catherine Spence  
Shipping Mon 23 Jan 2017

---

Issues:87590

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00003310 0014989 ORD86382**

Issue ID #95078

Date Created 18/May/2017

Issued To: General

Completed Status: 19 May 2017

Stock Ref: 0014989

Opera A/c: 00003310

Order Number: ORD86382

---

Notes:

18 May 2017 12:54:47 PM Catherine Spence  
See Attached Notes  
to re-work stock

---

Issues:95078

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00004640 0014650 ORD87652**

Issue ID #99964

Date Created 03/Aug/2017

Issued To: General

Completed Status: 03 Aug 2017

Stock Ref: 0014650

Opera A/c: 00004640

Order Number: ORD87652

---

Notes:

03 Aug 2017 07:43:52 AM Catherine Spence  
Shipping Thu 03 Aug 2017

---

Issues:99964

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00004729 0014994 ORD85209**

Issue ID #90481

Date Created 06/Mar/2017

Issued To: General

Completed Status: 06 Mar 2017

Stock Ref: 0014994

Opera A/c: 00004729

Order Number: ORD85209

---

Notes:

06 Mar 2017 07:52:37 AM Catherine Spence  
Shipping Mon 06 Mar 2017

06 Mar 2017 07:54:31 AM Catherine Spence

---

Issues:90481

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00005030 0014608 ORD80644**

Issue ID #72511

Date Created 06/Jun/2016

Issued To: General

Completed Status: 06 Jun 2016

Stock Ref: 0014608

Opera A/c: 00005030

Order Number: ORD80644

---

Notes:

06 Jun 2016 Catherine Spence  
Shipping Mon 06 Jun 2016

---

Issues:72511

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00005030 0014608 ORD80961**

Issue ID #73474

Date Created 23/Jun/2016

Issued To: General

Completed Status: 23 Jun 2016

Stock Ref: 0014608

Opera A/c: 00005030

Order Number: ORD80961

---

Notes:

23 Jun 2016 Catherine Spence  
Shipping Thu 23 Jun 2016

---

Issues:73474

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00005110 0014802 ORD82558**

Issue ID #80265

Date Created 29/Sep/2016

Issued To: General

Completed Status: 30 Sep 2016

Stock Ref: 0014802

Opera A/c: 00005110

Order Number: ORD82558

---

Notes:

29 Sep 2016 Catherine Spence  
Shipping Thu 29 Sep 2016

---

Issues:80265

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00005430 0014612 ORD81389**

Issue ID #75293

Date Created 19/Jul/2016

Issued To: General

Completed Status: 19 Jul 2016

Stock Ref: 0014612

Opera A/c: 00005430

Order Number: ORD81389

---

Notes:

19 Jul 2016 Catherine Spence  
Shipping Tue 19 Jul 2016

---

Issues:75293

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00005430 0014612 ORD88878**

Issue ID #105696

Date Created 18/Oct/2017

Issued To: General

Completed Status: 18 Oct 2017

Stock Ref: 0014612

Opera A/c: 00005430

Order Number: ORD88878

---

Notes:

18 Oct 2017 08:28:35 AM Catherine Spence  
Shipping Wed 18 Oct 2017

---

Issues:105696

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014601 ORD83265**

Issue ID #83115

Date Created 09/Nov/2016

Issued To: General

Completed Status: 09 Nov 2016

Stock Ref: 0014601

Opera A/c: 00006277

Order Number: ORD83265

---

Notes:

09 Nov 2016 08:04:30 AM Catherine Spence  
Shipping Wed 09 Nov 2016

---

Issues:83115

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014603 ORD86790**

Issue ID #96581

Date Created 13/Jun/2017

Issued To: General

Completed Status: 13 Jun 2017

Stock Ref: 0014603

Opera A/c: 00006277

Order Number: ORD86790

---

Notes:

13 Jun 2017 07:50:42 AM Catherine Spence  
Shipping Tue 13 Jun 2017

---

Issues:96581

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014604 ORD86608**

Issue ID #95901

Date Created 02/Jun/2017

Issued To: General

Completed Status: 02 Jun 2017

Stock Ref: 0014604

Opera A/c: 00006277

Order Number: ORD86608

---

Notes:

02 Jun 2017 07:59:15 AM Catherine Spence  
Shipping Fri 02 Jun 2017

---

Issues:95901

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014904 ORD83523**

Issue ID #84147

Date Created 24/Nov/2016

Issued To: General

Completed Status: 24 Nov 2016

Stock Ref: 0014904

Opera A/c: 00006277

Order Number: ORD83523

---

Notes:

24 Nov 2016 07:50:36 AM Catherine Spence  
Back Order - Supplier  
part shipment done as requested

24 Nov 2016 01:23:48 PM Katie Evans  
The order for the remaining 3 boxes has been cancelled. Please do not ship

---

Issues:84147

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014905 ORD80592**

Issue ID #72353

Date Created 02/Jun/2016

Issued To: General

Completed Status: 02 Jun 2016

Stock Ref: 0014905

Opera A/c: 00006277

Order Number: ORD80592

---

Notes:

02 Jun 2016 Catherine Spence  
Shipping Thu 02 Jun 2016

---

Issues:72353

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014905 ORD81123**

Issue ID #74166

Date Created 01/Jul/2016

Issued To: Catherine Spence

Completed Status: 01 Jul 2016

Stock Ref: 0014905

Opera A/c: 00006277

Order Number: ORD81123

---

Notes:

01 Jul 2016 Sarah Walton

See Attached Notes

Customer would like part shipment

---

Issues:74166

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014909 ORD83890**

Issue ID #85271

Date Created 13/Dec/2016

Issued To: General

Completed Status: 13 Dec 2016

Stock Ref: 0014909

Opera A/c: 00006277

Order Number: ORD83890

---

Notes:

13 Dec 2016 07:30:43 AM Catherine Spence  
Shipping Tue 13 Dec 2016

---

Issues:85271

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014909 ORD85235**

Issue ID #90564

Date Created 07/Mar/2017

Issued To: General

Completed Status: 07 Mar 2017

Stock Ref: 0014909

Opera A/c: 00006277

Order Number: ORD85235

---

Notes:

07 Mar 2017 09:17:57 AM Catherine Spence  
Shipping Tue 07 Mar 2017

---

Issues:90564

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006277 0014909 ORD85503**

Issue ID #91490

Date Created 23/Mar/2017

Issued To: General

Completed Status: 23 Mar 2017

Stock Ref: 0014909

Opera A/c: 00006277

Order Number: ORD85503

---

Notes:

23 Mar 2017 07:42:50 AM Catherine Spence  
Shipping Thu 23 Mar 2017

---

Issues:91490

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006297 0014590 ORD79055**

Issue ID #68257

Date Created 04/Mar/2016

Issued To: General

Completed Status: 26 Apr 2016

Stock Ref: 0014590

Opera A/c: 00006297

Order Number: ORD79055

---

Notes:

04 Mar 2016 Catherine Spence  
Back Order - Supplier

08 Mar 2016 Katie Evans  
See Attached Notes  
Customer aware of the delay, offered part shipment

08 Mar 2016 Katie Evans  
Please part ship remainder of order. Customer happy to wait for this product only

17 Mar 2016 Catherine Spence  
part shipment has been done last week

17 Mar 2016 Catherine Spence  
Back Order - Supplier

---

Issues:68257

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006343 0014750 ORD84507**

Issue ID #87589

Date Created 23/Jan/2017

Issued To: General

Completed Status: 23 Jan 2017

Stock Ref: 0014750

Opera A/c: 00006343

Order Number: ORD84507

---

Notes:

23 Jan 2017 07:46:05 AM Catherine Spence  
Shipping Mon 23 Jan 2017

---

Issues:87589

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006343 0014751 ORD83971**

Issue ID #85574

Date Created 16/Dec/2016

Issued To: General

Completed Status: 19 Dec 2016

Stock Ref: 0014751

Opera A/c: 00006343

Order Number: ORD83971

---

Notes:

16 Dec 2016 07:50:58 AM Catherine Spence  
Shipping Fri 16 Dec 2016

19 Dec 2016 07:34:28 AM Catherine Spence  
Shipping Mon 19 Dec 2016

---

Issues:85574

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006343 0014995 ORD88326**

Issue ID #102913

Date Created 14/Sep/2017

Issued To: General

Completed Status: 15 Sep 2017

Stock Ref: 0014995

Opera A/c: 00006343

Order Number: ORD88326

---

Notes:

14 Sep 2017 12:19:45 PM Katie Evans  
See Attached Notes  
Sent OC

14 Sep 2017 12:19:52 PM Katie Evans  
Happy to proceed

15 Sep 2017 08:15:17 AM Catherine Spence  
Shipping Fri 15 Sep 2017

---

Issues:102913

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006350 0014993 ORD85546**

Issue ID #91697

Date Created 27/Mar/2017

Issued To: General

Completed Status: 28 Mar 2017

Stock Ref: 0014993

Opera A/c: 00006350

Order Number: ORD85546

---

Notes:

27 Mar 2017 08:30:56 AM Catherine Spence  
Shipping Mon 27 Mar 2017

---

Issues:91697

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006350 0014993 ORD86758**

Issue ID #96465

Date Created 12/Jun/2017

Issued To: General

Completed Status: 12 Jun 2017

Stock Ref: 0014993

Opera A/c: 00006350

Order Number: ORD86758

---

Notes:

12 Jun 2017 08:03:48 AM Catherine Spence  
Shipping Mon 12 Jun 2017

---

Issues:96465

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006350 0014993 ORD87237**

Issue ID #98525

Date Created 11/Jul/2017

Issued To: General

Completed Status: 11 Jul 2017

Stock Ref: 0014993

Opera A/c: 00006350

Order Number: ORD87237

---

Notes:

11 Jul 2017 07:43:42 AM Catherine Spence  
Shipping Tue 11 Jul 2017

---

Issues:98525

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006439 0014750 ORD87572**

Issue ID #99649

Date Created 28/Jul/2017

Issued To: General

Completed Status: 20 Sep 2017

Stock Ref: 0014750

Opera A/c: 00006439

Order Number: ORD87572

---

Notes:

28 Jul 2017 11:46:39 AM Katie Evans

See Attached Notes

Please part ship sidestreams ASAP

03 Aug 2017 07:30:26 AM Catherine Spence

part shipment done

29 Aug 2017 08:38:59 AM Catherine Spence

Back Order - Supplier

20 Sep 2017 07:48:56 AM Catherine Spence

Shipping Wed 20 Sep 2017

---

Issues:99649

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006441 0014750 ORD87228**

Issue ID #98536

Date Created 11/Jul/2017

Issued To: General

Completed Status: 12 Jul 2017

Stock Ref: 0014750

Opera A/c: 00006441

Order Number: ORD87228

---

Notes:

11 Jul 2017 08:20:51 AM Katie Evans

See Attached Notes

Waiting to see how he'd like it shipping

11 Jul 2017 11:58:57 AM Katie Evans

Awaiting payment for PRO04513 for carriage

11 Jul 2017 04:54:33 PM Katie Evans

HL happy for this to ship before carriage payment is made, see PRO04513 paperwork for signature

12 Jul 2017 07:56:51 AM Catherine Spence

Shipping Wed 12 Jul 2017

---

Issues:98536

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00006673 0019002 ORD84743**

Issue ID #88519

Date Created 06/Feb/2017

Issued To: General

Completed Status: 07 Feb 2017

Stock Ref: 0019002

Opera A/c: 00006673

Order Number: ORD84743

---

Notes:

06 Feb 2017 02:07:04 PM Emma Clark

See Attached Notes

SHIP WITH OTHER ORDERS

---

Issues:88519

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007125 0014750 ORD80383**

Issue ID #71878

Date Created 23/May/2016

Issued To: General

Completed Status: 24 May 2016

Stock Ref: 0014750

Opera A/c: 00007125

Order Number: ORD80383

---

Notes:

23 May 2016 Katie Evans

See Attached Notes

DO NOT SHIP. Customer wants to put order on a new € account

---

Issues:71878

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007224 0014750 ORD86345**

Issue ID #94943

Date Created 17/May/2017

Issued To: General

Completed Status: 17 May 2017

Stock Ref: 0014750

Opera A/c: 00007224

Order Number: ORD86345

---

Notes:

17 May 2017 07:32:26 AM Catherine Spence  
Shipping Wed 17 May 2017

---

Issues:94943

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007246 0012101 ORD88417**

Issue ID #103379

Date Created 20/Sep/2017

Issued To: General

Completed Status: Still Outstanding

Stock Ref: 0012101

Opera A/c: 00007246

Order Number: ORD88417

---

Notes:

20 Sep 2017 03:33:04 PM Katie Evans

See Attached Notes

Please part ship 311 x 0012101 the VM2160`s and the 0014825`s ASAP

---

Issues:103379

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007246 0014750 ORD87220**

Issue ID #98468

Date Created 10/Jul/2017

Issued To: General

Completed Status: 21 Jul 2017

Stock Ref: 0014750

Opera A/c: 00007246

Order Number: ORD87220

---

Notes:

10 Jul 2017 12:28:36 PM Katie Evans

See Attached Notes

Ship now or with SRS?

17 Jul 2017 11:22:54 AM Katie Evans

Ongoing, customer uncertain at present

18 Jul 2017 11:50:40 AM Katie Evans

Please ship with ORD87372 and outstanding repair

---

Issues:98468

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007246 0014750 ORD87372**

Issue ID #98968

Date Created 18/Jul/2017

Issued To: General

Completed Status: 21 Jul 2017

Stock Ref: 0014750

Opera A/c: 00007246

Order Number: ORD87372

---

Notes:

18 Jul 2017 02:04:31 PM Katie Evans

See Attached Notes

Please ship with ORD87220 and outstanding repair

---

Issues:98968

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007246 0014751 ORD81447**

Issue ID #75485

Date Created 22/Jul/2016

Issued To: General

Completed Status: 22 Jul 2016

Stock Ref: 0014751

Opera A/c: 00007246

Order Number: ORD81447

---

Notes:

22 Jul 2016 Catherine Spence  
Shipping Fri 22 Jul 2016

---

Issues:75485

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007246 0014825 ORD80658**

Issue ID #72568

Date Created 07/Jun/2016

Issued To: General

Completed Status: 09 Jun 2016

Stock Ref: 0014825

Opera A/c: 00007246

Order Number: ORD80658

---

Notes:

07 Jun 2016 Catherine Spence  
Back Order - Supplier

---

Issues:72568

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014801 ORD83948**

Issue ID #85489

Date Created 15/Dec/2016

Issued To: General

Completed Status: 15 Dec 2016

Stock Ref: 0014801

Opera A/c: 00007356

Order Number: ORD83948

---

Notes:

15 Dec 2016 08:18:29 AM Catherine Spence  
Shipping Thu 15 Dec 2016

---

Issues:85489

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014903 ORD80179**

Issue ID #71170

Date Created 10/May/2016

Issued To: Catherine Spence

Completed Status: 17 Jun 2016

Stock Ref: 0014903

Opera A/c: 00007356

Order Number: ORD80179

---

Notes:

10 May 2016 Catherine Spence  
Back Order - Supplier  
not able to use stock

11 May 2016 Catherine Spence  
asked office to contact customer to see if they will accept a part shipment

11 May 2016 Emily Hanson  
emailed customer

11 May 2016 Emily Hanson  
please part ship.

12 May 2016 Catherine Spence  
Shipping Thu 12 May 2016

13 May 2016 Catherine Spence  
Back Order - Supplier  
part shipment done

31 May 2016 Emily Hanson  
do not ship yet, they are adding to the order.

---

Issues:71170

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014903 ORD81267**

Issue ID #74793

Date Created 12/Jul/2016

Issued To: General

Completed Status: 26 Jul 2016

Stock Ref: 0014903

Opera A/c: 00007356

Order Number: ORD81267

---

Notes:

12 Jul 2016 Catherine Spence  
See Attached Notes  
need to be re-labeled

13 Jul 2016 Catherine Spence  
with jc

25 Jul 2016 Katie Evans  
with CH

---

Issues:74793

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014903 ORD82225**

Issue ID #78610

Date Created 07/Sep/2016

Issued To: General

Completed Status: 15 Sep 2016

Stock Ref: 0014903

Opera A/c: 00007356

Order Number: ORD82225

---

Notes:

07 Sep 2016 Catherine Spence  
See Attached Notes  
ship with other order

14 Sep 2016 Catherine Spence  
Shipping Wed 14 Sep 2016

---

Issues:78610

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014905 ORD80179**

Issue ID #71171

Date Created 10/May/2016

Issued To: General

Completed Status: 12 May 2016

Stock Ref: 0014905

Opera A/c: 00007356

Order Number: ORD80179

---

Notes:

10 May 2016 Catherine Spence  
See Attached Notes  
to relabel

---

Issues:71171

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014905 ORD81371**

Issue ID #75217

Date Created 18/Jul/2016

Issued To: General

Completed Status: 15 Sep 2016

Stock Ref: 0014905

Opera A/c: 00007356

Order Number: ORD81371

---

Notes:

18 Jul 2016 Catherine Spence  
Shipping Mon 18 Jul 2016  
part ship

19 Jul 2016 Catherine Spence  
Back Order - Supplier  
part shipment done

06 Sep 2016 Sarah Walton  
See Attached Notes  
Customer aware that we aim to ship the rest of their order w/c 12th Sept

07 Sep 2016 Catherine Spence  
Back Order - Supplier

07 Sep 2016 Emily Hanson  
dont ship yet, adding to the order.

07 Sep 2016 Emily Hanson  
ship with ORD82225

14 Sep 2016 Catherine Spence  
Shipping Wed 14 Sep 2016

14 Sep 2016 Catherine Spence  
Testing/QA  
to label and qa

---

Issues:75217

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014905 ORD85040**

Issue ID #89856

Date Created 23/Feb/2017

Issued To: General

Completed Status: 24 Apr 2017

Stock Ref: 0014905

Opera A/c: 00007356

Order Number: ORD85040

---

Notes:

23 Feb 2017 07:58:47 AM Catherine Spence  
See Attached Notes  
part ship

23 Feb 2017 02:20:21 PM Catherine Spence  
Back Order - Supplier  
Part shipment done

---

Issues:89856

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014905 ORD85482**

Issue ID #91578

Date Created 24/Mar/2017

Issued To: General

Completed Status: 25 Apr 2017

Stock Ref: 0014905

Opera A/c: 00007356

Order Number: ORD85482

---

Notes:

24 Mar 2017 07:32:53 AM Catherine Spence  
Back Order - Supplier

---

Issues:91578

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014905 ORD86613**

Issue ID #96082

Date Created 06/Jun/2017

Issued To: General

Completed Status: 15 Jun 2017

Stock Ref: 0014905

Opera A/c: 00007356

Order Number: ORD86613

---

Notes:

06 Jun 2017 09:13:08 AM Katie Evans  
Please ship both orders and repair together

---

Issues:96082

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014905 ORD87330**

Issue ID #98800

Date Created 14/Jul/2017

Issued To: General

Completed Status: 20 Jul 2017

Stock Ref: 0014905

Opera A/c: 00007356

Order Number: ORD87330

---

Notes:

14 Jul 2017 11:36:50 AM Catherine Spence  
See Attached Notes  
shipping monday, need new barcodes

19 Jul 2017 07:55:40 AM Catherine Spence  
still waiting on new labels from jc

---

Issues:98800

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014905 ORD87734**

Issue ID #100361

Date Created 09/Aug/2017

Issued To: General

Completed Status: 09 Aug 2017

Stock Ref: 0014905

Opera A/c: 00007356

Order Number: ORD87734

---

Notes:

09 Aug 2017 07:39:31 AM Catherine Spence  
See Attached Notes  
to re label

---

Issues:100361

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014905 ORD88382**

Issue ID #104640

Date Created 05/Oct/2017

Issued To: General

Completed Status: 05 Oct 2017

Stock Ref: 0014905

Opera A/c: 00007356

Order Number: ORD88382

---

Notes:

05 Oct 2017 07:44:34 AM Catherine Spence  
Shipping Thu 05 Oct 2017

---

Issues:104640

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014911 ORD86613**

Issue ID #95915

Date Created 02/Jun/2017

Issued To: General

Completed Status: 15 Jun 2017

Stock Ref: 0014911

Opera A/c: 00007356

Order Number: ORD86613

---

Notes:

02 Jun 2017 01:58:11 PM Catherine Spence  
See Attached Notes  
need re-working

06 Jun 2017 08:16:09 AM Jonathan Connor  
SN working on new labels, symbols need changing

---

Issues:95915

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014912 ORD80538**

Issue ID #72305

Date Created 01/Jun/2016

Issued To: General

Completed Status: 17 Jun 2016

Stock Ref: 0014912

Opera A/c: 00007356

Order Number: ORD80538

---

Notes:

01 Jun 2016 Catherine Spence

See Attached Notes

ship with other order when stock is available

---

Issues:72305

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007356 0014912 ORD86085**

Issue ID #93754

Date Created 02/May/2017

Issued To: General

Completed Status: 02 May 2017

Stock Ref: 0014912

Opera A/c: 00007356

Order Number: ORD86085

---

Notes:

02 May 2017 07:56:06 AM Catherine Spence  
Shipping Tue 02 May 2017

---

Issues:93754

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014601 ORD81774**

Issue ID #76517

Date Created 10/Aug/2016

Issued To: General

Completed Status: 17 Aug 2016

Stock Ref: 0014601

Opera A/c: 00007374

Order Number: ORD81774

---

Notes:

10 Aug 2016 Katie Evans

See Attached Notes

Would like to wait until they have another order

---

Issues:76517

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014701 ORD81774**

Issue ID #76978

Date Created 16/Aug/2016

Issued To: General

Completed Status: 17 Aug 2016

Stock Ref: 0014701

Opera A/c: 00007374

Order Number: ORD81774

---

Notes:

16 Aug 2016 Catherine Spence  
Shipping Tue 16 Aug 2016

---

Issues:76978

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014704 ORD81306**

Issue ID #74881

Date Created 13/Jul/2016

Issued To: General

Completed Status: 13 Jul 2016

Stock Ref: 0014704

Opera A/c: 00007374

Order Number: ORD81306

---

Notes:

13 Jul 2016 Catherine Spence  
Shipping Wed 13 Jul 2016

---

Issues:74881

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014704 ORD84923**

Issue ID #89328

Date Created 15/Feb/2017

Issued To: General

Completed Status: 21 Mar 2017

Stock Ref: 0014704

Opera A/c: 00007374

Order Number: ORD84923

---

Notes:

15 Feb 2017 09:38:36 AM Catherine Spence  
Back Order - Supplier

16 Feb 2017 03:12:11 PM Catherine Spence  
See Attached Notes  
ship with other orders

20 Feb 2017 07:56:19 AM Catherine Spence  
Shipping Mon 20 Feb 2017

20 Feb 2017 08:01:17 AM Catherine Spence  
See Attached Notes  
ship with other orders

24 Feb 2017 07:46:35 AM Catherine Spence  
Back Order - Supplier  
part shipped as requested

20 Mar 2017 03:04:13 PM Catherine Spence  
checking to see if they want now or with next order

21 Mar 2017 07:32:22 AM Catherine Spence  
Shipping Tue 21 Mar 2017

---

Issues:89328

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014708 ORD82958**

Issue ID #82063

Date Created 21/Oct/2016

Issued To: General

Completed Status: 21 Oct 2016

Stock Ref: 0014708

Opera A/c: 00007374

Order Number: ORD82958

---

Notes:

21 Oct 2016 07:41:49 AM Catherine Spence  
Shipping Fri 21 Oct 2016

---

Issues:82063

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014711 ORD81192**

Issue ID #74362

Date Created 05/Jul/2016

Issued To: General

Completed Status: 12 Sep 2016

Stock Ref: 0014711

Opera A/c: 00007374

Order Number: ORD81192

---

Notes:

05 Jul 2016 Katie Evans  
Customer aware of delay

06 Jul 2016 Catherine Spence  
Back Order - Supplier

05 Sep 2016 Katie Evans  
Customer advised that we aim to ship this week

05 Sep 2016 Katie Evans  
When 0014711 is available, please ship with 2530004, 0110017, 0110022, 0110057 and 0110040

09 Sep 2016 Katie Evans  
Please could we ship SRS65562 with this order if at all possible?

---

Issues:74362

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014715 ORD82335**

Issue ID #79060

Date Created 14/Sep/2016

Issued To: General

Completed Status: 23 Sep 2016

Stock Ref: 0014715

Opera A/c: 00007374

Order Number: ORD82335

---

Notes:

14 Sep 2016 Katie Evans  
Please ship with ORD82176

15 Sep 2016 Catherine Spence  
See Attached Notes  
ord 82176 was shipped an hour before this order was shipped, please advise of they want this shipping now or with next order

15 Sep 2016 Emily Hanson  
Katie emailing them now

15 Sep 2016 Katie Evans  
As this was unable to ship with ORD82176, please hold off on shipping for now

23 Sep 2016 Catherine Spence  
Shipping Fri 23 Sep 2016

---

Issues:79060

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014715 ORD87025**

Issue ID #97642

Date Created 28/Jun/2017

Issued To: General

Completed Status: 28 Jun 2017

Stock Ref: 0014715

Opera A/c: 00007374

Order Number: ORD87025

---

Notes:

28 Jun 2017 07:51:01 AM Catherine Spence  
Shipping Wed 28 Jun 2017

28 Jun 2017 09:43:54 AM Katie Evans  
Please ship both orders together

---

Issues:97642

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014715 ORD87167**

Issue ID #98213

Date Created 05/Jul/2017

Issued To: General

Completed Status: 07 Jul 2017

Stock Ref: 0014715

Opera A/c: 00007374

Order Number: ORD87167

---

Notes:

05 Jul 2017 03:40:53 PM Katie Evans

See Attached Notes

Please ship with both outstanding orders

06 Jul 2017 07:59:46 AM Catherine Spence

Shipping Thu 06 Jul 2017

---

Issues:98213

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014724 ORD79898**

Issue ID #70423

Date Created 25/Apr/2016

Issued To: General

Completed Status: 28 Apr 2016

Stock Ref: 0014724

Opera A/c: 00007374

Order Number: ORD79898

---

Notes:

25 Apr 2016 Catherine Spence  
Back Order - Production

---

Issues:70423

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014724 ORD81413**

Issue ID #75587

Date Created 25/Jul/2016

Issued To: General

Completed Status: 29 Jul 2016

Stock Ref: 0014724

Opera A/c: 00007374

Order Number: ORD81413

---

Notes:

25 Jul 2016 Catherine Spence  
Back Order - Production

29 Jul 2016 Catherine Spence  
Shipping Fri 29 Jul 2016

---

Issues:75587

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014724 ORD82176**

Issue ID #78447

Date Created 06/Sep/2016

Issued To: General

Completed Status: 15 Sep 2016

Stock Ref: 0014724

Opera A/c: 00007374

Order Number: ORD82176

---

Notes:

06 Sep 2016 Catherine Spence  
Back Order - Supplier

14 Sep 2016 Katie Evans  
Please ship with ORD82335

---

Issues:78447

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014724 ORD83472**

Issue ID #83867

Date Created 18/Nov/2016

Issued To: General

Completed Status: 23 Nov 2016

Stock Ref: 0014724

Opera A/c: 00007374

Order Number: ORD83472

---

Notes:

18 Nov 2016 01:39:38 PM Katie Evans

See Attached Notes

Warehouse request sent, told customer we aim to ship Wednesday at the latest as per JC

21 Nov 2016 07:48:38 AM Catherine Spence

Back Order - Production

---

Issues:83867

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014724 ORD84744**

Issue ID #88508

Date Created 06/Feb/2017

Issued To: General

Completed Status: 15 Feb 2017

Stock Ref: 0014724

Opera A/c: 00007374

Order Number: ORD84744

---

Notes:

06 Feb 2017 11:53:47 AM Katie Evans

See Attached Notes

Please ship with ORD84662 and ORD84750

10 Feb 2017 08:56:00 AM Catherine Spence

Back Order - Supplier

---

Issues:88508

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014724 ORD84976**

Issue ID #89550

Date Created 17/Feb/2017

Issued To: General

Completed Status: 22 Feb 2017

Stock Ref: 0014724

Opera A/c: 00007374

Order Number: ORD84976

---

Notes:

17 Feb 2017 08:12:14 AM Catherine Spence  
Back Order - Supplier

22 Feb 2017 07:19:30 AM Catherine Spence  
Shipping Wed 22 Feb 2017  
now in stock

---

Issues:89550

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014775 ORD85077**

Issue ID #89959

Date Created 24/Feb/2017

Issued To: General

Completed Status: 24 Feb 2017

Stock Ref: 0014775

Opera A/c: 00007374

Order Number: ORD85077

---

Notes:

24 Feb 2017 07:49:59 AM Catherine Spence  
See Attached Notes  
do they want this now or shipping with other orders

24 Feb 2017 09:23:02 AM Katie Evans  
Please ship now

---

Issues:89959

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014801 ORD82201**

Issue ID #78550

Date Created 07/Sep/2016

Issued To: General

Completed Status: 12 Sep 2016

Stock Ref: 0014801

Opera A/c: 00007374

Order Number: ORD82201

---

Notes:

07 Sep 2016 Catherine Spence  
See Attached Notes  
ship with other orders

---

Issues:78550

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014802 ORD81842**

Issue ID #76979

Date Created 16/Aug/2016

Issued To: General

Completed Status: 17 Aug 2016

Stock Ref: 0014802

Opera A/c: 00007374

Order Number: ORD81842

---

Notes:

16 Aug 2016 Catherine Spence  
Shipping Tue 16 Aug 2016

---

Issues:76979

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014802 ORD85285**

Issue ID #90767

Date Created 09/Mar/2017

Issued To: General

Completed Status: 10 Mar 2017

Stock Ref: 0014802

Opera A/c: 00007374

Order Number: ORD85285

---

Notes:

09 Mar 2017 03:25:06 PM Emily Hanson  
ship with next order.

10 Mar 2017 07:36:13 AM Catherine Spence  
Shipping Fri 10 Mar 2017

---

Issues:90767

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014802 ORD85296**

Issue ID #90827

Date Created 10/Mar/2017

Issued To: General

Completed Status: 10 Mar 2017

Stock Ref: 0014802

Opera A/c: 00007374

Order Number: ORD85296

---

Notes:

10 Mar 2017 07:36:23 AM Catherine Spence  
Shipping Fri 10 Mar 2017

---

Issues:90827

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014957 ORD81974**

Issue ID #77657

Date Created 24/Aug/2016

Issued To: General

Completed Status: 24 Aug 2016

Stock Ref: 0014957

Opera A/c: 00007374

Order Number: ORD81974

---

Notes:

24 Aug 2016 Catherine Spence  
Shipping Wed 24 Aug 2016

---

Issues:77657

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014957 ORD84362**

Issue ID #87073

Date Created 13/Jan/2017

Issued To: General

Completed Status: 13 Jan 2017

Stock Ref: 0014957

Opera A/c: 00007374

Order Number: ORD84362

---

Notes:

13 Jan 2017 10:51:33 AM Emily Hanson  
DO NOT SHIP> they are adding more stuff to the order

13 Jan 2017 11:18:06 AM Emily Hanson  
okay to go, ship with other order/

---

Issues:87073

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014957 ORD87746**

Issue ID #100367

Date Created 09/Aug/2017

Issued To: General

Completed Status: 09 Aug 2017

Stock Ref: 0014957

Opera A/c: 00007374

Order Number: ORD87746

---

Notes:

09 Aug 2017 07:42:37 AM Catherine Spence  
Shipping Wed 09 Aug 2017

---

Issues:100367

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD80361**

Issue ID #71740

Date Created 19/May/2016

Issued To: General

Completed Status: 20 May 2016

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD80361

---

Notes:

19 May 2016 Katie Evans

See Attached Notes

Please part ship

---

Issues:71740

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD82606**

Issue ID #80442

Date Created 03/Oct/2016

Issued To: General

Completed Status: 05 Oct 2016

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD82606

---

Notes:

03 Oct 2016 Catherine Spence  
See Attached Notes  
ship with ord 82476

---

Issues:80442

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD84387**

Issue ID #87220

Date Created 17/Jan/2017

Issued To: General

Completed Status: 17 Jan 2017

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD84387

---

Notes:

17 Jan 2017 11:55:57 AM Catherine Spence

Shipping Tue 17 Jan 2017

Part Shipped as requested

---

Issues:87220

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD85328**

Issue ID #90970

Date Created 14/Mar/2017

Issued To: General

Completed Status: 14 Mar 2017

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD85328

---

Notes:

14 Mar 2017 09:37:01 AM Catherine Spence  
Shipping Tue 14 Mar 2017

---

Issues:90970

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD85840**

Issue ID #92991

Date Created 18/Apr/2017

Issued To: General

Completed Status: 19 Apr 2017

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD85840

---

Notes:

18 Apr 2017 02:29:44 PM Catherine Spence  
Shipping Tue 18 Apr 2017  
part shipped as requested

---

Issues:92991

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD86397**

Issue ID #95138

Date Created 19/May/2017

Issued To: General

Completed Status: 22 May 2017

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD86397

---

Notes:

19 May 2017 09:17:25 AM Katie Evans

See Attached Notes

Please ship with ORD85840

---

Issues:95138

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD87258**

Issue ID #98612

Date Created 12/Jul/2017

Issued To: General

Completed Status: 12 Jul 2017

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD87258

---

Notes:

12 Jul 2017 07:57:22 AM Catherine Spence  
Shipping Wed 12 Jul 2017

---

Issues:98612

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD87514**

Issue ID #99385

Date Created 25/Jul/2017

Issued To: General

Completed Status: 26 Jul 2017

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD87514

---

Notes:

25 Jul 2017 03:26:40 PM Katie Evans

See Attached Notes

Please ship with SRS66080

---

Issues:99385

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD88248**

Issue ID #102724

Date Created 12/Sep/2017

Issued To: General

Completed Status: 13 Sep 2017

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD88248

---

Notes:

12 Sep 2017 07:48:33 AM Catherine Spence  
Shipping Tue 12 Sep 2017

---

Issues:102724

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014958 ORD88976**

Issue ID #106381

Date Created 24/Oct/2017

Issued To: General

Completed Status: Still Outstanding

Stock Ref: 0014958

Opera A/c: 00007374

Order Number: ORD88976

---

Notes:

24 Oct 2017 07:42:00 AM Catherine Spence  
Shipping Tue 24 Oct 2017

25 Oct 2017 08:44:27 AM Catherine Spence  
Consolidate Orders  
Ship with ORD88918

---

Issues:106381

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014959 ORD83819**

Issue ID #85043

Date Created 08/Dec/2016

Issued To: General

Completed Status: 08 Dec 2016

Stock Ref: 0014959

Opera A/c: 00007374

Order Number: ORD83819

---

Notes:

08 Dec 2016 08:39:30 AM Catherine Spence  
Shipping Thu 08 Dec 2016

---

Issues:85043

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014959 ORD84987**

Issue ID #89651

Date Created 20/Feb/2017

Issued To: General

Completed Status: 22 Feb 2017

Stock Ref: 0014959

Opera A/c: 00007374

Order Number: ORD84987

---

Notes:

20 Feb 2017 08:01:08 AM Catherine Spence

See Attached Notes

ship with other orders

---

Issues:89651

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014959 ORD86171**

Issue ID #94175

Date Created 05/May/2017

Issued To: General

Completed Status: 05 May 2017

Stock Ref: 0014959

Opera A/c: 00007374

Order Number: ORD86171

---

Notes:

05 May 2017 07:33:42 AM Catherine Spence  
Shipping Fri 05 May 2017

---

Issues:94175

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014959 ORD87892**

Issue ID #101008

Date Created 18/Aug/2017

Issued To: General

Completed Status: 18 Aug 2017

Stock Ref: 0014959

Opera A/c: 00007374

Order Number: ORD87892

---

Notes:

18 Aug 2017 07:57:14 AM Catherine Spence  
Shipping Fri 18 Aug 2017

---

Issues:101008

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014981 ORD83046**

Issue ID #82304

Date Created 26/Oct/2016

Issued To: General

Completed Status: 02 Nov 2016

Stock Ref: 0014981

Opera A/c: 00007374

Order Number: ORD83046

---

Notes:

26 Oct 2016 04:08:32 PM Katie Evans

See Attached Notes

Please ship with ORD82861

01 Nov 2016 02:29:39 PM Katie Evans

Please ship ORD82861, 83046 and 83130 together

---

Issues:82304

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014981 ORD86464**

Issue ID #95369

Date Created 24/May/2017

Issued To: General

Completed Status: 24 May 2017

Stock Ref: 0014981

Opera A/c: 00007374

Order Number: ORD86464

---

Notes:

24 May 2017 07:56:08 AM Catherine Spence  
Shipping Wed 24 May 2017

---

Issues:95369

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014981 ORD88958**

Issue ID #106459

Date Created 25/Oct/2017

Issued To: General

Completed Status: Still Outstanding

Stock Ref: 0014981

Opera A/c: 00007374

Order Number: ORD88958

---

Notes:

25 Oct 2017 08:43:55 AM Catherine Spence  
Consolidate Orders  
ship ORD88918

---

Issues:106459

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014982 ORD79903**

Issue ID #70425

Date Created 25/Apr/2016

Issued To: General

Completed Status: 08 Jun 2016

Stock Ref: 0014982

Opera A/c: 00007374

Order Number: ORD79903

---

Notes:

25 Apr 2016 Catherine Spence  
Shipping Mon 25 Apr 2016

27 Apr 2016 Catherine Spence  
Back Order - Supplier  
part shipment done as requested by customer

03 Jun 2016 Katie Evans  
Amy aware of lead time

---

Issues:70425

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014982 ORD80361**

Issue ID #71850

Date Created 23/May/2016

Issued To: General

Completed Status: 08 Jun 2016

Stock Ref: 0014982

Opera A/c: 00007374

Order Number: ORD80361

---

Notes:

23 May 2016 Catherine Spence  
Back Order - Supplier

03 Jun 2016 Katie Evans  
Amy aware of lead time

---

Issues:71850

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014982 ORD85442**

Issue ID #91312

Date Created 21/Mar/2017

Issued To: General

Completed Status: 21 Mar 2017

Stock Ref: 0014982

Opera A/c: 00007374

Order Number: ORD85442

---

Notes:

21 Mar 2017 07:44:48 AM Catherine Spence  
Shipping Tue 21 Mar 2017

---

Issues:91312

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014982 ORD85840**

Issue ID #93484

Date Created 26/Apr/2017

Issued To: General

Completed Status: 22 May 2017

Stock Ref: 0014982

Opera A/c: 00007374

Order Number: ORD85840

---

Notes:

26 Apr 2017 07:46:28 AM Catherine Spence  
Back Order - Supplier

27 Apr 2017 09:53:24 AM Katie Evans  
Aware of lead time

18 May 2017 03:32:18 PM Katie Evans  
Please ship with ORD86397

---

Issues:93484

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014982 ORD86932**

Issue ID #97143

Date Created 21/Jun/2017

Issued To: General

Completed Status: 27 Jun 2017

Stock Ref: 0014982

Opera A/c: 00007374

Order Number: ORD86932

---

Notes:

21 Jun 2017 01:02:43 PM Katie Evans

See Attached Notes

Asked if they would like us to ship immediately or wait for their next order

26 Jun 2017 11:12:34 AM Katie Evans

No response, please ship

---

Issues:97143

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014982 ORD88088**

Issue ID #102015

Date Created 01/Sep/2017

Issued To: General

Completed Status: 01 Sep 2017

Stock Ref: 0014982

Opera A/c: 00007374

Order Number: ORD88088

---

Notes:

01 Sep 2017 07:42:01 AM Catherine Spence  
Shipping Fri 01 Sep 2017

---

Issues:102015

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014982 ORD88824**

Issue ID #105531

Date Created 16/Oct/2017

Issued To: General

Completed Status: 18 Oct 2017

Stock Ref: 0014982

Opera A/c: 00007374

Order Number: ORD88824

---

Notes:

16 Oct 2017 01:40:14 PM Sarah Walton

See Attached Notes

Emailed customer to advise them that their account has been placed on hold.

18 Oct 2017 10:23:08 AM Sarah Walton

SN has authorised this to be shipped.

I have spoken to Lisa in accounts and she is going to make payment.

---

Issues:105531

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014982 ORD89043**

Issue ID #106651

Date Created 27/Oct/2017

Issued To: General

Completed Status: Still Outstanding

Stock Ref: 0014982

Opera A/c: 00007374

Order Number: ORD89043

---

Notes:

27 Oct 2017 07:54:39 AM Catherine Spence

See Attached Notes

Ship with other orders

---

Issues:106651

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD81119**

Issue ID #74148

Date Created 01/Jul/2016

Issued To: General

Completed Status: 01 Jul 2016

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD81119

---

Notes:

01 Jul 2016 Catherine Spence  
Shipping Fri 01 Jul 2016

---

Issues:74148

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD83670**

Issue ID #84622

Date Created 01/Dec/2016

Issued To: General

Completed Status: 01 Dec 2016

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD83670

---

Notes:

01 Dec 2016 08:12:35 AM Catherine Spence  
Shipping Thu 01 Dec 2016

---

Issues:84622

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD84387**

Issue ID #87384

Date Created 19/Jan/2017

Issued To: General

Completed Status: 24 Apr 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD84387

---

Notes:

19 Jan 2017 07:40:57 AM Catherine Spence  
Back Order - Supplier

19 Jan 2017 04:26:50 PM Catherine Spence  
Part shipment has been done

---

Issues:87384

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD84745**

Issue ID #88569

Date Created 07/Feb/2017

Issued To: General

Completed Status: 24 Apr 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD84745

---

Notes:

07 Feb 2017 07:50:40 AM Catherine Spence  
Back Order - Supplier

---

Issues:88569

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD85043**

Issue ID #89857

Date Created 23/Feb/2017

Issued To: General

Completed Status: 24 Apr 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD85043

---

Notes:

23 Feb 2017 07:59:18 AM Catherine Spence  
Back Order - Supplier  
part shipped

---

Issues:89857

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD87328**

Issue ID #98799

Date Created 14/Jul/2017

Issued To: General

Completed Status: 19 Jul 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD87328

---

Notes:

14 Jul 2017 11:36:06 AM Catherine Spence  
See Attached Notes  
ship with other orders

18 Jul 2017 07:44:12 AM Catherine Spence  
Shipping Tue 18 Jul 2017

19 Jul 2017 07:54:50 AM Catherine Spence  
Shipping Wed 19 Jul 2017

---

Issues:98799

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD87514**

Issue ID #99505

Date Created 27/Jul/2017

Issued To: General

Completed Status: 20 Sep 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD87514

---

Notes:

27 Jul 2017 08:15:15 AM Catherine Spence  
Back Order - Supplier

27 Jul 2017 08:21:24 AM Katie Evans  
Customer aware of lead time

23 Aug 2017 09:16:57 AM Katie Evans  
Advised late September shipping date, customer would like to keep them on back order

---

Issues:99505

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD87779**

Issue ID #100577

Date Created 11/Aug/2017

Issued To: General

Completed Status: 20 Sep 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD87779

---

Notes:

11 Aug 2017 07:48:16 AM Catherine Spence  
Back Order - Supplier

14 Aug 2017 03:52:12 PM Katie Evans  
See Attached Notes  
To advise of lead time when we have one

15 Aug 2017 07:31:52 AM Catherine Spence  
Back Order - Supplier

23 Aug 2017 09:17:01 AM Katie Evans  
Advised late September shipping date, customer would like to keep them on back order

---

Issues:100577

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD87926**

Issue ID #101290

Date Created 22/Aug/2017

Issued To: General

Completed Status: 20 Sep 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD87926

---

Notes:

22 Aug 2017 07:39:11 AM Katie Evans  
See Attached Notes  
Customer aware of lead time

23 Aug 2017 09:17:06 AM Katie Evans  
Advised late September shipping date, customer would like to keep them on back order

29 Aug 2017 08:47:42 AM Catherine Spence  
Back Order - Supplier

---

Issues:101290

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD87973**

Issue ID #101358

Date Created 23/Aug/2017

Issued To: General

Completed Status: 20 Sep 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD87973

---

Notes:

23 Aug 2017 09:17:14 AM Katie Evans

See Attached Notes

Advised late September shipping date, customer would like to keep them on back order

29 Aug 2017 08:47:57 AM Catherine Spence

Back Order - Supplier

---

Issues:101358

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0014983 ORD88248**

Issue ID #102818

Date Created 13/Sep/2017

Issued To: General

Completed Status: 20 Sep 2017

Stock Ref: 0014983

Opera A/c: 00007374

Order Number: ORD88248

---

Notes:

13 Sep 2017 08:22:29 AM Catherine Spence

Back Order - Supplier

Part shipped as requested

---

Issues:102818

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0110017 ORD82172**

Issue ID #78402

Date Created 05/Sep/2016

Issued To: General

Completed Status: 12 Sep 2016

Stock Ref: 0110017

Opera A/c: 00007374

Order Number: ORD82172

---

Notes:

05 Sep 2016 Katie Evans

When 0014711 is available, please ship with 2530004, 0110017, 0110022, 0110057 and 0110040

06 Sep 2016 Catherine Spence

See Attached Notes

ship with ord81192

---

Issues:78402

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0110022 ORD82172**

Issue ID #78403

Date Created 05/Sep/2016

Issued To: General

Completed Status: 12 Sep 2016

Stock Ref: 0110022

Opera A/c: 00007374

Order Number: ORD82172

---

Notes:

05 Sep 2016 Katie Evans

When 0014711 is available, please ship with 2530004, 0110017, 0110022, 0110057 and 0110040

---

Issues:78403

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0110040 ORD82172**

Issue ID #78405

Date Created 05/Sep/2016

Issued To: General

Completed Status: 12 Sep 2016

Stock Ref: 0110040

Opera A/c: 00007374

Order Number: ORD82172

---

Notes:

05 Sep 2016 Katie Evans

When 0014711 is available, please ship with 2530004, 0110017, 0110022, 0110057 and 0110040

---

Issues:78405

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 0110057 ORD82172**

Issue ID #78406

Date Created 05/Sep/2016

Issued To: General

Completed Status: 12 Sep 2016

Stock Ref: 0110057

Opera A/c: 00007374

Order Number: ORD82172

---

Notes:

05 Sep 2016 Katie Evans

See Attached Notes

When 0014711 is available, please ship with 2530004, 0110017, 0110022, 0110057 and 0110040

---

Issues:78406

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007374 2530004 ORD82073**

Issue ID #78404

Date Created 05/Sep/2016

Issued To: General

Completed Status: 12 Sep 2016

Stock Ref: 2530004

Opera A/c: 00007374

Order Number: ORD82073

---

Notes:

05 Sep 2016 Katie Evans

When 0014711 is available, please ship with 2530004, 0110017, 0110022, 0110057 and 0110040

---

Issues:78404

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014650 ORD81473**

Issue ID #75597

Date Created 25/Jul/2016

Issued To: General

Completed Status: 25 Jul 2016

Stock Ref: 0014650

Opera A/c: 00007515

Order Number: ORD81473

---

Notes:

25 Jul 2016 Catherine Spence  
Shipping Mon 25 Jul 2016

---

Issues:75597

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014650 ORD81820**

Issue ID #77945

Date Created 30/Aug/2016

Issued To: General

Completed Status: 10 Oct 2016

Stock Ref: 0014650

Opera A/c: 00007515

Order Number: ORD81820

---

Notes:

30 Aug 2016 Catherine Spence  
Repair - In progress

02 Sep 2016 Sarah Hawker  
DO NOT SHIP CUSTOMER HAS EMAILED SAYING THEY DO NOT WANT THE PART BACK SARAHH  
02/09/2016

08 Sep 2016 Katie Evans  
Customer was mistaken, he would like repair back as it is covered by warranty (as per JC) Please ship  
entire repair with next order

---

Issues:77945

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014650 ORD84503**

Issue ID #87633

Date Created 23/Jan/2017

Issued To: General

Completed Status: 22 Mar 2017

Stock Ref: 0014650

Opera A/c: 00007515

Order Number: ORD84503

---

Notes:

23 Jan 2017 03:28:02 PM Katie Evans

See Attached Notes

Please ship with next order

---

Issues:87633

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014650 ORD86038**

Issue ID #94400

Date Created 09/May/2017

Issued To: General

Completed Status: 26 May 2017

Stock Ref: 0014650

Opera A/c: 00007515

Order Number: ORD86038

---

Notes:

09 May 2017 08:29:47 AM Katie Evans

See Attached Notes

Please ship with next order

---

Issues:94400

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014650 ORD87429**

Issue ID #99361

Date Created 25/Jul/2017

Issued To: General

Completed Status: 28 Jul 2017

Stock Ref: 0014650

Opera A/c: 00007515

Order Number: ORD87429

---

Notes:

25 Jul 2017 09:56:38 AM Katie Evans

See Attached Notes

Waiting for remaining repairs

27 Jul 2017 09:46:29 AM Katie Evans

Both repairs are now ready, please ship together

---

Issues:99361

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014650 ORD88430**

Issue ID #103440

Date Created 21/Sep/2017

Issued To: General

Completed Status: 22 Sep 2017

Stock Ref: 0014650

Opera A/c: 00007515

Order Number: ORD88430

---

Notes:

21 Sep 2017 07:58:32 AM Catherine Spence  
Shipping Thu 21 Sep 2017

22 Sep 2017 07:44:17 AM Catherine Spence  
Shipping Fri 22 Sep 2017

---

Issues:103440

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014750 ORD81822**

Issue ID #76973

Date Created 16/Aug/2016

Issued To: General

Completed Status: 23 Aug 2016

Stock Ref: 0014750

Opera A/c: 00007515

Order Number: ORD81822

---

Notes:

16 Aug 2016 Catherine Spence  
See Attached Notes  
ship with other repair

---

Issues:76973

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014750 ORD82949**

Issue ID #82078

Date Created 21/Oct/2016

Issued To: General

Completed Status: 04 Nov 2016

Stock Ref: 0014750

Opera A/c: 00007515

Order Number: ORD82949

---

Notes:

21 Oct 2016 10:25:32 AM Catherine Spence  
See Attached Notes  
ship with next order

---

Issues:82078

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014750 ORD83974**

Issue ID #85669

Date Created 19/Dec/2016

Issued To: General

Completed Status: 19 Dec 2016

Stock Ref: 0014750

Opera A/c: 00007515

Order Number: ORD83974

---

Notes:

19 Dec 2016 07:34:40 AM Catherine Spence  
Shipping Mon 19 Dec 2016

---

Issues:85669

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014825 ORD82703**

Issue ID #81166

Date Created 10/Oct/2016

Issued To: General

Completed Status: 04 Nov 2016

Stock Ref: 0014825

Opera A/c: 00007515

Order Number: ORD82703

---

Notes:

10 Oct 2016 Katie Evans  
See Attached Notes  
offered part shipment

10 Oct 2016 Katie Evans  
Please part ship

18 Oct 2016 07:40:48 AM Catherine Spence  
part shipment was done last week

18 Oct 2016 07:40:58 AM Catherine Spence  
Back Order - Supplier

---

Issues:81166

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014825 ORD83686**

Issue ID #84634

Date Created 01/Dec/2016

Issued To: General

Completed Status: 16 Dec 2016

Stock Ref: 0014825

Opera A/c: 00007515

Order Number: ORD83686

---

Notes:

01 Dec 2016 12:08:31 PM Katie Evans

See Attached Notes

Please ship with next order

---

Issues:84634

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014825 ORD85913**

Issue ID #93426

Date Created 25/Apr/2017

Issued To: General

Completed Status: 25 Apr 2017

Stock Ref: 0014825

Opera A/c: 00007515

Order Number: ORD85913

---

Notes:

25 Apr 2017 09:52:00 AM Katie Evans

See Attached Notes

Please part ship this item ASAP

25 Apr 2017 11:11:51 AM Catherine Spence

done

---

Issues:93426

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014835 ORD85229**

Issue ID #91120

Date Created 16/Mar/2017

Issued To: General

Completed Status: 11 Apr 2017

Stock Ref: 0014835

Opera A/c: 00007515

Order Number: ORD85229

---

Notes:

16 Mar 2017 12:23:18 PM Katie Evans  
Back Order - Supplier

---

Issues:91120

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014994 ORD82704**

Issue ID #81149

Date Created 10/Oct/2016

Issued To: General

Completed Status: 14 Oct 2016

Stock Ref: 0014994

Opera A/c: 00007515

Order Number: ORD82704

---

Notes:

10 Oct 2016 Catherine Spence  
Shipping Mon 10 Oct 2016

11 Oct 2016 Catherine Spence  
Shipping Tue 11 Oct 2016

---

Issues:81149

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014994 ORD85471**

Issue ID #91411

Date Created 22/Mar/2017

Issued To: General

Completed Status: 26 May 2017

Stock Ref: 0014994

Opera A/c: 00007515

Order Number: ORD85471

---

Notes:

22 Mar 2017 07:42:07 AM Catherine Spence  
Back Order - Supplier

22 Mar 2017 10:21:12 AM Katie Evans  
Please part ship

11 Apr 2017 08:26:17 AM Catherine Spence  
part shipment done

25 May 2017 08:39:50 AM Catherine Spence  
Shipping Next Week  
ship with other orders

---

Issues:91411

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007515 0014994 ORD86396**

Issue ID #95445

Date Created 25/May/2017

Issued To: General

Completed Status: 26 May 2017

Stock Ref: 0014994

Opera A/c: 00007515

Order Number: ORD86396

---

Notes:

25 May 2017 08:43:05 AM Catherine Spence  
See Attached Notes  
ship with other orders

---

Issues:95445

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014704 ORD82342**

Issue ID #79128

Date Created 15/Sep/2016

Issued To: General

Completed Status: 26 Sep 2016

Stock Ref: 0014704

Opera A/c: 00007530

Order Number: ORD82342

---

Notes:

15 Sep 2016 Katie Evans

See Attached Notes

Please do not part ship

---

Issues:79128

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014704 ORD84232**

Issue ID #86793

Date Created 10/Jan/2017

Issued To: General

Completed Status: 16 Feb 2017

Stock Ref: 0014704

Opera A/c: 00007530

Order Number: ORD84232

---

Notes:

10 Jan 2017 08:04:30 AM Catherine Spence  
Back Order - Supplier

24 Jan 2017 07:44:53 AM Catherine Spence  
Part shipment has been done as requested by customer last week.

07 Feb 2017 09:29:10 AM Katie Evans  
Amending order, please do not ship

08 Feb 2017 09:47:12 AM Katie Evans  
New order added instead, please ship everything together

---

Issues:86793

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014704 ORD84779**

Issue ID #88645

Date Created 08/Feb/2017

Issued To: General

Completed Status: 16 Feb 2017

Stock Ref: 0014704

Opera A/c: 00007530

Order Number: ORD84779

---

Notes:

08 Feb 2017 07:45:28 AM Catherine Spence  
Back Order - Supplier

---

Issues:88645

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014704 ORD87203**

Issue ID #98382

Date Created 07/Jul/2017

Issued To: General

Completed Status: 08 Sep 2017

Stock Ref: 0014704

Opera A/c: 00007530

Order Number: ORD87203

---

Notes:

07 Jul 2017 11:14:41 AM Katie Evans

See Attached Notes

Customer aware of lead time, please part ship

12 Jul 2017 07:54:40 AM Catherine Spence

Back Order - Supplier

---

Issues:98382

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014704 ORD87732**

Issue ID #100360

Date Created 09/Aug/2017

Issued To: General

Completed Status: 19 Sep 2017

Stock Ref: 0014704

Opera A/c: 00007530

Order Number: ORD87732

---

Notes:

09 Aug 2017 07:39:10 AM Catherine Spence  
Back Order - Supplier

14 Aug 2017 03:49:25 PM Katie Evans  
See Attached Notes  
SW advised of lead time

31 Aug 2017 07:39:59 AM Catherine Spence  
Back Order - Supplier

14 Sep 2017 02:04:18 PM Zoey Teal  
Please ship with ORD886877 and ORD88332.

---

Issues:100360

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014723 ORD82342**

Issue ID #79823

Date Created 23/Sep/2016

Issued To: General

Completed Status: 26 Sep 2016

Stock Ref: 0014723

Opera A/c: 00007530

Order Number: ORD82342

---

Notes:

23 Sep 2016 Katie Evans

See Attached Notes

JC confirmed these are ready, please ship asap

26 Sep 2016 Catherine Spence

Shipping Mon 26 Sep 2016

---

Issues:79823

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014723 ORD82916**

Issue ID #81976

Date Created 20/Oct/2016

Issued To: General

Completed Status: 24 Oct 2016

Stock Ref: 0014723

Opera A/c: 00007530

Order Number: ORD82916

---

Notes:

20 Oct 2016 07:50:15 AM Catherine Spence  
In Production

---

Issues:81976

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014723 ORD83326**

Issue ID #83350

Date Created 11/Nov/2016

Issued To: General

Completed Status: 18 Nov 2016

Stock Ref: 0014723

Opera A/c: 00007530

Order Number: ORD83326

---

Notes:

11 Nov 2016 08:09:04 AM Catherine Spence  
Back Order - Supplier

11 Nov 2016 12:20:26 PM Sarah Walton  
See Attached Notes  
Customer aware that there is a slight delay and advised that we will aim to ship by the end of next week.

---

Issues:83350

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014723 ORD85170**

Issue ID #90305

Date Created 02/Mar/2017

Issued To: General

Completed Status: 09 Mar 2017

Stock Ref: 0014723

Opera A/c: 00007530

Order Number: ORD85170

---

Notes:

02 Mar 2017 08:27:58 AM Catherine Spence  
Back Order - Supplier

03 Mar 2017 08:27:31 AM Katie Evans  
Advised that we expect to ship 6/7th as per JC

---

Issues:90305

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014723 ORD86252**

Issue ID #94506

Date Created 10/May/2017

Issued To: General

Completed Status: 16 May 2017

Stock Ref: 0014723

Opera A/c: 00007530

Order Number: ORD86252

---

Notes:

10 May 2017 01:45:43 PM Sarah Walton

See Attached Notes

Customer aware of the lead time and happy to wait for the full order to be shipped together.

---

Issues:94506

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014723 ORD86877**

Issue ID #97161

Date Created 21/Jun/2017

Issued To: General

Completed Status: 19 Sep 2017

Stock Ref: 0014723

Opera A/c: 00007530

Order Number: ORD86877

---

Notes:

21 Jun 2017 03:43:45 PM Zoey Teal  
Advised lead time.

21 Jun 2017 03:44:11 PM Zoey Teal  
See Attached Notes

26 Jun 2017 07:46:43 AM Catherine Spence  
Back Order - Supplier

11 Aug 2017 02:20:46 PM Sarah Walton  
See Attached Notes  
Waiting for part number 0014703 to arrive, due the end of August.

14 Aug 2017 07:57:23 AM Catherine Spence  
Shipping Mon 14 Aug 2017

16 Aug 2017 07:41:23 AM Catherine Spence  
Back Order - Production

14 Sep 2017 02:05:05 PM Zoey Teal  
Please ship with ORD88332 and ORD87732.

---

Issues:97161

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014723 ORD88332**

Issue ID #102997

Date Created 15/Sep/2017

Issued To: General

Completed Status: 19 Sep 2017

Stock Ref: 0014723

Opera A/c: 00007530

Order Number: ORD88332

---

Notes:

15 Sep 2017 08:15:46 AM Catherine Spence  
Back Order - Supplier

---

Issues:102997

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007530 0014902 ORD81247**

Issue ID #74662

Date Created 11/Jul/2016

Issued To: General

Completed Status: 11 Jul 2016

Stock Ref: 0014902

Opera A/c: 00007530

Order Number: ORD81247

---

Notes:

11 Jul 2016 Catherine Spence  
Shipping Mon 11 Jul 2016

---

Issues:74662

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007539 0014723 ORD86832**

Issue ID #96781

Date Created 15/Jun/2017

Issued To: General

Completed Status: 18 Sep 2017

Stock Ref: 0014723

Opera A/c: 00007539

Order Number: ORD86832

---

Notes:

15 Jun 2017 02:23:40 PM Sarah Walton  
See Attached Notes  
Customer has confirmed that they would like part shipment

20 Jun 2017 08:48:48 AM Catherine Spence  
Back Order - Supplier

20 Jun 2017 08:49:58 AM Catherine Spence  
no stock to part ship

28 Jun 2017 04:33:05 PM Sarah Walton  
See Attached Notes  
Waiting for a lead time so that I can update the customer.

30 Jun 2017 08:08:50 AM Catherine Spence  
Shipping Fri 30 Jun 2017

03 Jul 2017 07:33:40 AM Catherine Spence  
Back Order - Production

11 Aug 2017 01:10:10 PM Sarah Walton  
See Attached Notes  
Have asked CG to look into this for me as this has been on back order for some time now and not sure why.

11 Aug 2017 02:14:45 PM Sarah Walton  
Waiting for part number 0014703 to arrive from Bluepoint

06 Sep 2017 08:34:19 AM Catherine Spence  
Back Order - Production

07 Sep 2017 09:20:31 AM Sarah Walton

See Attached Notes  
Customer aware of the lead time.

---

Issues:96781

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007542 0014650 ORD80254**

Issue ID #71439

Date Created 16/May/2016

Issued To: General

Completed Status: 23 May 2016

Stock Ref: 0014650

Opera A/c: 00007542

Order Number: ORD80254

---

Notes:

16 May 2016 Catherine Spence  
Shipping Mon 16 May 2016

16 May 2016 Catherine Spence  
See Attached Notes  
ship with repairs

---

Issues:71439

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007542 0014825 ORD81705**

Issue ID #76284

Date Created 05/Aug/2016

Issued To: General

Completed Status: 15 Aug 2016

Stock Ref: 0014825

Opera A/c: 00007542

Order Number: ORD81705

---

Notes:

05 Aug 2016 Catherine Spence  
See Attached Notes  
Ship with SRS65541

05 Aug 2016 Katie Evans  
asked customer if they would like SRS65549 shipping with this too

08 Aug 2016 Katie Evans  
They would like 3 x SRS repairs sent together with this PO. Last SRS on it`s way to us

11 Aug 2016 Katie Evans  
Please ship this PO with SRS65557 and SRS65541, please do not ship SRS65549 yet

---

Issues:76284

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007542 0014825 ORD87276**

Issue ID #98637

Date Created 12/Jul/2017

Issued To: General

Completed Status: 15 Aug 2017

Stock Ref: 0014825

Opera A/c: 00007542

Order Number: ORD87276

---

Notes:

12 Jul 2017 02:26:52 PM Katie Evans

See Attached Notes

Offered part shipment

13 Jul 2017 07:45:27 AM Catherine Spence

Back Order - Supplier

11 Aug 2017 02:21:39 PM Sarah Walton

See Attached Notes

These have just arrived into the building.

---

Issues:98637

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007592 0012160 ORD82215**

Issue ID #78598

Date Created 07/Sep/2016

Issued To: General

Completed Status: 19 Sep 2016

Stock Ref: 0012160

Opera A/c: 00007592

Order Number: ORD82215

---

Notes:

07 Sep 2016 Sarah Hawker

Customer aware 0014825 we aim to ship w/c 12th sept and part number 2710155 aim to ship w/c 05th October.

Offered part shipment sarahh 07/09/2016

07 Sep 2016 Sarah Hawker

Please Part ship sarahh 07/09/2016

07 Sep 2016 Sarah Hawker

Please ship once part number 0014825 arrives week commencing 12th sept

14 Sep 2016 Sarah Hawker

Please give sarahh tracking number once shipped please sarahh 14.09.2016

---

Issues:78598

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007594 0014592 ORD84141**

Issue ID #86142

Date Created 29/Dec/2016

Issued To: General

Completed Status: 31 Jan 2017

Stock Ref: 0014592

Opera A/c: 00007594

Order Number: ORD84141

---

Notes:

29 Dec 2016 11:20:38 AM Emily Hanson  
emailed to see if they want par shipping.

29 Dec 2016 12:55:24 PM Emily Hanson  
please part ship

05 Jan 2017 11:52:03 AM Catherine Spence  
See Attached Notes  
Part shippment done

06 Jan 2017 07:37:23 AM Catherine Spence  
Back Order - Supplier

31 Jan 2017 08:57:10 AM Emily Hanson  
wanting 5 0014835 will change the order today.

---

Issues:86142

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007594 0014751 ORD88122**

Issue ID #102491

Date Created 07/Sep/2017

Issued To: General

Completed Status: 12 Sep 2017

Stock Ref: 0014751

Opera A/c: 00007594

Order Number: ORD88122

---

Notes:

07 Sep 2017 04:11:35 PM Sarah Walton

See Attached Notes

Customer aware of the lead time.

---

Issues:102491

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007594 0014835 ORD84651**

Issue ID #88148

Date Created 31/Jan/2017

Issued To: General

Completed Status: 02 Feb 2017

Stock Ref: 0014835

Opera A/c: 00007594

Order Number: ORD84651

---

Notes:

31 Jan 2017 11:35:33 AM Catherine Spence

01 Feb 2017 07:52:36 AM Catherine Spence  
Back Order - Supplier

---

Issues:88148

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007594 0014835 ORD85113**

Issue ID #90149

Date Created 28/Feb/2017

Issued To: General

Completed Status: 28 Feb 2017

Stock Ref: 0014835

Opera A/c: 00007594

Order Number: ORD85113

---

Notes:

28 Feb 2017 07:36:06 AM Catherine Spence  
Shipping Tue 28 Feb 2017

---

Issues:90149

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007594 0014835 ORD88122**

Issue ID #102032

Date Created 01/Sep/2017

Issued To: General

Completed Status: 12 Sep 2017

Stock Ref: 0014835

Opera A/c: 00007594

Order Number: ORD88122

---

Notes:

01 Sep 2017 02:48:04 PM Catherine Spence  
Consolidate Orders  
ship with other order

---

Issues:102032

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007688 0014596 ORD83093**

Issue ID #82620

Date Created 01/Nov/2016

Issued To: Catherine Spence

Completed Status: 08 Dec 2016

Stock Ref: 0014596

Opera A/c: 00007688

Order Number: ORD83093

---

Notes:

01 Nov 2016 07:32:30 AM Catherine Spence  
Back Order - Supplier

02 Nov 2016 08:37:25 AM Emily Hanson  
ship everything we have in stock.

02 Nov 2016 08:39:27 AM Emily Hanson  
please

08 Nov 2016 07:32:55 AM Catherine Spence  
part shipment done

08 Nov 2016 07:32:55 AM Catherine Spence  
part shipment done

08 Dec 2016 08:56:34 AM Emily Hanson  
please ship the alternative of 0014751

---

Issues:82620

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007688 0014751 ORD83093**

Issue ID #85572

Date Created 16/Dec/2016

Issued To: General

Completed Status: 16 Dec 2016

Stock Ref: 0014751

Opera A/c: 00007688

Order Number: ORD83093

---

Notes:

16 Dec 2016 07:41:55 AM Catherine Spence  
Shipping Fri 16 Dec 2016

---

Issues:85572

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007688 0014996 ORD81150**

Issue ID #74308

Date Created 05/Jul/2016

Issued To: General

Completed Status: 05 Jul 2016

Stock Ref: 0014996

Opera A/c: 00007688

Order Number: ORD81150

---

Notes:

05 Jul 2016 Catherine Spence  
Shipping Tue 05 Jul 2016

---

Issues:74308

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007688 0014996 ORD82674**

Issue ID #80876

Date Created 06/Oct/2016

Issued To: General

Completed Status: 10 Oct 2016

Stock Ref: 0014996

Opera A/c: 00007688

Order Number: ORD82674

---

Notes:

06 Oct 2016 Catherine Spence  
Shipping Thu 06 Oct 2016

06 Oct 2016 Catherine Spence  
See Attached Notes  
to -relabel

10 Oct 2016 Catherine Spence  
Shipping Mon 10 Oct 2016

---

Issues:80876

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007688 0014996 ORD83093**

Issue ID #82674

Date Created 01/Nov/2016

Issued To: General

Completed Status: 16 Dec 2016

Stock Ref: 0014996

Opera A/c: 00007688

Order Number: ORD83093

---

Notes:

01 Nov 2016 02:13:17 PM Emily Hanson

asked if they want a part shipment as this item isnt in stock- lead time around 3-4 weeks. WC 5th December

---

Issues:82674

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007688 0014996 ORD84947**

Issue ID #89425

Date Created 16/Feb/2017

Issued To: General

Completed Status: 17 Feb 2017

Stock Ref: 0014996

Opera A/c: 00007688

Order Number: ORD84947

---

Notes:

16 Feb 2017 09:24:44 AM Catherine Spence  
Shipping Thu 16 Feb 2017

---

Issues:89425

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007786 0014750 ORD87397**

Issue ID #99014

Date Created 19/Jul/2017

Issued To: General

Completed Status: 20 Jul 2017

Stock Ref: 0014750

Opera A/c: 00007786

Order Number: ORD87397

---

Notes:

19 Jul 2017 07:51:14 AM Catherine Spence  
Shipping Wed 19 Jul 2017

---

Issues:99014

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007885 0014912 ORD81993**

Issue ID #77762

Date Created 25/Aug/2016

Issued To: General

Completed Status: 14 Oct 2016

Stock Ref: 0014912

Opera A/c: 00007885

Order Number: ORD81993

---

Notes:

25 Aug 2016 Emily Hanson  
SW on with this..they need some documents.

31 Aug 2016 Sarah Walton  
See Attached Notes  
Customer is aware of a lead time of 8 weeks - 19th October.

09 Sep 2016 Catherine Spence  
Back Order - Supplier

---

Issues:77762

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00007995 0014750 ORD85173**

Issue ID #90475

Date Created 06/Mar/2017

Issued To: General

Completed Status: 06 Mar 2017

Stock Ref: 0014750

Opera A/c: 00007995

Order Number: ORD85173

---

Notes:

06 Mar 2017 07:45:50 AM Catherine Spence  
Shipping Mon 06 Mar 2017

---

Issues:90475

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00008043 0014750 ORD85310**

Issue ID #90906

Date Created 13/Mar/2017

Issued To: General

Completed Status: 13 Mar 2017

Stock Ref: 0014750

Opera A/c: 00008043

Order Number: ORD85310

---

Notes:

13 Mar 2017 07:40:57 AM Catherine Spence  
Shipping Mon 13 Mar 2017

---

Issues:90906

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00008044 0014816 ORD85915**

Issue ID #93131

Date Created 20/Apr/2017

Issued To: General

Completed Status: 21 Apr 2017

Stock Ref: 0014816

Opera A/c: 00008044

Order Number: ORD85915

---

Notes:

20 Apr 2017 08:18:38 AM Catherine Spence  
Shipping Thu 20 Apr 2017

---

Issues:93131

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00009044 0014750 ORD80380**

Issue ID #72293

Date Created 01/Jun/2016

Issued To: General

Completed Status: 09 Jun 2016

Stock Ref: 0014750

Opera A/c: 00009044

Order Number: ORD80380

---

Notes:

01 Jun 2016 Sarah Walton  
See Attached Notes  
customer aware of the delay

09 Jun 2016 Catherine Spence  
Shipping Thu 09 Jun 2016

---

Issues:72293

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00009044 0014750 ORD86442**

Issue ID #95284

Date Created 23/May/2017

Issued To: General

Completed Status: 23 May 2017

Stock Ref: 0014750

Opera A/c: 00009044

Order Number: ORD86442

---

Notes:

23 May 2017 07:41:13 AM Catherine Spence  
Shipping Tue 23 May 2017

---

Issues:95284

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00009130 0014835 ORD87390**

Issue ID #99077

Date Created 20/Jul/2017

Issued To: General

Completed Status: 01 Aug 2017

Stock Ref: 0014835

Opera A/c: 00009130

Order Number: ORD87390

---

Notes:

20 Jul 2017 10:20:35 AM Katie Evans

See Attached Notes

SJH advised customer that 7 of each will ship this week and the remainder will be 2-3 weeks

25 Jul 2017 07:56:02 AM Catherine Spence

part shipped last week

---

Issues:99077

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00011982 0014994 ORD88301**

Issue ID #102810

Date Created 13/Sep/2017

Issued To: General

Completed Status: 13 Sep 2017

Stock Ref: 0014994

Opera A/c: 00011982

Order Number: ORD88301

---

Notes:

13 Sep 2017 08:17:30 AM Catherine Spence  
Shipping Wed 13 Sep 2017

---

Issues:102810

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00011982 0014996 ORD82902**

Issue ID #81888

Date Created 19/Oct/2016

Issued To: General

Completed Status: 26 Oct 2016

Stock Ref: 0014996

Opera A/c: 00011982

Order Number: ORD82902

---

Notes:

19 Oct 2016 07:42:24 AM Catherine Spence  
Shipping Wed 19 Oct 2016

19 Oct 2016 09:36:33 AM Catherine Spence  
See Attached Notes  
re-label stock

25 Oct 2016 04:19:50 PM Jonathan Connor  
Please part ship minus the teccare

---

Issues:81888

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012605 0014601 ORD86073**

Issue ID #93753

Date Created 02/May/2017

Issued To: General

Completed Status: 03 May 2017

Stock Ref: 0014601

Opera A/c: 00012605

Order Number: ORD86073

---

Notes:

02 May 2017 07:53:42 AM Catherine Spence  
Back Order - Supplier

---

Issues:93753

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012605 0014625 ORD84074**

Issue ID #85821

Date Created 21/Dec/2016

Issued To: General

Completed Status: 22 Dec 2016

Stock Ref: 0014625

Opera A/c: 00012605

Order Number: ORD84074

---

Notes:

21 Dec 2016 07:28:32 AM Catherine Spence  
Shipping Wed 21 Dec 2016

---

Issues:85821

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012605 0014625 ORD85443**

Issue ID #91604

Date Created 24/Mar/2017

Issued To: General

Completed Status: 24 Mar 2017

Stock Ref: 0014625

Opera A/c: 00012605

Order Number: ORD85443

---

Notes:

24 Mar 2017 10:23:01 AM Sophie Lines  
Okay to proceed

---

Issues:91604

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012605 0014625 ORD87567**

Issue ID #99639

Date Created 28/Jul/2017

Issued To: General

Completed Status: 28 Jul 2017

Stock Ref: 0014625

Opera A/c: 00012605

Order Number: ORD87567

---

Notes:

28 Jul 2017 09:08:43 AM Catherine Spence  
Shipping Fri 28 Jul 2017

---

Issues:99639

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014751 ORD89038**

Issue ID #106654

Date Created 27/Oct/2017

Issued To: General

Completed Status: Still Outstanding

Stock Ref: 0014751

Opera A/c: 00012672

Order Number: ORD89038

---

Notes:

27 Oct 2017 07:57:39 AM Catherine Spence  
Shipping Fri 27 Oct 2017

---

Issues:106654

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD79888**

Issue ID #70427

Date Created 25/Apr/2016

Issued To: General

Completed Status: 26 Apr 2016

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD79888

---

Notes:

25 Apr 2016 Catherine Spence  
Back Order - Production  
to be re-labeled

---

Issues:70427

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD81159**

Issue ID #74269

Date Created 04/Jul/2016

Issued To: General

Completed Status: 08 Jul 2016

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD81159

---

Notes:

04 Jul 2016 Sarah Hawker

Emailed customer to say we aim to ship by 08/07/2016 sarahh 04/07/2016

04 Jul 2016 Sarah Hawker

Emailed customer to say we aim to ship by 08/07/2016 sarahh 04/07/2016

05 Jul 2016 Catherine Spence

See Attached Notes

to be re-labeled

08 Jul 2016 Catherine Spence

Shipping Fri 08 Jul 2016

---

Issues:74269

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD81455**

Issue ID #75489

Date Created 22/Jul/2016

Issued To: General

Completed Status: 15 Sep 2016

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD81455

---

Notes:

22 Jul 2016 Catherine Spence  
Back Order - Supplier

22 Jul 2016 Sarah Hawker  
Please part ship sarahh 22/07/2016

27 Jul 2016 Catherine Spence  
part shipment done

---

Issues:75489

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD82041**

Issue ID #77951

Date Created 30/Aug/2016

Issued To: General

Completed Status: 15 Sep 2016

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD82041

---

Notes:

30 Aug 2016 Catherine Spence  
Back Order - Supplier

05 Sep 2016 Sarah Walton  
See Attached Notes  
Customer aware of the lead time - week beginning 5th Sept.

06 Sep 2016 Catherine Spence  
Back Order - Supplier

---

Issues:77951

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD82064**

Issue ID #77990

Date Created 30/Aug/2016

Issued To: General

Completed Status: 16 Sep 2016

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD82064

---

Notes:

30 Aug 2016 Katie Evans

See Attached Notes

Advised of 1-2 week lead time (6-13th September)

31 Aug 2016 Catherine Spence

Back Order - Supplier

---

Issues:77990

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD83721**

Issue ID #84810

Date Created 05/Dec/2016

Issued To: General

Completed Status: 05 Dec 2016

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD83721

---

Notes:

05 Dec 2016 08:33:24 AM Catherine Spence  
Shipping Mon 05 Dec 2016

---

Issues:84810

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD84517**

Issue ID #87602

Date Created 23/Jan/2017

Issued To: General

Completed Status: 24 Jan 2017

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD84517

---

Notes:

23 Jan 2017 11:51:25 AM Emily Hanson  
OKAY TO GO

---

Issues:87602

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD84837**

Issue ID #88814

Date Created 10/Feb/2017

Issued To: General

Completed Status: 24 Apr 2017

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD84837

---

Notes:

10 Feb 2017 08:57:58 AM Catherine Spence  
Back Order - Supplier

14 Feb 2017 09:10:44 AM Emily Hanson  
aware of 2-3 weeks lead time.

10 Mar 2017 09:31:27 AM Emily Hanson  
advised mid april lead time.

---

Issues:88814

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD84922**

Issue ID #89326

Date Created 15/Feb/2017

Issued To: General

Completed Status: 24 Apr 2017

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD84922

---

Notes:

15 Feb 2017 09:37:51 AM Catherine Spence  
Back Order - Supplier

10 Mar 2017 09:31:30 AM Emily Hanson  
advised mid april lead time.

---

Issues:89326

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD87518**

Issue ID #99437

Date Created 26/Jul/2017

Issued To: General

Completed Status: 26 Jul 2017

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD87518

---

Notes:

26 Jul 2017 07:51:41 AM Catherine Spence  
Shipping Wed 26 Jul 2017

---

Issues:99437

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014990 ORD88995**

Issue ID #106463

Date Created 25/Oct/2017

Issued To: General

Completed Status: 25 Oct 2017

Stock Ref: 0014990

Opera A/c: 00012672

Order Number: ORD88995

---

Notes:

25 Oct 2017 08:50:17 AM Catherine Spence  
Shipping Wed 25 Oct 2017

---

Issues:106463

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014994 ORD85208**

Issue ID #90480

Date Created 06/Mar/2017

Issued To: General

Completed Status: 06 Mar 2017

Stock Ref: 0014994

Opera A/c: 00012672

Order Number: ORD85208

---

Notes:

06 Mar 2017 07:50:43 AM Catherine Spence  
Shipping Mon 06 Mar 2017

---

Issues:90480

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012672 0014994 ORD85311**

Issue ID #90907

Date Created 13/Mar/2017

Issued To: General

Completed Status: 13 Mar 2017

Stock Ref: 0014994

Opera A/c: 00012672

Order Number: ORD85311

---

Notes:

13 Mar 2017 07:41:12 AM Catherine Spence  
Shipping Mon 13 Mar 2017

---

Issues:90907

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012748 0014750 ORD80758**

Issue ID #72814

Date Created 13/Jun/2016

Issued To: General

Completed Status: 13 Jun 2016

Stock Ref: 0014750

Opera A/c: 00012748

Order Number: ORD80758

---

Notes:

13 Jun 2016 Catherine Spence  
Shipping Mon 13 Jun 2016

---

Issues:72814

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012906 0014904 ORD82470**

Issue ID #79876

Date Created 26/Sep/2016

Issued To: General

Completed Status: 26 Sep 2016

Stock Ref: 0014904

Opera A/c: 00012906

Order Number: ORD82470

---

Notes:

26 Sep 2016 Catherine Spence  
Shipping Mon 26 Sep 2016

---

Issues:79876

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012906 0014904 ORD83406**

Issue ID #85050

Date Created 08/Dec/2016

Issued To: General

Completed Status: 05 Jan 2017

Stock Ref: 0014904

Opera A/c: 00012906

Order Number: ORD83406

---

Notes:

08 Dec 2016 10:48:37 AM Katie Evans

See Attached Notes

Please do not ship, ship date moving to 5th January. Needs to be called on the 4th of January to confirm he is ready to receive them

04 Jan 2017 12:22:26 PM Katie Evans

Elan to call me back to confirm shipment date

04 Jan 2017 02:55:35 PM Katie Evans

Please ship 05/01/17 as planned

05 Jan 2017 07:50:57 AM Catherine Spence

Shipping Thu 05 Jan 2017

---

Issues:85050

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012906 0014904 ORD85297**

Issue ID #90828

Date Created 10/Mar/2017

Issued To: General

Completed Status: 14 Mar 2017

Stock Ref: 0014904

Opera A/c: 00012906

Order Number: ORD85297

---

Notes:

10 Mar 2017 07:36:34 AM Catherine Spence  
Back Order - Supplier

10 Mar 2017 10:03:54 AM Katie Evans  
See Attached Notes  
To part ship 42 please

---

Issues:90828

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012907 0014996 ORD82919**

Issue ID #81913

Date Created 19/Oct/2016

Issued To: General

Completed Status: 20 Oct 2016

Stock Ref: 0014996

Opera A/c: 00012907

Order Number: ORD82919

---

Notes:

19 Oct 2016 02:27:59 PM Catherine Spence  
See Attached Notes  
needs to be re-labeled

---

Issues:81913

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012923 0014750 ORD79668**

Issue ID #69929

Date Created 14/Apr/2016

Issued To: General

Completed Status: 28 Apr 2016

Stock Ref: 0014750

Opera A/c: 00012923

Order Number: ORD79668

---

Notes:

14 Apr 2016 Sarah Hawker

Aware of delay does not want part shipment 14/04/16 sarahh

19 Apr 2016 Sarah Hawker

advised customer that there order should be ready for dispatch week commencing 2nd may. if further delays please contact customer

neil.hayward@wellsoffshore.co.uk

28 Apr 2016 Catherine Spence

Shipping Thu 28 Apr 2016

---

Issues:69929

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012923 0014750 ORD81910**

Issue ID #77969

Date Created 30/Aug/2016

Issued To: General

Completed Status: 05 Sep 2016

Stock Ref: 0014750

Opera A/c: 00012923

Order Number: ORD81910

---

Notes:

30 Aug 2016 Sarah Hawker

awaiting a date from suppliers for part number 4420621 - spoke to JS and EH who confirmed we dont have a date as of yet sarahh 30/08/16

---

Issues:77969

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012923 4410520 ORD79668**

Issue ID #69665

Date Created 08/Apr/2016

Issued To: General

Completed Status: 28 Apr 2016

Stock Ref: 4410520

Opera A/c: 00012923

Order Number: ORD79668

---

Notes:

08 Apr 2016 Sarah Hawker

Emailed customer to ask which free silicone sensor they require as we have 3 and to see if they would like part shipment as versastream-nasal sampling line is not in stock 08/04/16 sarahh

08 Apr 2016 Sarah Hawker

Emailed customer to see which silicone sensor they require as we have 3 types and to see if they require part shipment as the versastram-nasal sampling line is out of stock sarahh 08/04/16

12 Apr 2016 Katie Evans

customer confirmed they want 0014750. SH

12 Apr 2016 Katie Evans

Advised customer of lead time. SH

---

Issues:69665

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00012969 0014750 ORD83065**

Issue ID #82400

Date Created 27/Oct/2016

Issued To: General

Completed Status: 06 Dec 2016

Stock Ref: 0014750

Opera A/c: 00012969

Order Number: ORD83065

---

Notes:

27 Oct 2016 05:15:42 PM Katie Evans  
See Attached Notes  
With SJH

28 Oct 2016 10:31:26 AM Katie Evans  
ok to part ship

01 Nov 2016 07:40:12 AM Catherine Spence  
part ship when we get some capnographs in

24 Nov 2016 07:49:23 AM Catherine Spence  
Back Order - Supplier  
part shipment done

---

Issues:82400

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00013011 0014835 ORD85771**

Issue ID #92630

Date Created 11/Apr/2017

Issued To: General

Completed Status: 11 Apr 2017

Stock Ref: 0014835

Opera A/c: 00013011

Order Number: ORD85771

---

Notes:

11 Apr 2017 08:29:26 AM Catherine Spence  
Shipping Tue 11 Apr 2017

---

Issues:92630

---

VIAMED Stock Meeting - Sales Back Orders Review - By Customer

**Backorder 00013041 0014835 ORD88704**

Issue ID #104944

Date Created 10/Oct/2017

Issued To: General

Completed Status: 10 Oct 2017

Stock Ref: 0014835

Opera A/c: 00013041

Order Number: ORD88704

---

Notes:

10 Oct 2017 07:41:49 AM Catherine Spence  
Shipping Tue 10 Oct 2017

---

Issues:104944

---

VIAMED Sales And Marketing - Mailshots UK

**BCI compatible disposables**

Issue ID #69492

Date Created 05/Apr/2016

Issued To: Steve Nixon

Completed Status: 21 Apr 2016

---

**Notes:**

Copy Issue Main Issue #69486

Added by Catrin Hird sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

I am currently mailing a number of existing customers and distributors regarding the BCI plaster disposable sensors. All correspondence will show in Goldmine as all emails are being sent through there.

For reference the following is being offered, should you received orders or enquiries regarding this product please let me know. I have also added this memo to the stock records in Opera. Thanks

Crown Agents and UK end users (end user price minus 30%):

0014957 £86.80

0014958 £86.80

0014959 £89.60

0014960 £89.60

UK Distributors (UK distributor price minus 10%):

0014957 £72.52

0014958 £72.52

0014959 £77.12,

0014960 £77.12

---

Issues:69492

---

VIAMED Sales And Marketing - Mailshots UK

**BCI compatible disposables**

Issue ID #69490

Date Created 05/Apr/2016

Issued To: Sarah Walton

Completed Status: 06 Oct 2016

---

**Notes:**

Copy Issue Main Issue #69486

Added by Catrin Hird sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

I am currently mailing a number of existing customers and distributors regarding the BCI plaster disposable sensors. All correspondence will show in Goldmine as all emails are being sent through there.

For reference the following is being offered, should you received orders or enquiries regarding this product please let me know. I have also added this memo to the stock records in Opera. Thanks

Crown Agents and UK end users (end user price minus 30%):

0014957 £86.80

0014958 £86.80

0014959 £89.60

0014960 £89.60

UK Distributors (UK distributor price minus 10%):

0014957 £72.52

0014958 £72.52

0014959 £77.12,

0014960 £77.12

---

Issues:69490

---

VIAMED Sales And Marketing - Mailshots UK

**BCI compatible disposables**

Issue ID #69488

Date Created 05/Apr/2016

Issued To: Katie Evans

Completed Status: 06 Oct 2016

---

Notes:

Copy Issue Main Issue #69486

Added by Catrin Hird sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

I am currently mailing a number of existing customers and distributors regarding the BCI plaster disposable sensors. All correspondence will show in Goldmine as all emails are being sent through there.

For reference the following is being offered, should you received orders or enquiries regarding this product please let me know. I have also added this memo to the stock records in Opera. Thanks

Crown Agents and UK end users (end user price minus 30%):

0014957 £86.80

0014958 £86.80

0014959 £89.60

0014960 £89.60

UK Distributors (UK distributor price minus 10%):

0014957 £72.52

0014958 £72.52

0014959 £77.12,

0014960 £77.12

---

Issues:69488

---

VIAMED Stock Meeting - Sales Order Issues

**Bulk packing 0014904 disposable sensors**

Issue ID #81840

Date Created 18/Oct/2016

Issued To: Derek Lamb

Completed Status: 01 Nov 2016

---

Notes:

Copy Issue Main Issue #81839

Added by Steve Hardaker sent to Steve Nixon

I have a customer, Acorn Distributions, that has been ordering 0014904 disposable sensors 21 boxes at a time.

He ships these on by removing the boxes and bulk packing the sensors loose in a larger shipping box, he asked if we could do that for the ones we ship to him to reduce the size and weight and to prevent waste.

He also wants to up this to 1000 sensors per month, which works out at 42 boxes, on a regular call-off (still awaiting written confirmation on that).

For labelling and traceability reasons, I would suggest we put each box of 24 into a self-sealing plastic bag and print an adhesive label with the same info as the box and stick those onto the bags.

Steve, can you discuss with Derek as per our conversation to see if this is viable and allowed under ISO?

---

19 Oct 2016 Derek Lamb

If you have it in writing from Acorn Distributions, we can do what we want regarding ISO.

I agree to bagging and labelling each batch 24.

I would also suggest on the first time we do this we get the customer to photograph the delivery and send us a copy of the picture,

as part of the CE marking there is packaging validity tests. I can then add the photos to the CE file as proving the got there undamaged out of their normal boxes.

---

01 Nov 2016 Steve Hardaker

Spoke to Elan at Acorn, he accepts that this will create us work so is happy to take them in standard boxes and unpack them himself.

---

Issues:81840

---

VIAMED Stock Meeting - Sales Order Issues  
**Bulk packing 0014904 disposable sensors**  
Issue ID #81839  
Date Created 18/Oct/2016  
Issued To: Jonathan Connor  
Completed Status: 01 Nov 2016

---

Notes:

Added by Steve Hardaker sent to Steve Nixon

I have a customer, Acorn Distributions, that has been ordering 0014904 disposable sensors 21 boxes at a time.

He ships these on by removing the boxes and bulk packing the sensors loose in a larger shipping box, he asked if we could do that for the ones we ship to him to reduce the size and weight and to prevent waste.

He also wants to up this to 1000 sensors per month, which works out at 42 boxes, on a regular call-off (still awaiting written confirmation on that).

For labelling and traceability reasons, I would suggest we put each box of 24 into a self-sealing plastic bag and print an adhesive label with the same info as the box and stick those onto the bags.

Steve, can you discuss with Derek as per our conversation to see if this is viable and allowed under ISO?

---

25 Oct 2016 Steve Nixon

As discussed please work with JC and work out a suitable box and quantity it will contain. I will then set up Opera part number to cover...

Looks like we have already committed to buying 100 more boxes but no customer orders. We need to look at trying to reduce number of shipments and associated costs.

---

25 Oct 2016 Steve Hardaker

Next Action Changed From Steve Nixon To Jonathan Connor Jonathan, please can you estimate whether the following quantities of single SpO2 sensors in bags of 24 are viable in standard shipping boxes that we use?

504 - 21 boxes

1008 - 42 boxes

---

28 Oct 2016 Jonathan Connor

We can only fit 25 Boxes per cut down, there is a small amount of empty space. 21 would take up the same amount of room if they were boxed. 61 x 47 x 47.

If we unboxed them you could fit 21 per cut down at 61 x 47 x 42. This would take approx 40 minutes to re-pack plus whatever label and bag costs need to be added.

The customers carriage would be worked out based on the volumetric weight. The physical weight even boxed is smaller than the volumetric so it would save them a small amount. Based on what it would cost us to ship to the middle east it would be a saving of £9

Personally if the customer wants to remove the packaging to reduce carriage costs they can, but I don't think it is the best idea to do it for them, they already get the lowest price for these so not a fan of adding extra work to goods in/QA

---

28 Oct 2016 Jonathan Connor

Volumetric weight for 61 x 47 x 47 is 27kg. For 42 high it is 24.08, UPS usually round up to nearest Kg, so 2kg lighter

---

01 Nov 2016 Steve Hardaker

Spoke to Elan at Acorn, he accepts that this will create us work so is happy to take them in standard boxes and unpack them himself.

---

Issues:81839

---

VIAMED Sales And Marketing - Telesales

**Call Log #2149 - Acom cancelling back order**

Issue ID #90977

Date Created 14/Mar/2017

Issued To: Katie Evans

Completed Status: 14 Mar 2017

---

**Notes:**

Added by Steve Hardaker sent to Katie Evans

Elan has just rung to say to cancel his back order for 42 boxes of 0014904 on back order (ORD85297). He has received 42 boxes already against this order, I get the impression that he is concerned that if Jamjoom make a fuss with the hospital or his Saudi contact, he could get stuck with these. Please can you amend the order paperwork and remove the back order, thanks.

---

14 Mar 2017 Katie Evans

Back order removed, JC aware not to order them

---

14 Mar 2017 Steve Hardaker

---

Issues:90977

---

VIAMED Management Meeting - Other

**Certificates of origin: 0012160...**

Issue ID #70803

Date Created 03/May/2016

Issued To: Derek Lamb

Completed Status: 23 Aug 2016

---

Notes:

Added by Steve Nixon sent to Derek Lamb

As discussed can you please generate certificates of origin for:

0012160

0014590

0014750

0014756

0014835

---

19 May 2016 Derek Lamb

In the Pen Icon of stock pages

---

Issues:70803

---

VIAMED Sales And Marketing - Telesales

**Change of shipping date for back order 83406**

Issue ID #85000

Date Created 07/Dec/2016

Issued To: Sarah Walton

Completed Status: 08 Dec 2016

---

**Notes:**

Copy Issue Main Issue #84998

Added by Steve Hardaker sent to Katie Evans

I spoke to Elan at Acorn to confirm that he still wants ORD83406 for 21 boxes of 0014904 shipping on the 13th Dec, he advised that he is currently awaiting confirmation from his customer and wants to delay this.

He suggested changing shipping date to 5th Jan with a note to call him the day before to verify that he is ready to take them. Please can you amend the order date and put a note on the order checking as appropriate, thanks.

---

Issues:85000

---

VIAMED Sales And Marketing - Telesales

**Change of shipping date for back order 83406**

Issue ID #84999

Date Created 07/Dec/2016

Issued To: Emily Hanson

Completed Status: 08 Dec 2016

---

**Notes:**

Copy Issue Main Issue #84998

Added by Steve Hardaker sent to Katie Evans

I spoke to Elan at Acorn to confirm that he still wants ORD83406 for 21 boxes of 0014904 shipping on the 13th Dec, he advised that he is currently awaiting confirmation from his customer and wants to delay this.

He suggested changing shipping date to 5th Jan with a note to call him the day before to verify that he is ready to take them. Please can you amend the order date and put a note on the order checking as appropriate, thanks.

---

Issues:84999

---

VIAMED Sales And Marketing - Telesales

**Change of shipping date for back order 83406**

Issue ID #84998

Date Created 07/Dec/2016

Issued To: Sarah Walton

Completed Status: 09 Dec 2016

---

**Notes:**

Added by Steve Hardaker sent to Katie Evans

I spoke to Elan at Acorn to confirm that he still wants ORD83406 for 21 boxes of 0014904 shipping on the 13th Dec, he advised that he is currently awaiting confirmation from his customer and wants to delay this.

He suggested changing shipping date to 5th Jan with a note to call him the day before to verify that he is ready to take them. Please can you amend the order date and put a note on the order checking as appropriate, thanks.

---

08 Dec 2016 Katie Evans

Not in the office, will leave in the hands of SW and EH

---

09 Dec 2016 Steve Hardaker

Next Action Changed From Katie Evans To Sarah Walton Please can you amend the forward order due date, its currently still showing as 13/12/16 on the active list, thanks.

---

09 Dec 2016 Sarah Walton

The date had been amended by Emily, however each line still had the old date on. Have amended to the correct one.

---

09 Dec 2016 Steve Hardaker

---

Issues:84998

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Cortessis - Greek Ambulance Tender**

Issue ID #98189

Date Created 05/Jul/2017

Issued To: Katie Evans

Completed Status: 29 Sep 2017

---

**Notes:**

Copy Issue Main Issue #98187

Added by Ryan Swaine sent to Steve Nixon

Cortessis has received his PO from Peugeot and we should receive our PO next Tuesday. The quantities have been reduced slightly to the following:

143x

0012162 - VM-2160

0014750 - Included Silicone Adult SpO2 Sensor

0014895 - 1.2m Extension Cable

0022171 - Universal Mounting Plate

0121180 - V Wall Mount

Delivery will be to the ambulance manufacturer in Portugal, address details will be confirmed next week in the PO. Invoiced to Cortessis in Greece.

Delivery: Required in Portugal for middle of September.

Katie: Please update proforma invoice accordingly

---

05 Jul 2017 Katie Evans

Updated quote and emailed to Diamond

---

28 Jul 2017 Katie Evans

5 x shipping today as they requested, the remaining are on track for September

---

29 Sep 2017 Ryan Swaine

---

Issues:98189

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Cortessis - Greek Ambulance Tender**

Issue ID #98188

Date Created 05/Jul/2017

Issued To: Jonathan Connor

Completed Status: 06 Sep 2017

---

Notes:

Copy Issue Main Issue #98187

Added by Ryan Swaine sent to Steve Nixon

Cortessis has received his PO from Peugeot and we should receive our PO next Tuesday. The quantities have been reduced slightly to the following:

143x

0012162 - VM-2160

0014750 - Included Silicone Adult SpO2 Sensor

0014895 - 1.2m Extension Cable

0022171 - Universal Mounting Plate

0121180 - V Wall Mount

Delivery will be to the ambulance manufacturer in Portugal, address details will be confirmed next week in the PO. Invoiced to Cortessis in Greece.

Delivery: Required in Portugal for middle of September.

Katie: Please update proforma invoice accordingly

---

07 Jul 2017 Jonathan Connor  
Ordered

---

Issues:98188

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Cortessis - Greek Ambulance Tender**

Issue ID #98187

Date Created 05/Jul/2017

Issued To: Jonathan Connor

Completed Status: 06 Sep 2017

---

**Notes:**

Added by Ryan Swaine sent to Steve Nixon

Cortessis has received his PO from Peugeot and we should receive our PO next Tuesday. The quantities have been reduced slightly to the following:

143x

0012162 - VM-2160

0014750 - Included Silicone Adult SpO2 Sensor

0014895 - 1.2m Extension Cable

0022171 - Universal Mounting Plate

0121180 - V Wall Mount

Delivery will be to the ambulance manufacturer in Portugal, address details will be confirmed next week in the PO. Invoiced to Cortessis in Greece.

Delivery: Required in Portugal for middle of September.

Katie: Please update proforma invoice accordingly

---

06 Jul 2017 Steve Nixon

Next Action Changed From Steve Nixon To Jonathan Connor JC, please order. BPM have 100 ext. cables.

NOTE: Software needs updating to resolve screen Greek translation errors.

---

07 Jul 2017 Jonathan Connor  
Ordered

---

Issues:98187

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Cortessis Order (Viamed VM-2160)**

Issue ID #103723

Date Created 25/Sep/2017

Issued To: Jonathan Connor

Completed Status: 05 Oct 2017

---

Notes:

Added by Ryan Swaine sent to Jonathan Connor

We are entering in to another 2x tenders, which we have a good chance of winning, but the devices need to be labelled Viamed. Please order the devices and place on a hold shelf, please see SN for further details.

The total requirement for both tenders together will be:

31x VM-2160 (0012162)

31x Silicone adult sensors (to be included with devices) (0014750)

31x 1.2m Ext Cables (0014895)

6x Universal Mounting plates (0022171)

6x V Wall mount (0121180)

---

05 Oct 2017 Jonathan Connor  
ordered

---

05 Oct 2017 Ryan Swaine

---

Issues:103723

---

VIAMED Stock Meeting - Stock Queries

**Declarations of Origin Required**

Issue ID #70940

Date Created 05/May/2016

Issued To: Derek Lamb

Completed Status: 11 May 2016

---

**Notes:**

Copy Issue Main Issue #70939

Added by Katie Evans sent to Steve Nixon

Please could we create declarations of origin for part numbers:

0012160

0014590

0014750

0014756

0014835

Needed for Ryan`s customer, Golden Medical Tech, Peru

---

05 May 2016 Derek Lamb

should be linked, from the intrastats page

click the (i) Country of Origin,

should be a couple links

---

06 May 2016 Katie Evans

My apologies, part number 0014756 is unavailable, says, "needs checking";

---

09 May 2016 Derek Lamb

Fixed

---

11 May 2016 Katie Evans

Thank you

---

Issues:70940

---

VIAMED Stock Meeting - Stock Queries

**Declarations of Origin Required**

Issue ID #70939

Date Created 05/May/2016

Issued To: Steve Nixon

Completed Status: 12 May 2016

---

Notes:

Added by Katie Evans sent to Steve Nixon

Please could we create declarations of origin for part numbers:

0012160

0014590

0014750

0014756

0014835

Needed for Ryan`s customer, Golden Medical Tech, Peru

---

12 May 2016 Katie Evans

---

Issues:70939

---

VIAMED General Issue - VIAMED General Issue

**Demo stock to move**

Issue ID #74834

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 13 Jul 2016

---

Notes:

Added by Steve Hardaker sent to Jonathan Connor

Please can you scan 0111265 UltraMaxO2 ID956542 back to main stock from my demo stock. Never left the building, perfect condition, just needs batteries re-bagging.

Also, 0012160 ID842231 to Main SOR to go out to a UK distributor on SOR.

Note: the VM-2160 box also contains a sensor 0014750 ID711777 which is already scanned to Main SOR.

---

13 Jul 2016 Jonathan Connor

Returned 0111265 to QA, MG to put through QA system. 2160 returned to SOR and given back to Cathy

---

13 Jul 2016 Steve Hardaker

---

Issues:74834

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #92549

Date Created 10/Apr/2017

Issued To: Catrin Hird

Completed Status: 24 Apr 2017

---

Notes:

Added by Catrin Hird sent to Catrin Hird

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

remove from literature and website

---

10 Apr 2017 Catrin Hird

0014596 not 0014506

Removed all from website, now just need removing from literature

---

24 Apr 2017 Catrin Hird

Updated Pulse oximetry sensors for viamed oximeters leaflet and added to Z and intrastats.

Updated VM-2160 leaflet and added to z and intrastats.

---

24 Apr 2017 Catrin Hird

---

Issues:92549

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87993

Date Created 27/Jan/2017

Issued To: Steve Hardaker

Completed Status: 30 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87993

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87992

Date Created 27/Jan/2017

Issued To: Sarah Walton

Completed Status: 30 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87992

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87991

Date Created 27/Jan/2017

Issued To: Ryan Swaine

Completed Status: 27 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87991

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87990

Date Created 27/Jan/2017

Issued To: Robert Connor

Completed Status: 30 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87990

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87989

Date Created 27/Jan/2017

Issued To: Philip Crossley

Completed Status: 30 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87989

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87988

Date Created 27/Jan/2017

Issued To: Michael Green

Completed Status: 01 Feb 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87988

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87987

Date Created 27/Jan/2017

Issued To: Katie Evans

Completed Status: 14 Sep 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87987

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87986

Date Created 27/Jan/2017

Issued To: Jonathan Connor

Completed Status: 30 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87986

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87985

Date Created 27/Jan/2017

Issued To: Helen Lamb

Completed Status: 31 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87985

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87984

Date Created 27/Jan/2017

Issued To: Emma Clark

Completed Status: 14 Feb 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87984

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87983

Date Created 27/Jan/2017

Issued To: Emily Hanson

Completed Status: 30 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87983

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87982

Date Created 27/Jan/2017

Issued To: Derek Lamb

Completed Status: 31 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87982

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87981

Date Created 27/Jan/2017

Issued To: Catherine Spence

Completed Status: 30 Jan 2017

---

Notes:

Copy Issue Main Issue #87980

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

Issues:87981

---

VIAMED Sales And Marketing - Any other business

**Discontinued - Autoclavable pulse oximetry sensors**

Issue ID #87980

Date Created 27/Jan/2017

Issued To: Catrin Hird

Completed Status: 10 Apr 2017

---

Notes:

Added by Steve Nixon sent to Catrin Hird

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

Unfortunately we have had to discontinue the autoclavable pulse oximetry sensors for use with the VM-2160 and VM-2500 Series.

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

This is mainly due to lack of demand. The high cost of additional materials and associated ISO costs have also been a contributing factor.

Please refer customer enquiries to the standard non-autoclavable version sensors.

---

10 Apr 2017 Catrin Hird

Created Related Issue #92549

Added by Catrin Hird sent to Catrin Hird

0014590 1 sensor left in stock

0014591 5 sensors left in stock

0014592 0 sensors in stock

0014595 0 sensors in stock

0014506 0 sensors in stock

remove from literature and website

---

Issues:87980

---

VIAMED Stock Meeting - Stock Processing

**Disposable probes**

Issue ID #74259

Date Created 04/Jul/2016

Issued To: Catrin Hird

Completed Status: 25 Jul 2016

---

Notes:

Added by Jonathan Connor sent to Catrin Hird

Please can you update the Disposable probe labels in my T drive to include the additional 2 years use by date. 0014990 and 0014996 are needed for orders. If possible to do ready to ship end of this week. The others we will need in the future but probably better to have sooner rather than later. Thanks

---

04 Jul 2016 Catrin Hird

I need to know the lot numbers for each batch please

---

05 Jul 2016 Catrin Hird

Done 0014990, 0014996 wasn't in the original batch so please can you provide me with the lot number for this one, preferably a sample to copy. Thanks

---

05 Jul 2016 Catrin Hird

Done both in your T drive

---

11 Jul 2016 Jonathan Connor

Thanks, will you be able to do an update on the others in my T drive using the lot numbers that are on them already. I am going to give the job to label them to Callum over the next few weeks, want to get ahead of them so don't have to relabel for an order

---

11 Jul 2016 Catrin Hird

Believe I have done them all, please can you check when you are doing each box that the correct ones are used. Thanks

---

12 Jul 2016 Catrin Hird

Are you happy for me to close this issue and just reopen if you find any that need doing?

---

22 Jul 2016 Jonathan Connor

Please can you redo 0014903. Ref should be 1-IF not 8-IF

---

25 Jul 2016 Catrin Hird

Done

---

25 Jul 2016 Jonathan Connor

Thanks

---

Issues:74259

---

BS EN ISO 13485:2003 - Document Control Amendments needed

**Document Update Request (6321)**

Issue ID #81818

Date Created 18/Oct/2016

Issued To: Derek Lamb

Completed Status: 19 Oct 2016

---

**Notes:**

Added by Sarah Walton sent to Derek Lamb

Controlled Document (6321) Viamed Declaration of Conformity 0014605 Requires Updating

I have a customer who requires this document asap.

---

18 Oct 2016 Derek Lamb

technically that ones still valid,  
they wanting a newer one?

---

18 Oct 2016 Sarah Walton

The one on the system is for internal use only and when I asked Ryan he said it needs updating?

---

19 Oct 2016 Sarah Walton

---

Issues:81818

---

VIAMED Sales And Marketing - GHX Web Pricing

**GHX Stock Group Master**

Issue ID #84461

Date Created 29/Nov/2016

Issued To: Steve Hardaker

Completed Status: 06 Dec 2016

---

**Notes:**

Added by Catrin Hird sent to Steve Hardaker

Please can you remove the following parts from your stock group as they are discontinued, thanks

0014913

0014914

0014915

0014916

0014985

0014986

0014987

0014988

---

05 Dec 2016 Steve Hardaker

Removed.

---

06 Dec 2016 Catrin Hird

Thanks

---

Issues:84461

---

Intrastats - Cross Reference Database Tables Updates

**Maxtec SpO2 sensor p/n R134P09-001 is 0014701**

Issue ID #83989

Date Created 22/Nov/2016

Issued To: Steve Nixon

Completed Status: Still Outstanding

---

Notes:

Added by Steve Hardaker sent to Steve Nixon

Maxtec SpO2 sensor p/n R134P09-001 (aka model no R-3212-9) is a Nellcor non-Oximax compatible soft silicone sensor, 0.9m, same as 0014701.

Found for a customer by using Maxtecs cross reference and selecting older Nellcor oximeters.

<https://www.maxtec.com/spo2/>

---

Issues:83989

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Oxylink and 0114723 stock**

Issue ID #96890

Date Created 16/Jun/2017

Issued To: Sarah Walton

Completed Status: 25 Jul 2017

---

Notes:

Copy Issue Main Issue #96887

Added by Steve Nixon sent to Ryan Swaine

Looking at the memo on the Oxylink EUR account the following is written:

NOTE: When Piotr orders 0014723, discuss with warehouse about replenishing the stock, i.e. ensure that we have 5 in stock in readiness for his next order and ask for the sensors to be scanned to the Oxylink stock shelf.

SN 16/04/15

1) It appears that we are not currently maintaining this stock. Please discuss with me in the future with situations like this.

2) However, looking at the sales figures Piotr hasn't ordered the sensors since 11/11/2015, so it looks like he has effectively cancelled this requirement. Bearing in mind that we currently don't have any stock, can someone tactfully ask him to confirm that he no longer requires stock to be maintained for him.

---

19 Jun 2017 Sarah Walton

---

Issues:96890

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Oxylink and 0114723 stock**

Issue ID #96889

Date Created 16/Jun/2017

Issued To: Katie Evans

Completed Status: 25 Jul 2017

---

Notes:

Copy Issue Main Issue #96887

Added by Steve Nixon sent to Ryan Swaine

Looking at the memo on the Oxylink EUR account the following is written:

NOTE: When Piotr orders 0014723, discuss with warehouse about replenishing the stock, i.e. ensure that we have 5 in stock in readiness for his next order and ask for the sensors to be scanned to the Oxylink stock shelf.

SN 16/04/15

1) It appears that we are not currently maintaining this stock. Please discuss with me in the future with situations like this.

2) However, looking at the sales figures Piotr hasn't ordered the sensors since 11/11/2015, so it looks like he has effectively cancelled this requirement. Bearing in mind that we currently don't have any stock, can someone tactfully ask him to confirm that he no longer requires stock to be maintained for him.

---

20 Jun 2017 Katie Evans

---

Issues:96889

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Oxylink and 0114723 stock**

Issue ID #96888

Date Created 16/Jun/2017

Issued To: Jonathan Connor

Completed Status: 25 Jul 2017

---

Notes:

Copy Issue Main Issue #96887

Added by Steve Nixon sent to Ryan Swaine

Looking at the memo on the Oxylink EUR account the following is written:

NOTE: When Piotr orders 0014723, discuss with warehouse about replenishing the stock, i.e. ensure that we have 5 in stock in readiness for his next order and ask for the sensors to be scanned to the Oxylink stock shelf.

SN 16/04/15

1) It appears that we are not currently maintaining this stock. Please discuss with me in the future with situations like this.

2) However, looking at the sales figures Piotr hasn't ordered the sensors since 11/11/2015, so it looks like he has effectively cancelled this requirement. Bearing in mind that we currently don't have any stock, can someone tactfully ask him to confirm that he no longer requires stock to be maintained for him.

---

19 Jun 2017 Jonathan Connor

ok thanks

---

Issues:96888

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Oxylink and 0114723 stock**

Issue ID #96887

Date Created 16/Jun/2017

Issued To: Ryan Swaine

Completed Status: 25 Jul 2017

---

Notes:

Added by Steve Nixon sent to Ryan Swaine

Looking at the memo on the Oxylink EUR account the following is written:

NOTE: When Piotr orders 0014723, discuss with warehouse about replenishing the stock, i.e. ensure that we have 5 in stock in readiness for his next order and ask for the sensors to be scanned to the Oxylink stock shelf.

SN 16/04/15

1) It appears that we are not currently maintaining this stock. Please discuss with me in the future with situations like this.

2) However, looking at the sales figures Piotr hasn't ordered the sensors since 11/11/2015, so it looks like he has effectively cancelled this requirement. Bearing in mind that we currently don't have any stock, can someone tactfully ask him to confirm that he no longer requires stock to be maintained for him.

---

16 Jun 2017 Ryan Swaine

Discussed with Piotr, please cancel the memo.

---

Issues:96887

---

VIAMED Stock Meeting - Any Other Business

**Parts for SBRI SpO2 trial**

Issue ID #94343

Date Created 08/May/2017

Issued To: Jonathan Connor

Completed Status: 05 Jun 2017

---

**Notes:**

Added by Steve Nixon sent to Jonathan Connor

As discussed, can you please produce labels and supply pouches for disposable sensors, so that I can repackage the sensors that have been provided by Bluepoint. Sensors are to be used for the R&D trial at Sheffield Children`s Hospital. Don`t affix the labels.

Paediatric: 20 x 0014994, lot 242313 date 2017-12

Infant: 27 x 0014995, lot 242315 date 2017-12

Neonatal: 24 x 0014996, lot 242314 date 2017-12

I also need:

1 x 0014895 ext cable

1 x 0014896 ext cable

1 x 0014889 sensor wraps

2 x 9950051 AA batteries, pack of 10

10 x 9950053 AA batteries, pack of 3

---

18 May 2017 Jonathan Connor  
done

---

Issues:94343

---

Purchase Back Orders Review - Purchase Back Orders Review

**POR10638 0014904 Have Been Delivered or Cancelled in Opera**

Issue ID #70852

Date Created 04/May/2016

Issued To: Steve Hardaker

Completed Status: 04 May 2016

---

Notes:

Have Been Delivered or Cancelled in Opera

---

Issues:70852

---

Purchase Back Orders Review - Purchase Back Orders Review

**POR10655 0014904 Have Been Delivered or Cancelled in Opera**

Issue ID #70853

Date Created 04/May/2016

Issued To: Steve Hardaker

Completed Status: 04 May 2016

---

Notes:

Have Been Delivered or Cancelled in Opera

---

Issues:70853

---

- Back Order Report

**POR10721 0014825 //**

Issue ID #70544

Date Created 27/Apr/2016

Issued To: Jonathan Connor

Completed Status: 27 Apr 2016

---

Notes:

POR10721 0014825 Estimated ship to us date changed to / /

---

Issues:70544

---

- Back Order Report

**POR10721 0014825 31/05/2016**

Issue ID #70694

Date Created 29/Apr/2016

Issued To: Jonathan Connor

Completed Status: 29 Apr 2016

---

Notes:

POR10721 0014825 Estimated ship to us date changed to 31 / 05 / 2016

---

Issues:70694

---

- Back Order Report

**POR10721 0014982 //**

Issue ID #70545

Date Created 27/Apr/2016

Issued To: Jonathan Connor

Completed Status: 27 Apr 2016

---

Notes:

POR10721 0014982 Estimated ship to us date changed to / /

---

Issues:70545

---

- Back Order Report

**POR10721 0014982 31/05/2016**

Issue ID #70695

Date Created 29/Apr/2016

Issued To: Jonathan Connor

Completed Status: 29 Apr 2016

---

Notes:

POR10721 0014982 Estimated ship to us date changed to 31 / 05 / 2016

---

Issues:70695

---

- Back Order Report

**POR10737 0014650 //**

Issue ID #71158

Date Created 10/May/2016

Issued To: Jonathan Connor

Completed Status: 10 May 2016

---

Notes:

POR10737 0014650 Estimated ship to us date changed to / /

---

Issues:71158

---

- Back Order Report

**POR10737 0014650 17/06/2016**

Issue ID #71209

Date Created 11/May/2016

Issued To: Jonathan Connor

Completed Status: 11 May 2016

---

Notes:

POR10737 0014650 Estimated ship to us date changed to 17 / 06 / 2016

---

Issues:71209

---

- Back Order Report

**POR10743 0014902 //**

Issue ID #71200

Date Created 11/May/2016

Issued To: Jonathan Connor

Completed Status: 11 May 2016

---

Notes:

POR10743 0014902 Estimated ship to us date changed to / /

---

Issues:71200

---

- Back Order Report

**POR10743 0014902 17/06/2016**

Issue ID #71336

Date Created 13/May/2016

Issued To: Jonathan Connor

Completed Status: 13 May 2016

---

Notes:

POR10743 0014902 Estimated ship to us date changed to 17 / 06 / 2016

---

Issues:71336

---

- Back Order Report

**POR10781 0014756 //**

Issue ID #72435

Date Created 03/Jun/2016

Issued To: Jonathan Connor

Completed Status: 03 Jun 2016

---

Notes:

POR10781 0014756 Estimated ship to us date changed to / /

---

Issues:72435

---

- Back Order Report

**POR10781 0014756 10/06/2016**

Issue ID #72439

Date Created 03/Jun/2016

Issued To: Jonathan Connor

Completed Status: 03 Jun 2016

---

Notes:

POR10781 0014756 Estimated ship to us date changed to 10 / 06 / 2016

---

Issues:72439

---

- Back Order Report

**POR10832 0014711 //**

Issue ID #74811

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 12 Jul 2016

---

Notes:

POR10832 0014711 Estimated ship to us date changed to / /

---

Issues:74811

---

- Back Order Report

**POR10832 0014711 //**

Issue ID #74672

Date Created 11/Jul/2016

Issued To: Jonathan Connor

Completed Status: 11 Jul 2016

---

Notes:

POR10832 0014711 Estimated ship to us date changed to / /

---

Issues:74672

---

- Back Order Report

**POR10832 0014711 31/08/2016**

Issue ID #75079

Date Created 15/Jul/2016

Issued To: Jonathan Connor

Completed Status: 15 Jul 2016

---

Notes:

POR10832 0014711 Estimated ship to us date changed to 31 / 08 / 2016

---

Issues:75079

---

- Back Order Report

**POR10832 0014711 31/08/2016**

Issue ID #74801

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 12 Jul 2016

---

Notes:

POR10832 0014711 Estimated ship to us date changed to 31 / 08 / 2016

---

Issues:74801

---

- Back Order Report

**POR10832 0014905 //**

Issue ID #74812

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 12 Jul 2016

---

Notes:

POR10832 0014905 Estimated ship to us date changed to / /

---

Issues:74812

---

- Back Order Report

**POR10832 0014905 //**

Issue ID #74673

Date Created 11/Jul/2016

Issued To: Jonathan Connor

Completed Status: 11 Jul 2016

---

Notes:

POR10832 0014905 Estimated ship to us date changed to / /

---

Issues:74673

---

- Back Order Report

**POR10832 0014905 31/08/2016**

Issue ID #75080

Date Created 15/Jul/2016

Issued To: Jonathan Connor

Completed Status: 15 Jul 2016

---

Notes:

POR10832 0014905 Estimated ship to us date changed to 31 / 08 / 2016

---

Issues:75080

---

- Back Order Report

**POR10832 0014905 31/08/2016**

Issue ID #74802

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 12 Jul 2016

---

Notes:

POR10832 0014905 Estimated ship to us date changed to 31 / 08 / 2016

---

Issues:74802

---

Purchase Back Orders Review - Purchase Back Orders Review

**POR10832 0014905 Have Been Delivered or Cancelled in Opera**

Issue ID #79123

Date Created 15/Sep/2016

Issued To: Sarah Walton

Completed Status: 15 Sep 2016

---

Notes:

Have Been Delivered or Cancelled in Opera

---

Issues:79123

---

- Back Order Report

**POR10832 0014912 //**

Issue ID #74813

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 12 Jul 2016

---

Notes:

POR10832 0014912 Estimated ship to us date changed to / /

---

Issues:74813

---

- Back Order Report

**POR10832 0014912 //**

Issue ID #74674

Date Created 11/Jul/2016

Issued To: Jonathan Connor

Completed Status: 11 Jul 2016

---

Notes:

POR10832 0014912 Estimated ship to us date changed to / /

---

Issues:74674

---

- Back Order Report

**POR10832 0014912 31/08/2016**

Issue ID #75081

Date Created 15/Jul/2016

Issued To: Jonathan Connor

Completed Status: 15 Jul 2016

---

Notes:

POR10832 0014912 Estimated ship to us date changed to 31 / 08 / 2016

---

Issues:75081

---

- Back Order Report

**POR10832 0014912 31/08/2016**

Issue ID #74803

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 12 Jul 2016

---

Notes:

POR10832 0014912 Estimated ship to us date changed to 31 / 08 / 2016

---

Issues:74803

---

- Back Order Report

**POR10832 0014990 //**

Issue ID #74814

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 12 Jul 2016

---

Notes:

POR10832 0014990 Estimated ship to us date changed to / /

---

Issues:74814

---

- Back Order Report

**POR10832 0014990 //**

Issue ID #74675

Date Created 11/Jul/2016

Issued To: Jonathan Connor

Completed Status: 11 Jul 2016

---

Notes:

POR10832 0014990 Estimated ship to us date changed to / /

---

Issues:74675

---

- Back Order Report

**POR10832 0014990 31/08/2016**

Issue ID #75082

Date Created 15/Jul/2016

Issued To: Jonathan Connor

Completed Status: 15 Jul 2016

---

Notes:

POR10832 0014990 Estimated ship to us date changed to 31 / 08 / 2016

---

Issues:75082

---

- Back Order Report

**POR10832 0014990 31/08/2016**

Issue ID #74804

Date Created 12/Jul/2016

Issued To: Jonathan Connor

Completed Status: 12 Jul 2016

---

Notes:

POR10832 0014990 Estimated ship to us date changed to 31 / 08 / 2016

---

Issues:74804

---

- Back Order Report

**POR10845 0014825 //**

Issue ID #75146

Date Created 15/Jul/2016

Issued To: Jonathan Connor

Completed Status: 15 Jul 2016

---

Notes:

POR10845 0014825 Estimated ship to us date changed to / /

---

Issues:75146

---

- Back Order Report

**POR10845 0014825 06/09/2016**

Issue ID #75416

Date Created 21/Jul/2016

Issued To: Jonathan Connor

Completed Status: 21 Jul 2016

---

Notes:

POR10845 0014825 Estimated ship to us date changed to 06 / 09 / 2016

---

Issues:75416

---

- Back Order Report

**POR10845 0014895 //**

Issue ID #75147

Date Created 15/Jul/2016

Issued To: Jonathan Connor

Completed Status: 15 Jul 2016

---

Notes:

POR10845 0014895 Estimated ship to us date changed to / /

---

Issues:75147

---

- Back Order Report

**POR10845 0014895 06/09/2016**

Issue ID #75417

Date Created 21/Jul/2016

Issued To: Jonathan Connor

Completed Status: 21 Jul 2016

---

Notes:

POR10845 0014895 Estimated ship to us date changed to 06 / 09 / 2016

---

Issues:75417

---

- Back Order Report

**POR10901 0014912 //**

Issue ID #77756

Date Created 25/Aug/2016

Issued To: Jonathan Connor

Completed Status: 26 Aug 2016

---

Notes:

POR10901 0014912 Estimated ship to us date changed to / /

---

Issues:77756

---

- Back Order Report

**POR10901 0014912 //**

Issue ID #77754

Date Created 25/Aug/2016

Issued To: Jonathan Connor

Completed Status: 26 Aug 2016

---

Notes:

POR10901 0014912 Estimated ship to us date changed to / /

---

Issues:77754

---

- Back Order Report

**POR10901 0014912 05/10/2016**

Issue ID #78039

Date Created 31/Aug/2016

Issued To: Jonathan Connor

Completed Status: 31 Aug 2016

---

Notes:

POR10901 0014912 Estimated ship to us date changed to 05 / 10 / 2016

---

Issues:78039

---

- Back Order Report

**POR10917 0014904 //**

Issue ID #78678

Date Created 08/Sep/2016

Issued To: Jonathan Connor

Completed Status: 08 Sep 2016

---

Notes:

POR10917 0014904 Estimated ship to us date changed to / /

---

Issues:78678

---

- Back Order Report

**POR10917 0014904 //**

Issue ID #78070

Date Created 31/Aug/2016

Issued To: Jonathan Connor

Completed Status: 01 Sep 2016

---

Notes:

POR10917 0014904 Estimated ship to us date changed to / /

---

Issues:78070

---

- Back Order Report

**POR10917 0014904 //**

Issue ID #78051

Date Created 31/Aug/2016

Issued To: Jonathan Connor

Completed Status: 31 Aug 2016

---

Notes:

POR10917 0014904 Estimated ship to us date changed to / /

---

Issues:78051

---

- Back Order Report

**POR10927 0014825 //**

Issue ID #80068

Date Created 27/Sep/2016

Issued To: Jonathan Connor

Completed Status: 27 Sep 2016

---

Notes:

POR10927 0014825 Estimated ship to us date changed to / /

---

Issues:80068

---

- Back Order Report

**POR10927 0014825 //**

Issue ID #79911

Date Created 26/Sep/2016

Issued To: Jonathan Connor

Completed Status: 26 Sep 2016

---

Notes:

POR10927 0014825 Estimated ship to us date changed to / /

---

Issues:79911

---

- Back Order Report

**POR10927 0014825 //**

Issue ID #78474

Date Created 06/Sep/2016

Issued To: Jonathan Connor

Completed Status: 06 Sep 2016

---

Notes:

POR10927 0014825 Estimated ship to us date changed to / /

---

Issues:78474

---

- Back Order Report

**POR10934 0014825 //**

Issue ID #78854

Date Created 12/Sep/2016

Issued To: Jonathan Connor

Completed Status: 13 Sep 2016

---

Notes:

POR10934 0014825 Estimated ship to us date changed to / /

---

Issues:78854

---

- Back Order Report

**POR10934 0014825 //**

Issue ID #78849

Date Created 12/Sep/2016

Issued To: Jonathan Connor

Completed Status: 13 Sep 2016

---

Notes:

POR10934 0014825 Estimated ship to us date changed to / /

---

Issues:78849

---

- Back Order Report

**POR10934 0014825 24/10/2016**

Issue ID #79046

Date Created 14/Sep/2016

Issued To: Jonathan Connor

Completed Status: 15 Sep 2016

---

Notes:

POR10934 0014825 Estimated ship to us date changed to 24 / 10 / 2016

---

Issues:79046

---

- Back Order Report

**POR10935 0014750 //**

Issue ID #78856

Date Created 12/Sep/2016

Issued To: Jonathan Connor

Completed Status: 13 Sep 2016

---

Notes:

POR10935 0014750 Estimated ship to us date changed to / /

---

Issues:78856

---

- Back Order Report

**POR10935 0014750 //**

Issue ID #78848

Date Created 12/Sep/2016

Issued To: Jonathan Connor

Completed Status: 13 Sep 2016

---

Notes:

POR10935 0014750 Estimated ship to us date changed to / /

---

Issues:78848

---

- Back Order Report

**POR10935 0014750 24/10/2016**

Issue ID #79045

Date Created 14/Sep/2016

Issued To: Jonathan Connor

Completed Status: 15 Sep 2016

---

Notes:

POR10935 0014750 Estimated ship to us date changed to 24 / 10 / 2016

---

Issues:79045

---

- Back Order Report

**POR10958 0014902 //**

Issue ID #79914

Date Created 26/Sep/2016

Issued To: Jonathan Connor

Completed Status: 26 Sep 2016

---

Notes:

POR10958 0014902 Estimated ship to us date changed to / /

---

Issues:79914

---

- Back Order Report

**POR10958 0014902 //**

Issue ID #79372

Date Created 19/Sep/2016

Issued To: Jonathan Connor

Completed Status: 19 Sep 2016

---

Notes:

POR10958 0014902 Estimated ship to us date changed to / /

---

Issues:79372

---

- Back Order Report

**POR10958 0014902 //**

Issue ID #79367

Date Created 19/Sep/2016

Issued To: Jonathan Connor

Completed Status: 19 Sep 2016

---

Notes:

POR10958 0014902 Estimated ship to us date changed to / /

---

Issues:79367

---

- Back Order Report

**POR10958 0014902 25/10/2016**

Issue ID #80043

Date Created 27/Sep/2016

Issued To: Jonathan Connor

Completed Status: 27 Sep 2016

---

Notes:

POR10958 0014902 Estimated ship to us date changed to 25 / 10 / 2016

---

Issues:80043

---

- Back Order Report

**POR10974 0014895 //**

Issue ID #80645

Date Created 04/Oct/2016

Issued To: Jonathan Connor

Completed Status: 04 Oct 2016

---

Notes:

POR10974 0014895 Estimated ship to us date changed to / /

---

Issues:80645

---

- Back Order Report

**POR10974 0014895 04/11/2016**

Issue ID #80726

Date Created 05/Oct/2016

Issued To: Jonathan Connor

Completed Status: 05 Oct 2016

---

Notes:

POR10974 0014895 Estimated ship to us date changed to 04 / 11 / 2016

---

Issues:80726

---

- Back Order Report

**POR10974 0014909 //**

Issue ID #80646

Date Created 04/Oct/2016

Issued To: Jonathan Connor

Completed Status: 04 Oct 2016

---

Notes:

POR10974 0014909 Estimated ship to us date changed to / /

---

Issues:80646

---

- Back Order Report

**POR10974 0014909 30/11/2016**

Issue ID #80727

Date Created 05/Oct/2016

Issued To: Jonathan Connor

Completed Status: 05 Oct 2016

---

Notes:

POR10974 0014909 Estimated ship to us date changed to 30 / 11 / 2016

---

Issues:80727

---

- Back Order Report

**POR10974 0014993 //**

Issue ID #80647

Date Created 04/Oct/2016

Issued To: Jonathan Connor

Completed Status: 04 Oct 2016

---

Notes:

POR10974 0014993 Estimated ship to us date changed to / /

---

Issues:80647

---

- Back Order Report

**POR10974 0014993 30/11/2016**

Issue ID #80728

Date Created 05/Oct/2016

Issued To: Jonathan Connor

Completed Status: 05 Oct 2016

---

Notes:

POR10974 0014993 Estimated ship to us date changed to 30 / 11 / 2016

---

Issues:80728

---

- Back Order Report

**POR10974 0014994 //**

Issue ID #80648

Date Created 04/Oct/2016

Issued To: Jonathan Connor

Completed Status: 04 Oct 2016

---

Notes:

POR10974 0014994 Estimated ship to us date changed to / /

---

Issues:80648

---

- Back Order Report

**POR10974 0014994 30/11/2016**

Issue ID #80729

Date Created 05/Oct/2016

Issued To: Jonathan Connor

Completed Status: 05 Oct 2016

---

Notes:

POR10974 0014994 Estimated ship to us date changed to 30 / 11 / 2016

---

Issues:80729

---

- Back Order Report

**POR11004 0014904 //**

Issue ID #82785

Date Created 03/Nov/2016

Issued To: Jonathan Connor

Completed Status: 03 Nov 2016

---

Notes:

POR11004 0014904 Estimated ship to us date changed to / /

---

Issues:82785

---

- Back Order Report

**POR11004 0014904 //**

Issue ID #81892

Date Created 19/Oct/2016

Issued To: Jonathan Connor

Completed Status: 19 Oct 2016

---

Notes:

POR11004 0014904 Estimated ship to us date changed to / /

---

Issues:81892

---

- Back Order Report

**POR11004 0014904 //**

Issue ID #81845

Date Created 18/Oct/2016

Issued To: Jonathan Connor

Completed Status: 19 Oct 2016

---

Notes:

POR11004 0014904 Estimated ship to us date changed to / /

---

Issues:81845

---

- Back Order Report

**POR11027 0014596 //**

Issue ID #83254

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014596 Estimated ship to us date changed to / /

---

Issues:83254

---

- Back Order Report

**POR11027 0014596 //**

Issue ID #83245

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014596 Estimated ship to us date changed to / /

---

Issues:83245

---

- Back Order Report

**POR11027 0014596 //**

Issue ID #82802

Date Created 03/Nov/2016

Issued To: Jonathan Connor

Completed Status: 03 Nov 2016

---

Notes:

POR11027 0014596 Estimated ship to us date changed to / /

---

Issues:82802

---

- Back Order Report

**POR11027 0014596 12/12/2016**

Issue ID #83365

Date Created 11/Nov/2016

Issued To: Jonathan Connor

Completed Status: 11 Nov 2016

---

Notes:

POR11027 0014596 Estimated ship to us date changed to 12 / 12 / 2016

---

Issues:83365

---

- Back Order Report

**POR11027 0014703 //**

Issue ID #83255

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014703 Estimated ship to us date changed to / /

---

Issues:83255

---

- Back Order Report

**POR11027 0014703 //**

Issue ID #83246

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014703 Estimated ship to us date changed to / /

---

Issues:83246

---

- Back Order Report

**POR11027 0014703 //**

Issue ID #82803

Date Created 03/Nov/2016

Issued To: Jonathan Connor

Completed Status: 03 Nov 2016

---

Notes:

POR11027 0014703 Estimated ship to us date changed to / /

---

Issues:82803

---

- Back Order Report

**POR11027 0014703 12/12/2016**

Issue ID #83366

Date Created 11/Nov/2016

Issued To: Jonathan Connor

Completed Status: 11 Nov 2016

---

Notes:

POR11027 0014703 Estimated ship to us date changed to 12 / 12 / 2016

---

Issues:83366

---

- Back Order Report

**POR11027 0014751 //**

Issue ID #83256

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014751 Estimated ship to us date changed to / /

---

Issues:83256

---

- Back Order Report

**POR11027 0014751 //**

Issue ID #83247

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014751 Estimated ship to us date changed to / /

---

Issues:83247

---

- Back Order Report

**POR11027 0014751 //**

Issue ID #82804

Date Created 03/Nov/2016

Issued To: Jonathan Connor

Completed Status: 03 Nov 2016

---

Notes:

POR11027 0014751 Estimated ship to us date changed to / /

---

Issues:82804

---

- Back Order Report

**POR11027 0014751 10/11/2016**

Issue ID #83367

Date Created 11/Nov/2016

Issued To: Jonathan Connor

Completed Status: 11 Nov 2016

---

Notes:

POR11027 0014751 Estimated ship to us date changed to 10 / 11 / 2016

---

Issues:83367

---

- Back Order Report

**POR11027 0014751 12/12/2016**

Issue ID #84985

Date Created 07/Dec/2016

Issued To: Jonathan Connor

Completed Status: 07 Dec 2016

---

Notes:

POR11027 0014751 Estimated ship to us date changed to 12 / 12 / 2016

---

Issues:84985

---

- Back Order Report

**POR11027 0014835 //**

Issue ID #83257

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014835 Estimated ship to us date changed to / /

---

Issues:83257

---

- Back Order Report

**POR11027 0014835 //**

Issue ID #83248

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014835 Estimated ship to us date changed to / /

---

Issues:83248

---

- Back Order Report

**POR11027 0014835 //**

Issue ID #82805

Date Created 03/Nov/2016

Issued To: Jonathan Connor

Completed Status: 03 Nov 2016

---

Notes:

POR11027 0014835 Estimated ship to us date changed to / /

---

Issues:82805

---

- Back Order Report

**POR11027 0014835 12/12/2016**

Issue ID #83368

Date Created 11/Nov/2016

Issued To: Jonathan Connor

Completed Status: 11 Nov 2016

---

Notes:

POR11027 0014835 Estimated ship to us date changed to 12 / 12 / 2016

---

Issues:83368

---

- Back Order Report

**POR11027 0014996 //**

Issue ID #83258

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014996 Estimated ship to us date changed to / /

---

Issues:83258

---

- Back Order Report

**POR11027 0014996 //**

Issue ID #83249

Date Created 10/Nov/2016

Issued To: Jonathan Connor

Completed Status: 10 Nov 2016

---

Notes:

POR11027 0014996 Estimated ship to us date changed to / /

---

Issues:83249

---

- Back Order Report

**POR11027 0014996 //**

Issue ID #82806

Date Created 03/Nov/2016

Issued To: Jonathan Connor

Completed Status: 03 Nov 2016

---

Notes:

POR11027 0014996 Estimated ship to us date changed to / /

---

Issues:82806

---

- Back Order Report

**POR11027 0014996 12/12/2016**

Issue ID #83369

Date Created 11/Nov/2016

Issued To: Jonathan Connor

Completed Status: 11 Nov 2016

---

Notes:

POR11027 0014996 Estimated ship to us date changed to 12 / 12 / 2016

---

Issues:83369

---

- Back Order Report

**POR11052 0014650 //**

Issue ID #83845

Date Created 18/Nov/2016

Issued To: Jonathan Connor

Completed Status: 22 Nov 2016

---

Notes:

POR11052 0014650 Estimated ship to us date changed to / /

---

Issues:83845

---

- Back Order Report

**POR11052 0014650 06/01/2017**

Issue ID #84720

Date Created 02/Dec/2016

Issued To: Jonathan Connor

Completed Status: 02 Dec 2016

---

Notes:

POR11052 0014650 Estimated ship to us date changed to 06 / 01 / 2017

---

Issues:84720

---

- Back Order Report

**POR11052 0014704 //**

Issue ID #83846

Date Created 18/Nov/2016

Issued To: Jonathan Connor

Completed Status: 22 Nov 2016

---

Notes:

POR11052 0014704 Estimated ship to us date changed to / /

---

Issues:83846

---

- Back Order Report

**POR11052 0014704 27/01/2017**

Issue ID #84721

Date Created 02/Dec/2016

Issued To: Jonathan Connor

Completed Status: 02 Dec 2016

---

Notes:

POR11052 0014704 Estimated ship to us date changed to 27 / 01 / 2017

---

Issues:84721

---

- Back Order Report

**POR11052 0014750 //**

Issue ID #83847

Date Created 18/Nov/2016

Issued To: Jonathan Connor

Completed Status: 22 Nov 2016

---

Notes:

POR11052 0014750 Estimated ship to us date changed to / /

---

Issues:83847

---

- Back Order Report

**POR11052 0014750 06/01/2017**

Issue ID #84722

Date Created 02/Dec/2016

Issued To: Jonathan Connor

Completed Status: 02 Dec 2016

---

Notes:

POR11052 0014750 Estimated ship to us date changed to 06 / 01 / 2017

---

Issues:84722

---

- Back Order Report

**POR11109 0014704 //**

Issue ID #86829

Date Created 10/Jan/2017

Issued To: Jonathan Connor

Completed Status: 10 Jan 2017

---

Notes:

POR11109 0014704 Estimated ship to us date changed to / /

---

Issues:86829

---

- Back Order Report

**POR11109 0014704 10/03/2017**

Issue ID #87065

Date Created 13/Jan/2017

Issued To: Jonathan Connor

Completed Status: 13 Jan 2017

---

Notes:

POR11109 0014704 Estimated ship to us date changed to 10 / 03 / 2017

---

Issues:87065

---

- Back Order Report

**POR11109 0014835 //**

Issue ID #86830

Date Created 10/Jan/2017

Issued To: Jonathan Connor

Completed Status: 10 Jan 2017

---

Notes:

POR11109 0014835 Estimated ship to us date changed to / /

---

Issues:86830

---

- Back Order Report

**POR11109 0014835 10/03/2017**

Issue ID #87066

Date Created 13/Jan/2017

Issued To: Jonathan Connor

Completed Status: 13 Jan 2017

---

Notes:

POR11109 0014835 Estimated ship to us date changed to 10 / 03 / 2017

---

Issues:87066

---

- Back Order Report

**POR11130 0014601 //**

Issue ID #88833

Date Created 10/Feb/2017

Issued To: Jonathan Connor

Completed Status: 10 Feb 2017

---

Notes:

POR11130 0014601 Estimated ship to us date changed to / /

---

Issues:88833

---

- Back Order Report

**POR11130 0014601 //**

Issue ID #87781

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014601 Estimated ship to us date changed to / /

---

Issues:87781

---

- Back Order Report

**POR11130 0014601 //**

Issue ID #87773

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014601 Estimated ship to us date changed to / /

---

Issues:87773

---

- Back Order Report

**POR11130 0014601 //**

Issue ID #87425

Date Created 19/Jan/2017

Issued To: Jonathan Connor

Completed Status: 19 Jan 2017

---

Notes:

POR11130 0014601 Estimated ship to us date changed to / /

---

Issues:87425

---

- Back Order Report

**POR11130 0014601 24/03/2017**

Issue ID #89674

Date Created 20/Feb/2017

Issued To: Jonathan Connor

Completed Status: 20 Feb 2017

---

Notes:

POR11130 0014601 Estimated ship to us date changed to 24 / 03 / 2017

---

Issues:89674

---

- Back Order Report

**POR11130 0014905 //**

Issue ID #88834

Date Created 10/Feb/2017

Issued To: Jonathan Connor

Completed Status: 10 Feb 2017

---

Notes:

POR11130 0014905 Estimated ship to us date changed to / /

---

Issues:88834

---

- Back Order Report

**POR11130 0014905 //**

Issue ID #87782

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014905 Estimated ship to us date changed to / /

---

Issues:87782

---

- Back Order Report

**POR11130 0014905 //**

Issue ID #87774

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014905 Estimated ship to us date changed to / /

---

Issues:87774

---

- Back Order Report

**POR11130 0014905 //**

Issue ID #87426

Date Created 19/Jan/2017

Issued To: Jonathan Connor

Completed Status: 19 Jan 2017

---

Notes:

POR11130 0014905 Estimated ship to us date changed to / /

---

Issues:87426

---

- Back Order Report

**POR11130 0014905 13/04/2017**

Issue ID #89675

Date Created 20/Feb/2017

Issued To: Jonathan Connor

Completed Status: 20 Feb 2017

---

Notes:

POR11130 0014905 Estimated ship to us date changed to 13 / 04 / 2017

---

Issues:89675

---

- Back Order Report

**POR11130 0014912 //**

Issue ID #88835

Date Created 10/Feb/2017

Issued To: Jonathan Connor

Completed Status: 10 Feb 2017

---

Notes:

POR11130 0014912 Estimated ship to us date changed to / /

---

Issues:88835

---

- Back Order Report

**POR11130 0014912 //**

Issue ID #87783

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014912 Estimated ship to us date changed to / /

---

Issues:87783

---

- Back Order Report

**POR11130 0014912 //**

Issue ID #87775

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014912 Estimated ship to us date changed to / /

---

Issues:87775

---

- Back Order Report

**POR11130 0014912 //**

Issue ID #87427

Date Created 19/Jan/2017

Issued To: Jonathan Connor

Completed Status: 19 Jan 2017

---

Notes:

POR11130 0014912 Estimated ship to us date changed to / /

---

Issues:87427

---

- Back Order Report

**POR11130 0014912 24/03/2017**

Issue ID #89676

Date Created 20/Feb/2017

Issued To: Jonathan Connor

Completed Status: 20 Feb 2017

---

Notes:

POR11130 0014912 Estimated ship to us date changed to 24 / 03 / 2017

---

Issues:89676

---

- Back Order Report

**POR11130 0014960 //**

Issue ID #88836

Date Created 10/Feb/2017

Issued To: Jonathan Connor

Completed Status: 10 Feb 2017

---

Notes:

POR11130 0014960 Estimated ship to us date changed to / /

---

Issues:88836

---

- Back Order Report

**POR11130 0014960 //**

Issue ID #87784

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014960 Estimated ship to us date changed to / /

---

Issues:87784

---

- Back Order Report

**POR11130 0014960 //**

Issue ID #87776

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014960 Estimated ship to us date changed to / /

---

Issues:87776

---

- Back Order Report

**POR11130 0014960 //**

Issue ID #87428

Date Created 19/Jan/2017

Issued To: Jonathan Connor

Completed Status: 19 Jan 2017

---

Notes:

POR11130 0014960 Estimated ship to us date changed to / /

---

Issues:87428

---

- Back Order Report

**POR11130 0014960 13/04/2017**

Issue ID #89677

Date Created 20/Feb/2017

Issued To: Jonathan Connor

Completed Status: 20 Feb 2017

---

Notes:

POR11130 0014960 Estimated ship to us date changed to 13 / 04 / 2017

---

Issues:89677

---

- Back Order Report

**POR11130 0014983 //**

Issue ID #88838

Date Created 10/Feb/2017

Issued To: Katie Evans

Completed Status: 10 Feb 2017

---

Notes:

POR11130 0014983 Estimated ship to us date changed to / /

---

Issues:88838

---

- Back Order Report

**POR11130 0014983 //**

Issue ID #88837

Date Created 10/Feb/2017

Issued To: Jonathan Connor

Completed Status: 10 Feb 2017

---

Notes:

POR11130 0014983 Estimated ship to us date changed to / /

---

Issues:88837

---

- Back Order Report

**POR11130 0014983 //**

Issue ID #87785

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014983 Estimated ship to us date changed to / /

---

Issues:87785

---

- Back Order Report

**POR11130 0014983 //**

Issue ID #87777

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014983 Estimated ship to us date changed to / /

---

Issues:87777

---

- Back Order Report

**POR11130 0014983 //**

Issue ID #87429

Date Created 19/Jan/2017

Issued To: Jonathan Connor

Completed Status: 19 Jan 2017

---

Notes:

POR11130 0014983 Estimated ship to us date changed to / /

---

Issues:87429

---

- Back Order Report

**POR11130 0014983 13/04/2017**

Issue ID #89679

Date Created 20/Feb/2017

Issued To: Katie Evans

Completed Status: 20 Feb 2017

---

Notes:

POR11130 0014983 Estimated ship to us date changed to 13 / 04 / 2017

---

Issues:89679

---

- Back Order Report

**POR11130 0014983 13/04/2017**

Issue ID #89678

Date Created 20/Feb/2017

Issued To: Jonathan Connor

Completed Status: 20 Feb 2017

---

Notes:

POR11130 0014983 Estimated ship to us date changed to 13 / 04 / 2017

---

Issues:89678

---

Purchase Back Orders Review - Purchase Back Orders Review

**POR11130 0014983 Have Been Delivered or Cancelled in Opera**

Issue ID #93228

Date Created 21/Apr/2017

Issued To: Katie Evans

Completed Status: 24 Apr 2017

---

Notes:

Have Been Delivered or Cancelled in Opera

---

Issues:93228

---

- Back Order Report

**POR11130 0014990 //**

Issue ID #88839

Date Created 10/Feb/2017

Issued To: Jonathan Connor

Completed Status: 10 Feb 2017

---

Notes:

POR11130 0014990 Estimated ship to us date changed to / /

---

Issues:88839

---

- Back Order Report

**POR11130 0014990 //**

Issue ID #87786

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014990 Estimated ship to us date changed to / /

---

Issues:87786

---

- Back Order Report

**POR11130 0014990 //**

Issue ID #87778

Date Created 25/Jan/2017

Issued To: Jonathan Connor

Completed Status: 25 Jan 2017

---

Notes:

POR11130 0014990 Estimated ship to us date changed to / /

---

Issues:87778

---

- Back Order Report

**POR11130 0014990 //**

Issue ID #87430

Date Created 19/Jan/2017

Issued To: Jonathan Connor

Completed Status: 19 Jan 2017

---

Notes:

POR11130 0014990 Estimated ship to us date changed to / /

---

Issues:87430

---

- Back Order Report

**POR11130 0014990 13/04/2017**

Issue ID #89680

Date Created 20/Feb/2017

Issued To: Jonathan Connor

Completed Status: 20 Feb 2017

---

Notes:

POR11130 0014990 Estimated ship to us date changed to 13 / 04 / 2017

---

Issues:89680

---

- Back Order Report

**POR11189 0014835 //**

Issue ID #91046

Date Created 15/Mar/2017

Issued To: Jonathan Connor

Completed Status: 15 Mar 2017

---

Notes:

POR11189 0014835 Estimated ship to us date changed to / /

---

Issues:91046

---

- Back Order Report

**POR11189 0014835 //**

Issue ID #90678

Date Created 08/Mar/2017

Issued To: Jonathan Connor

Completed Status: 09 Mar 2017

---

Notes:

POR11189 0014835 Estimated ship to us date changed to / /

---

Issues:90678

---

- Back Order Report

**POR11215 0014905 //**

Issue ID #91519

Date Created 23/Mar/2017

Issued To: Jonathan Connor

Completed Status: 23 Mar 2017

---

Notes:

POR11215 0014905 Estimated ship to us date changed to / /

---

Issues:91519

---

- Back Order Report

**POR11215 0014905 12/05/2017**

Issue ID #91761

Date Created 28/Mar/2017

Issued To: Jonathan Connor

Completed Status: 30 Mar 2017

---

Notes:

POR11215 0014905 Estimated ship to us date changed to 12 / 05 / 2017

---

Issues:91761

---

- Back Order Report

**POR11215 0014909 //**

Issue ID #91520

Date Created 23/Mar/2017

Issued To: Jonathan Connor

Completed Status: 23 Mar 2017

---

Notes:

POR11215 0014909 Estimated ship to us date changed to / /

---

Issues:91520

---

- Back Order Report

**POR11215 0014909 12/05/2017**

Issue ID #91762

Date Created 28/Mar/2017

Issued To: Jonathan Connor

Completed Status: 30 Mar 2017

---

Notes:

POR11215 0014909 Estimated ship to us date changed to 12 / 05 / 2017

---

Issues:91762

---

- Back Order Report

**POR11215 0014982 //**

Issue ID #91521

Date Created 23/Mar/2017

Issued To: Jonathan Connor

Completed Status: 23 Mar 2017

---

Notes:

POR11215 0014982 Estimated ship to us date changed to / /

---

Issues:91521

---

- Back Order Report

**POR11215 0014982 12/05/2017**

Issue ID #91763

Date Created 28/Mar/2017

Issued To: Jonathan Connor

Completed Status: 30 Mar 2017

---

Notes:

POR11215 0014982 Estimated ship to us date changed to 12 / 05 / 2017

---

Issues:91763

---

- Back Order Report

**POR11215 0014990 //**

Issue ID #91522

Date Created 23/Mar/2017

Issued To: Jonathan Connor

Completed Status: 23 Mar 2017

---

Notes:

POR11215 0014990 Estimated ship to us date changed to / /

---

Issues:91522

---

- Back Order Report

**POR11215 0014990 12/05/2017**

Issue ID #91764

Date Created 28/Mar/2017

Issued To: Jonathan Connor

Completed Status: 30 Mar 2017

---

Notes:

POR11215 0014990 Estimated ship to us date changed to 12 / 05 / 2017

---

Issues:91764

---

- Back Order Report

**POR11215 0014994 //**

Issue ID #91523

Date Created 23/Mar/2017

Issued To: Jonathan Connor

Completed Status: 23 Mar 2017

---

Notes:

POR11215 0014994 Estimated ship to us date changed to / /

---

Issues:91523

---

- Back Order Report

**POR11215 0014994 12/05/2017**

Issue ID #91765

Date Created 28/Mar/2017

Issued To: Jonathan Connor

Completed Status: 30 Mar 2017

---

Notes:

POR11215 0014994 Estimated ship to us date changed to 12 / 05 / 2017

---

Issues:91765

---

- Back Order Report

**POR11215 0014996 //**

Issue ID #91524

Date Created 23/Mar/2017

Issued To: Jonathan Connor

Completed Status: 23 Mar 2017

---

Notes:

POR11215 0014996 Estimated ship to us date changed to / /

---

Issues:91524

---

- Back Order Report

**POR11215 0014996 12/05/2017**

Issue ID #91766

Date Created 28/Mar/2017

Issued To: Jonathan Connor

Completed Status: 30 Mar 2017

---

Notes:

POR11215 0014996 Estimated ship to us date changed to 12 / 05 / 2017

---

Issues:91766

---

- Back Order Report

**POR11321 0014904 //**

Issue ID #97739

Date Created 29/Jun/2017

Issued To: Jonathan Connor

Completed Status: 29 Jun 2017

---

Notes:

POR11321 0014904 Estimated ship to us date changed to / /

---

Issues:97739

---

- Back Order Report

**POR11321 0014904 //**

Issue ID #97220

Date Created 22/Jun/2017

Issued To: Jonathan Connor

Completed Status: 22 Jun 2017

---

Notes:

POR11321 0014904 Estimated ship to us date changed to / /

---

Issues:97220

---

- Back Order Report

**POR11321 0014912 //**

Issue ID #97740

Date Created 29/Jun/2017

Issued To: Jonathan Connor

Completed Status: 29 Jun 2017

---

Notes:

POR11321 0014912 Estimated ship to us date changed to / /

---

Issues:97740

---

- Back Order Report

**POR11321 0014912 //**

Issue ID #97221

Date Created 22/Jun/2017

Issued To: Jonathan Connor

Completed Status: 22 Jun 2017

---

Notes:

POR11321 0014912 Estimated ship to us date changed to / /

---

Issues:97221

---

- Back Order Report

**POR11322 0014650 //**

Issue ID #97734

Date Created 29/Jun/2017

Issued To: Jonathan Connor

Completed Status: 29 Jun 2017

---

Notes:

POR11322 0014650 Estimated ship to us date changed to / /

---

Issues:97734

---

- Back Order Report

**POR11322 0014650 //**

Issue ID #97215

Date Created 22/Jun/2017

Issued To: Jonathan Connor

Completed Status: 22 Jun 2017

---

Notes:

POR11322 0014650 Estimated ship to us date changed to / /

---

Issues:97215

---

- Back Order Report

**POR11322 0014703 //**

Issue ID #97735

Date Created 29/Jun/2017

Issued To: Jonathan Connor

Completed Status: 29 Jun 2017

---

Notes:

POR11322 0014703 Estimated ship to us date changed to / /

---

Issues:97735

---

- Back Order Report

**POR11322 0014703 //**

Issue ID #97216

Date Created 22/Jun/2017

Issued To: Jonathan Connor

Completed Status: 22 Jun 2017

---

Notes:

POR11322 0014703 Estimated ship to us date changed to / /

---

Issues:97216

---

- Back Order Report

**POR11322 0014703 30/08/2017**

Issue ID #99379

Date Created 25/Jul/2017

Issued To: Jonathan Connor

Completed Status: 26 Jul 2017

---

Notes:

POR11322 0014703 Estimated ship to us date changed to 30 / 08 / 2017

---

Issues:99379

---

- Back Order Report

**POR11322 0014704 //**

Issue ID #97736

Date Created 29/Jun/2017

Issued To: Jonathan Connor

Completed Status: 29 Jun 2017

---

Notes:

POR11322 0014704 Estimated ship to us date changed to / /

---

Issues:97736

---

- Back Order Report

**POR11322 0014704 //**

Issue ID #97217

Date Created 22/Jun/2017

Issued To: Jonathan Connor

Completed Status: 22 Jun 2017

---

Notes:

POR11322 0014704 Estimated ship to us date changed to / /

---

Issues:97217

---

- Back Order Report

**POR11322 0014704 30/08/2017**

Issue ID #99380

Date Created 25/Jul/2017

Issued To: Jonathan Connor

Completed Status: 26 Jul 2017

---

Notes:

POR11322 0014704 Estimated ship to us date changed to 30 / 08 / 2017

---

Issues:99380

---

- Back Order Report

**POR11322 0014750 //**

Issue ID #97737

Date Created 29/Jun/2017

Issued To: Jonathan Connor

Completed Status: 29 Jun 2017

---

Notes:

POR11322 0014750 Estimated ship to us date changed to / /

---

Issues:97737

---

- Back Order Report

**POR11322 0014750 //**

Issue ID #97218

Date Created 22/Jun/2017

Issued To: Jonathan Connor

Completed Status: 22 Jun 2017

---

Notes:

POR11322 0014750 Estimated ship to us date changed to / /

---

Issues:97218

---

- Back Order Report

**POR11322 0014750 30/08/2017**

Issue ID #99381

Date Created 25/Jul/2017

Issued To: Jonathan Connor

Completed Status: 26 Jul 2017

---

Notes:

POR11322 0014750 Estimated ship to us date changed to 30 / 08 / 2017

---

Issues:99381

---

- Back Order Report

**POR11322 0014825 //**

Issue ID #97738

Date Created 29/Jun/2017

Issued To: Jonathan Connor

Completed Status: 29 Jun 2017

---

Notes:

POR11322 0014825 Estimated ship to us date changed to / /

---

Issues:97738

---

- Back Order Report

**POR11322 0014825 //**

Issue ID #97219

Date Created 22/Jun/2017

Issued To: Jonathan Connor

Completed Status: 22 Jun 2017

---

Notes:

POR11322 0014825 Estimated ship to us date changed to / /

---

Issues:97219

---

- Back Order Report

**POR11342 0014750 //**

Issue ID #100700

Date Created 14/Aug/2017

Issued To: Jonathan Connor

Completed Status: 14 Aug 2017

---

Notes:

POR11342 0014750 Estimated ship to us date changed to / /

---

Issues:100700

---

- Back Order Report

**POR11342 0014750 //**

Issue ID #99288

Date Created 24/Jul/2017

Issued To: Jonathan Connor

Completed Status: 26 Jul 2017

---

Notes:

POR11342 0014750 Estimated ship to us date changed to / /

---

Issues:99288

---

- Back Order Report

**POR11342 0014750 //**

Issue ID #98639

Date Created 12/Jul/2017

Issued To: Jonathan Connor

Completed Status: 12 Jul 2017

---

Notes:

POR11342 0014750 Estimated ship to us date changed to / /

---

Issues:98639

---

- Back Order Report

**POR11342 0014895 //**

Issue ID #100701

Date Created 14/Aug/2017

Issued To: Jonathan Connor

Completed Status: 14 Aug 2017

---

Notes:

POR11342 0014895 Estimated ship to us date changed to / /

---

Issues:100701

---

- Back Order Report

**POR11342 0014895 //**

Issue ID #99289

Date Created 24/Jul/2017

Issued To: Jonathan Connor

Completed Status: 26 Jul 2017

---

Notes:

POR11342 0014895 Estimated ship to us date changed to / /

---

Issues:99289

---

- Back Order Report

**POR11342 0014895 //**

Issue ID #98640

Date Created 12/Jul/2017

Issued To: Jonathan Connor

Completed Status: 12 Jul 2017

---

Notes:

POR11342 0014895 Estimated ship to us date changed to / /

---

Issues:98640

---

- Back Order Report

**POR11389 0014983 //**

Issue ID #101300

Date Created 22/Aug/2017

Issued To: Jonathan Connor

Completed Status: 24 Aug 2017

---

Notes:

POR11389 0014983 Estimated ship to us date changed to / /

---

Issues:101300

---

- Back Order Report

**POR11389 0014983 //**

Issue ID #101245

Date Created 21/Aug/2017

Issued To: Jonathan Connor

Completed Status: 24 Aug 2017

---

Notes:

POR11389 0014983 Estimated ship to us date changed to / /

---

Issues:101245

---

- Back Order Report

**POR11389 0014983 //**

Issue ID #100697

Date Created 14/Aug/2017

Issued To: Jonathan Connor

Completed Status: 14 Aug 2017

---

Notes:

POR11389 0014983 Estimated ship to us date changed to / /

---

Issues:100697

---

- Back Order Report

**POR11432 0014835 //**

Issue ID #104659

Date Created 05/Oct/2017

Issued To: Jonathan Connor

Completed Status: 05 Oct 2017

---

Notes:

POR11432 0014835 Estimated ship to us date changed to / /

---

Issues:104659

---

- Back Order Report

**POR11432 0014835 //**

Issue ID #102897

Date Created 14/Sep/2017

Issued To: Jonathan Connor

Completed Status: 14 Sep 2017

---

Notes:

POR11432 0014835 Estimated ship to us date changed to / /

---

Issues:102897

---

- Back Order Report

**POR11432 0014835 07/11/2017**

Issue ID #104871

Date Created 09/Oct/2017

Issued To: Jonathan Connor

Completed Status: 11 Oct 2017

---

Notes:

POR11432 0014835 Estimated ship to us date changed to 07 / 11 / 2017

---

Issues:104871

---

- Back Order Report

**POR11444 0014750 //**

Issue ID #104870

Date Created 09/Oct/2017

Issued To: Jonathan Connor

Completed Status: 11 Oct 2017

---

Notes:

POR11444 0014750 Estimated ship to us date changed to / /

---

Issues:104870

---

- Back Order Report

**POR11444 0014750 //**

Issue ID #104658

Date Created 05/Oct/2017

Issued To: Jonathan Connor

Completed Status: 05 Oct 2017

---

Notes:

POR11444 0014750 Estimated ship to us date changed to / /

---

Issues:104658

---

- Back Order Report

**POR11444 0014750 //**

Issue ID #103704

Date Created 25/Sep/2017

Issued To: Jonathan Connor

Completed Status: 25 Sep 2017

---

Notes:

POR11444 0014750 Estimated ship to us date changed to / /

---

Issues:103704

---

- Back Order Report

**POR11451 0014895 //**

Issue ID #104294

Date Created 02/Oct/2017

Issued To: Jonathan Connor

Completed Status: 02 Oct 2017

---

Notes:

POR11451 0014895 Estimated ship to us date changed to / /

---

Issues:104294

---

- Back Order Report

**POR11451 0014895 //**

Issue ID #104055

Date Created 29/Sep/2017

Issued To: Jonathan Connor

Completed Status: 02 Oct 2017

---

Notes:

POR11451 0014895 Estimated ship to us date changed to / /

---

Issues:104055

---

- Back Order Report

**POR11475 0014825 //**

Issue ID #105535

Date Created 16/Oct/2017

Issued To: Jonathan Connor

Completed Status: 17 Oct 2017

---

Notes:

POR11475 0014825 Estimated ship to us date changed to / /

---

Issues:105535

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Production Statistics (551)**

Issue ID #84795

Date Created 05/Dec/2016

Issued To: Jonathan Connor

Completed Status: 19 Dec 2016

---

Notes:

System Generated

Task To be Completed

Review recent 30-40 Days completed Production runs,

Check the last 3 Completed runs.

1. Check the pick parts against the previous builds pick parts.

List any where the quantities don't match

List production jobs any where Days to complete are greater than 30 Days.

---

08 Dec 2016 Jonathan Connor

Missed scanning connectors on job

PS2242- To Build 2 Units

Used 2 of 0014714

Used 2 of 0031256

PS2206- To Build 1 Units

Used 1 of 0014714

PS2184- To Build 1 Units

Used 1 of 0014714

Used 1 of 0031256

PS2240- To Build 5 Units

Used 5 of 0014703

PS2231- To Build 5 Units

Used 5 of 0014703

Used 5 of 0031238

PS2209- To Build 4 Units

Used 4 of 0014703

Used 4 of 0031238

---

Issues:84795

---

VIAMED Stock Meeting - VIAMED Stock Meeting

**Production Statistics (551)**

Issue ID #80702

Date Created 05/Oct/2016

Issued To: Derek Lamb

Completed Status: 06 Oct 2016

---

**Notes:**

System Generated

Task To be Completed

Review recent 30-40 Days completed Production runs,

Check the last 3 Completed runs.

1. Check the pick parts against the previous builds pick parts.

List any where the quantities don't match

List production jobs any where Days to complete are greater than 30 Days.

---

05 Oct 2016 Jonathan Connor

Not matching Production jobs

PS2225- To Build 30 Units - They dont match but either sensor can be used

Used 30 of 0110214

PS2218- To Build 6 Units

Used 6 of 0110215

PS2213- To Build 1 Units

Used 1 of 0110215

PS2220- To Build 40 Units - They dont match but either sensor can be used

Used 40 of 0110132

PS2199- To Build 20 Units

Used 20 of 0110122

PS2183- To Build 50 Units

Used 50 of 0110132

PS2219- To Build 25 Units - They dont match but either sensor can be used

Used 25 of 0110132

PS2105- To Build 25 Units

Used 25 of 9711022

PS2087- To Build 6 Units

Used 6 of 0110132

PS2208- To Build 400 Units

Used 1 of 8030502

Used 3 of 8050010

PS2165- To Build 400 Units

Used 3 of 8050010

PS2130- To Build 400 Units

Used 3 of 0050040

Used 1 of 8030502

Used 3 of 8050010

PS2206- To Build 1 Units - Connector not scanned

Used 1 of 0014714

PS2184- To Build 1 Units

Used 1 of 0014714

Used 1 of 0031256

PS2129- To Build 2 Units

Used 2 of 0014714

Used 2 of 0031256

PS2205- To Build 200 Units

Used 200 of 0110122

Used 1 of 0110301      0110301 mis-scan

PS2175- To Build 200 Units

Used 200 of 0110132

PS2133- To Build 200 Units

Used 200 of 0110132

PS2220- To Build 40 Units - They dont match but either sensor can be used

Used 40 of 0110132

PS2199- To Build 20 Units

Used 20 of 0110122

PS2183- To Build 50 Units

Used 50 of 0110132

PS2196- To Build 5 Units - They dont match but either sensor can be used

Used 5 of 0110122

PS2147- To Build 6 Units

Used 6 of 0110132

PS2021- To Build 10 Units

Used 10 of 0110132

PS2194 took longer than 30 days - Middle of service visits

---

05 Oct 2016 Derek Lamb

Next Action Changed From Jonathan Connor To Derek Lamb first run thro of this task, see if next one shows any further issues

---

05 Oct 2016 Derek Lamb

---

Issues:80702

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #106710

Date Created 27/Oct/2017

Issued To: Jonathan Connor

Completed Status: Still Outstanding

---

**Notes:**

Copy Issue Main Issue #106709

Added by Steve Hardaker sent to Catherine Spence

Please scan to Sale or Returns Location ID34132 Stock Reference:

1x 4410500 - WM-2500-M

1x 0014750 - Adult silicone finger sensor

5x 4420512 - Adult/paed airway adapter

Required by end of day on Monday 30/11/17 for sales appointment on Tuesday.

Ref SOR825, Salford Royal.

---

Issues:106710

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #106709

Date Created 27/Oct/2017

Issued To: Catherine Spence

Completed Status: Still Outstanding

---

**Notes:**

Added by Steve Hardaker sent to Catherine Spence

Please scan to Sale or Returns Location ID34132 Stock Reference:

1x 4410500 - WM-2500-M

1x 0014750 - Adult silicone finger sensor

5x 4420512 - Adult/paed airway adapter

Required by end of day on Monday 30/11/17 for sales appointment on Tuesday.

Ref SOR825, Salford Royal.

---

Issues:106709

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #103215

Date Created 19/Sep/2017

Issued To: Jonathan Connor

Completed Status: 20 Sep 2017

---

**Notes:**

Added by Katie Evans sent to Jonathan Connor

Please scan to Sale or Returns Location ID34132 Stock Reference: 1 x 4410520, 1 x 0014750 and 1 x 4421025

---

20 Sep 2017 Jonathan Connor  
scanned to SOR

---

20 Sep 2017 Katie Evans  
Thank you

---

Issues:103215

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #96392

Date Created 09/Jun/2017

Issued To: Jonathan Connor

Completed Status: 12 Jun 2017

---

**Notes:**

Added by Steve Hardaker sent to Jonathan Connor

Please scan to Sale or Returns Location ID34132 Stock Reference:

1x 4410520 - VM-2500-S

with

1x 0014835 - Wrap sensor

5x 4421021 - Adult/paed airway adapter sampling line, single.

I need this on Monday 12/6 as I'm out on Tuesday, thanks. SOR794

---

12 Jun 2017 Jonathan Connor

4410520= 1043513

0014835 = 1075336

4421021 =

1070819

1070833

1070820

1070834

1070832

---

12 Jun 2017 Steve Hardaker

Taken in person and left for trial, SOR794.

---

Issues:96392

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #84633

Date Created 01/Dec/2016

Issued To: Jonathan Connor

Completed Status: 01 Dec 2016

---

Notes:

Added by Katie Evans sent to Jonathan Connor

Please scan to Sale or Returns Location ID34132 Stock Reference: 1 x 4410500, 1 x 0014750 and 3 x 4420512

---

01 Dec 2016 Jonathan Connor  
scanned

965559  
965572  
965571  
994857  
908438

---

01 Dec 2016 Katie Evans  
Completed

---

Issues:84633

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #80567

Date Created 03/Oct/2016

Issued To: Jonathan Connor

Completed Status: 21 Oct 2016

---

**Notes:**

Added by Katie Evans sent to Catherine Spence

Please scan to Sale or Returns Location ID34132 Stock Reference: 4410520 x 1. 4421021 x 2. 4421023 x 2. 4421001 x 2. 4421007 x 2 please

---

04 Oct 2016 Katie Evans and 1 x 0014750 as standard please

---

04 Oct 2016 Catherine Spence  
part number 4421007 is not available, no parts available

---

04 Oct 2016 Katie Evans  
no probs, please scrap that part, still require the rest please :)

---

04 Oct 2016 Katie Evans  
Please send 4 x 4421001

---

21 Oct 2016 Katie Evans

Next Action Changed From Catherine Spence To Jonathan Connor Please scan

---

21 Oct 2016 Jonathan Connor  
Complete

---

21 Oct 2016 Katie Evans

---

Issues:80567

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #78892

Date Created 12/Sep/2016

Issued To: Jonathan Connor

Completed Status: 13 Sep 2016

---

**Notes:**

Added by Steve Hardaker sent to Jonathan Connor

Please scan to Sale or Returns Location ID34132 Stock Reference:

1x 4410520 VM-2500-S

1x 0014750 Silicone adult sensor

5x 4421021 Versastream adult/paed airway adapter sampling lines.

If you can't spare a new one, a demo one would suffice. Could do with this asap to send out on Tue to EMRTS Wales to support a bid for 15 capnographs, thanks.

---

13 Sep 2016 Steve Hardaker

Forgot to add a case:

4420621

---

13 Sep 2016 Jonathan Connor

all scanned

666460

978622

950035

997287

997284

997285

997280

997283

---

13 Sep 2016 Steve Hardaker

Urgent Flag Changed To Off Sent on SOR748 13/9/16.

---

Issues:78892

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #78396

Date Created 05/Sep/2016

Issued To: Catherine Spence

Completed Status: 06 Sep 2016

---

**Notes:**

Added by Katie Evans sent to Catherine Spence

Please scan to Sale or Returns Location ID34132 Stock Reference: 0014750 x 1 and 4421026 x 1

---

06 Sep 2016 Catherine Spence  
id 944179,997257

---

06 Sep 2016 Katie Evans  
Please could you scan an additional 4 x 4421026 to SOR shelf?

---

06 Sep 2016 Catherine Spence  
id 997280,997259,997258,997256

---

06 Sep 2016 Katie Evans  
Thank you

---

Issues:78396

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #74840

Date Created 12/Jul/2016

Issued To: Catherine Spence

Completed Status: 18 Jul 2016

---

Notes:

Added by Katie Evans sent to Catherine Spence

Please scan to Sale or Returns Location ID34132 Stock Reference: 1 x 0012160 and 1 x 0014750

---

14 Jul 2016 Catherine Spence  
ids944161,947153

---

18 Jul 2016 Katie Evans  
Thank you

---

Issues:74840

---

VIAMED Stock Meeting - Sale or Returns  
**Request Scanning Item to Sale or return**  
Issue ID #74399  
Date Created 06/Jul/2016  
Issued To: Catherine Spence  
Completed Status: 06 Jul 2016

---

Notes:

Added by Steve Hardaker sent to Catherine Spence  
Please scan to Sale or Returns Location ID34132 Stock Reference:

2x 4410520 - VM2500-S with  
2x 0014750 - Silicone adult finger sensor  
10 4421021 - Airway adapter sampling lines  
1x 4420620 - Carrying case  
1x 4420621 - Carrying case with clear panel

For 2 separate SORs with 5 sampling lines each.

1 VM-2500-S with sensor, 5 sampling lines and the standard carrying case is required today for a sales visit tomorrow, thanks.

My ref: Scunthorpe General and Great North Air Ambulance.

---

06 Jul 2016 Catherine Spence  
id: 944154,944155,978620,9786198,975130  
975130  
66487,584501, 806330-806333,874097,874100,874103,874107-9

---

06 Jul 2016 Steve Hardaker  
Scanned to SOR733 and SOR734

---

Issues:74399

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #70813

Date Created 03/May/2016

Issued To: Catherine Spence

Completed Status: 05 May 2016

---

**Notes:**

Added by Katie Evans sent to Catherine Spence

Please scan to Sale or Returns Location ID34132 Stock Reference: 1 x 4410500, 1 x 0014750 and 2 x 420512

---

03 May 2016 Catherine Spence

id: 985552,944143,908436

---

05 May 2016 Katie Evans

Please can we return these to the sellable shelves as Medivent have changed their mind

---

Issues:70813

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #70151

Date Created 19/Apr/2016

Issued To: Catherine Spence

Completed Status: 25 Apr 2016

---

**Notes:**

Copy Issue Main Issue #70150

Added by Sarah Walton sent to Sarah Walton

Please scan to Sale or Returns Location ID34132 Stock Reference: 4410520 x1 & 0014750 x1

---

22 Apr 2016 Catherine Spence  
done

---

25 Apr 2016 Sarah Walton  
thank you

---

Issues:70151

---

VIAMED Stock Meeting - Sale or Returns

**Request Scanning Item to Sale or return**

Issue ID #70150

Date Created 19/Apr/2016

Issued To: Sarah Walton

Completed Status: 25 Apr 2016

---

**Notes:**

Added by Sarah Walton sent to Sarah Walton

Please scan to Sale or Returns Location ID34132 Stock Reference: 4410520 x1 & 0014750 x1

---

25 Apr 2016 Sarah Walton

---

25 Apr 2016 Sarah Walton

---

Issues:70150

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR457**

Issue ID #58501

Date Created 29/Apr/2015

Issued To: Steve Nixon

Completed Status: 10 Aug 2016

Linked Files

15 Jul 2015 Missing pixels image

15 Jul 2015 Missing pixels image 2

---

**Notes:**

Added by Jonathan Connor sent to Steve Hardaker

SOR ID457

Originally added by :Steve Hardaker

London Ambulance Service

Shaun Rock

Hart Base

Unit 14 Clock Tower Industrial Estate

Clock Tower Road

Isleworth

TW7 6GF

Returned : 0014750 EBN01837

Returned : 4410520 51020143

Returned : 4420620 Tracking14

Unit, carry case and sensor returned. Will book in to SRS64859 ( No notes added yet, but will update SRS as I book it in)

At first glance the screen looks to be faulty, there are pixels missing

---

29 Apr 2015 Steve Hardaker

Noted, thanks. If it can be economically repaired re-sold as ex-demo, let me know.

---

30 Apr 2015 Jonathan Connor

ok, will keep issue active till repair is complete

---

14 Jul 2015 Steve Hardaker

There are no notes on the SRS as yet, did you determine if this can be economically repaired

Also, is this an example of OLED failure and should it be sent to Bluepoint to illustrate a potential problem

If Steve N hasnt seen this, you should make him aware of it to keep him in the picture.

---

15 Jul 2015 Jonathan Connor

Updated Notes on the unit. Unit is beyond economical repair. The display is has missing pixels, looks the same as the liquid damage display from another unit. The pump does not work, sounds like there is a blockage. The unit will be returned to BPM to see if we can get any parts replaced under warranty. The main PCB is the old style so is not compatible with new units.

Images attached

---

04 Aug 2015 Steve Hardaker

Next Action Changed From Steve Hardaker To Steve Nixon Steve, see notes above. FYI in case you need

to discuss with Bluepoint.

This unit suffered an OLED display failure, so far we have 9 display failures logged in the system across various models.

---

16 Feb 2016 Steve Nixon

---

10 Aug 2016 Jonathan ConnorThanks

---

Issues:58501

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR688**

Issue ID #70814

Date Created 03/May/2016

Issued To: Jonathan Connor

Completed Status: 03 May 2016

---

Notes:

Added by Jonathan Connor sent to Jonathan Connor

SOR ID688

Originally added by :Sarah Walton

Morgate Biomedical

Patrick Le Mouellic

117 Route De Carrieres

Bat B2 Chatou

France

78400

Returned : 0014750 EWA01350

Returned : 0014750 EWF02919

Returned : 4410500 40130166

Returned : 4410550 60910166

---

03 May 2016 Jonathan Connor

---

03 May 2016 Jonathan Connor

Thanks

---

Issues:70814

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR709**

Issue ID #71961

Date Created 24/May/2016

Issued To: Ryan Swaine

Completed Status: 04 Jul 2016

---

**Notes:**

Added by Catherine Spence sent to Katie Evans

SOR ID709

Originally added by :Sarah Walton

Medicoline ApS

Lottie Persson

Vaerkstedsvej 24 B

Koege

Denmark

VAT No: 33046960

DK-4600

sold ord80401

Returned : 0014750 ESGA00516

Returned : 4410520 81020149

---

24 May 2016 Katie Evans

Next Action Changed From Katie Evans To Ryan Swaine

---

25 May 2016 Ryan Swaine

---

04 Jul 2016 Catherine Spence

---

Issues:71961

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR733**

Issue ID #82082

Date Created 21/Oct/2016

Issued To: Steve Hardaker

Completed Status: 21 Oct 2016

---

**Notes:**

Added by Jonathan Connor sent to Steve Hardaker

SOR ID733

Originally added by :Steve Hardaker

Scunthorpe General Hospital

Nick Harrison (Ward 11)

c/o Medical Engineering Department

Cliff Gardens

Scunthorpe

DN15 7BH

Returned : 0014750 ESGA00530

Returned : 4410520 A1603160003

Returned : 4420620 Tracking11

Returned : 4421021 Tracking

Returned : 4421021 Tracking1

Returned : 4421021 Tracking10

Returned : 4421021 Tracking12

Returned : 4421021 Tracking2

Returned : 4421021 Tracking2

Returned : 4421021 Tracking20

Returned : 4421021 Tracking3

Returned : 4421021 Tracking4

Returned : 4421021 Tracking5

---

21 Oct 2016 Steve Hardaker

No sale, does not meet their requirements as wanted to use with NIV and it didnt work.

---

21 Oct 2016 Jonathan Connor

Thanks

---

Issues:82082

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR734**

Issue ID #78566

Date Created 07/Sep/2016

Issued To: Steve Hardaker

Completed Status: 15 Sep 2016

---

Notes:

Added by Jonathan Connor sent to Steve Hardaker

SOR ID734

Originally added by :Steve Hardaker

Great North Air Ambulance

Colin Clark

Gate 8

Durham Tees Valley Airport

Darlington

DL2 1LU

Returned : 0014750 ESGA00529

Returned : 4410520 A1603160002

Returned : 4420621 Tracking16

Returned : 4421021 Tracking11

Returned : 4421021 Tracking14

Returned : 4421021 Tracking18

Returned : 4421021 Tracking19

Returned : 4421021 Tracking8

In repair queue to book in as test only, at first glance unit looks to work perfectly

---

07 Sep 2016 Steve Hardaker

---

15 Sep 2016 Jonathan Connor

Thanks

---

Issues:78566

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR747**

Issue ID #79896

Date Created 26/Sep/2016

Issued To: Katie Evans

Completed Status: 28 Sep 2016

---

Notes:

Added by Jonathan Connor sent to Steve Hardaker

SOR ID747

Originally added by :Katie Evans

Torbay Hospital

Mark Elliott

Medical Electronics

Lawes Bridge

Newton Rd

Torquay

TQ2 7AA

Returned : 0014750 ESGA00554

Returned : 4410520 70740155

---

26 Sep 2016 Jonathan Connor

Returned - SRS65624 - Test only

---

26 Sep 2016 Steve Hardaker

Next Action Changed From Steve Hardaker To Katie Evans Hi Katie, it looks like your Sale Or Return VM-2500-S to Torbay has come back in. Please can you follow up with the customer and add some notes into the CRM as to the outcome.

In particular, we need to know whether they intend to buy one and if not, why it was not suitable. If we can get a completed feedback form, that would be useful, thanks.

---

26 Sep 2016 Katie Evans

Completed, will chase when he returns from leave

---

28 Sep 2016 Jonathan ConnorThanks, was tagged as SH

---

Issues:79896

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR748**

Issue ID #84009

Date Created 22/Nov/2016

Issued To: Steve Hardaker

Completed Status: 24 Nov 2016

---

**Notes:**

Added by Catherine Spence sent to Steve Hardaker

SOR ID748

Originally added by :Steve Hardaker

Emergency Medical and Retrieval Transfer Service

c/o Rhyan Curtin

Ty Elusen

Ffordd Angel

Llanelli Gate

Dafen

SA14 8LQ

Returned : 0014750 ESGA00599

sold on ord83065

Returned : 4410520 A1603160005

---

22 Nov 2016 Steve Hardaker

---

24 Nov 2016 Catherine Spence

---

Issues:84009

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR758**

Issue ID #87705

Date Created 24/Jan/2017

Issued To: Katie Evans

Completed Status: 25 Jan 2017

---

Notes:

Added by Jonathan Connor sent to Katie Evans  
SOR ID758  
Originally added by :Katie Evans  
Mediq Danmark A/S  
FAO: Jacob Mark Sorensen and Kasper Larsen

Kornmarksvej 15-19

Brondby  
DK-2605

Returned : 0014750 ESGA00619  
Returned : 4410520 80940147

---

24 Jan 2017 Katie Evans  
Thank you

---

25 Jan 2017 Jonathan Connor  
np

---

Issues:87705

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR768**

Issue ID #88507

Date Created 06/Feb/2017

Issued To: Ryan Swaine

Completed Status: 07 Feb 2017

---

Notes:

Added by Jonathan Connor sent to Ryan Swaine

SOR ID768

Originally added by :Katie Evans

Jacomedic

Oivin Ek Jacobsen

Solgaard Skog 131

1599 Moss

Norway

1599

Returned : 0014750 EFG01360

Returned : 4410500 90730147

Returned : 4420512 Tracking754433\_19

Returned : 4420512 Tracking754433\_20

Returned : 4420512 Tracking754433\_7

---

06 Feb 2017 Ryan Swaine

---

07 Feb 2017 Jonathan Connor

Thanks

---

Issues:88507

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR769**

Issue ID #93816

Date Created 02/May/2017

Issued To: Ryan Swaine

Completed Status: 04 May 2017

---

**Notes:**

Added by Jonathan Connor sent to Ryan Swaine

SOR ID769

Originally added by :Jonathan Connor

Medilink Yorkshire and Humber Ltd

Peter Wheeler

3 Smithy Wood Drive

Smithy Wood Business Park

Sheffield

0114 232 9273

S35 1QN

Returned : 0012165 T90500203

Returned : 0014751 EEL01309

Returned : 0121182 Tracking22

Returned : 0121182 Tracking3

Returned : 4410500 61130142

---

03 May 2017 Ryan Swaine

---

04 May 2017 Jonathan Connor

thanks

---

Issues:93816

---

VIAMED Stock Meeting - Sale or Returns

**Returned Items SOR778**

Issue ID #92639

Date Created 11/Apr/2017

Issued To: Steve Hardaker

Completed Status: 03 May 2017

---

**Notes:**

Added by Catherine Spence sent to Steve Hardaker

SOR ID778

Originally added by :Steve Hardaker

James Paget University Hospital NHS Trust

Julie Averies

Purchasing Department

Lowestoft Road

Gorleston

Great Yarmouth

NR31 6LA

Returned : 4410520 A1610160004

sold on ord85771

---

11 Apr 2017 Steve Hardaker

1x 4410520 & 1x 0014835 wrap sensor on IN150220 11/4/17 to AGA Equipment.

---

03 May 2017 Catherine Spence

---

Issues:92639

---

VIAMED General Issue - VIAMED General Issue

**Sale or return giving**

Issue ID #57855

Date Created 01/Apr/2015

Issued To: Derek Lamb

Completed Status: 01 Jul 2016

---

**Notes:**

Added by Jonathan Connor sent to Jonathan Connor

Test, does sale or return giving add to opera adjustment list? 4421021 x 20 added to SOR for customer  
Issue Hidden on 01 Apr 2015 by Jonathan Connor - until 09/04/2015

---

09 Apr 2015 Jonathan Connor

Next Action Changed From Jonathan Connor To Derek Lamb Please can you add sale or return giving to the opera adjustments list.

---

09 Apr 2015 Derek Lamb  
really sor's been missed.

---

09 Apr 2015 Jonathan Connoryes, only just started testing it the other week when I noticed some things were a couple out on capnograph accessories.

---

05 Aug 2015 Derek Lamb  
Done the code,  
going forwards new items will be listed correctly for opera correction,  
suggest the following Items are Stock checked as they have been given away in the past 2 years

Stock Reference	Description	Quantity Used
0014602	SpO2 Finger Sensor - Adult.	2
0014650	SpO2 Finger Sensor - Adult.	1
0021011	Sensor wrap single. Sample pack.	1
0021013	Sensor wraps box of 12.	8
0021016	Sensor wrap single. Sample pack.	1
0034996	Disposable SpO2 Sensor - Neonatal.	6
1134011	Sample Pack. EyeMax 25	
1134014	NeoMask Sample Pack	1
2030005	DryMax Neo 330mm 280mm. Sample pack.	1
2520000	Microstim DB3 - Patient Lead.	1
2710150	Gel Head Pillow with Centring Dish.	1
3210011	NeoPEEP neonatal single patient use	2
3830061	Posey Foam Trach Tie - Sample Pack	4
4330002	Flowsensor SpiroTrue A (PC).	5
4330003	Flowsensor SpiroTrue H.	4
4420525	Nomo adapter for use with VM-2500-S3	
4420527	Nomoline PhaseIn.	10
4420532	Sidestream disposable airway adapter	22
4420551	Disposable sampling line.	12
4420561	Disposable nasal sampling line.	2
4420780	CO2 Nasal/Oral Sampling Cannula	2
4421021	VersaStream - Airway Adapter Sampling	29
4421051	VersaStream - Nasal Sampling Line	1
4421071	VersaStream - Airway Adapter Sampling	1
4421154	VersaStream - Nasal Sampling Line with	1
4421165	VersaStream - Oral/Nasal Sampling Line	1
4421171	VersaStream - Airway Adapter Sampling	1

4421173VersaStream - Airway Adapter Sampling1  
4421200VersaStream - Sample Pack Viamed Adult5  
4421210VersaStream - Sample Pack LLM Adult1  
4421220VersaStream - Sample Pack Respironics1  
4421221VersaStream - Sample Pack Respironics1

---

05 Aug 2015 Derek Lamb

---

Issues:57855

---

VIAMED Sales And Marketing - Any other business  
**Sensormedics SpO2 sensors are discontinued**

Issue ID #96779

Date Created 15/Jun/2017

Issued To: Steve Hardaker

Completed Status: 25 Jul 2017

---

Notes:

Copy Issue Main Issue #96778

Added by Steve Nixon sent to Catrin Hird

Sensormedics SpO2 sensors (800 & 4000 Series) are discontinued, if applicable please remove from price lists, leaflets, web site. Should all show has deleted after the next update.

0014220

0014221

0014620

0014720

0014780

0014843

0018650

---

15 Jun 2017 Steve Hardaker

I've tidied the GHX stock group.

Steve: you may also want to add to the Opera stock memos to say discontinued.

---

16 Jun 2017 Steve Nixon

Thanks, I've done so for all of them.

---

16 Jun 2017 Steve Hardaker

---

Issues:96779

---

VIAMED Sales And Marketing - Any other business  
**Sensormedics SpO2 sensors are discontinued**

Issue ID #96778

Date Created 15/Jun/2017

Issued To: Catrin Hird

Completed Status: 25 Jul 2017

---

Notes:

Added by Steve Nixon sent to Catrin Hird

Sensormedics SpO2 sensors (800 & 4000 Series) are discontinued, if applicable please remove from price lists, leaflets, web site. Should all show has deleted after the next update.

0014220

0014221

0014620

0014720

0014780

0014843

0018650

---

24 Jul 2017 Catrin Hird

0014220 - not on price list or website

0014221 - not on price list or website

0014620 - removed from website, removed from price lists, saved and reprinted to Z.

0014720 - removed from website, removed from price lists, saved and reprinted to Z.

0014780 - removed from website, removed from price lists, saved and reprinted to Z.

0014843 - removed from website, removed from price lists, saved and reprinted to Z.

0018650 - No longer have price list for 800 series as all discontinued I thought. Not on website, no leaflets live for 800 series.

---

Issues:96778

---

VIAMED Stock Meeting - Sale or Returns  
**SMC Northern Ireland exhibition stock - Urgent**  
Issue ID #103719  
Date Created 25/Sep/2017  
Issued To: Jonathan Connor  
Completed Status: 29 Sep 2017

---

Notes:

Added by Steve Hardaker sent to Jonathan Connor  
Richard Quigley at SMC in Northern Ireland has requested some demo stock to allow them to do an exhibition on Saturday 30th.

I have prepared a box of leaflets, brochures and pens, and have pulled the following items out of the ESS exhibition stock for you to send:

1x 4410520 VM-2500-S  
1x 0012160 VM-2160  
1x 0014750 Adult sensor  
1x 0014751 Paed sensor  
1x 0014835 Wrap sensor  
1x 2810006 MD-300C52  
1x 4430001 Flocap sample pack

Please can you also add:

1x 0310030 TT480 (demo unit is fine) complete with a Neopeep circuit with mask and reusable adapter (I have put my non-stock demo one in).

Please can these go today, sorry it's not more notice but I only got this request on Friday as I was away last week when he called.

I have created SOR817 to track it all. Thanks.

---

25 Sep 2017 Jonathan Connor  
Done

---

29 Sep 2017 Steve Hardaker  
Confirmed receipt and discussed at length over the phone with Richard and Karen prior to their exhibition.

---

Issues:103719

---

VIAMED Sales And Marketing - Sales Viamed Medical Export  
**Smiths Medical - BCI compatible Paediatric Finger Sensor**

Issue ID #87719

Date Created 24/Jan/2017

Issued To: Derek Lamb

Completed Status: 14 Feb 2017

---

**Notes:**

Copy Issue Main Issue #87718

Added by Ryan Swaine sent to Steve Nixon

Further to my meeting with Smiths Medical, I have been asked to quote for the BCI compatible paediatric finger sensor, part no. 0014769. They are looking for a quantity of 1000/year.

Please can you let me know if we can offer a special price or should we run with distributor pricing?

---

03 Feb 2017 Derek Lamb

---

Issues:87719

---

VIAMED Sales And Marketing - Sales Viamed Medical Export  
**Smiths Medical - BCI compatible Paediatric Finger Sensor**

Issue ID #87718

Date Created 24/Jan/2017

Issued To: Steve Nixon

Completed Status: 23 Aug 2017

---

**Notes:**

Added by Ryan Swaine sent to Steve Nixon

Further to my meeting with Smiths Medical, I have been asked to quote for the BCI compatible paediatric finger sensor, part no. 0014769. They are looking for a quantity of 1000/year.

Please can you let me know if we can offer a special price or should we run with distributor pricing?

---

24 Jan 2017 Ryan Swaine

Rainer @ Smiths has also confirmed their interest in the OEM Flow Sensors. I have informed him that we are a few weeks away from having something to present, but I will try and introduce you at Arab Health and you will be available to visit them in the UK.

I have a meeting booked offsite on Tuesday @ 1400 to present the sensor and the Capnograph. Please let me know if you would like to come with me to meet them and speak to them briefly about the flow sensors?

---

23 Aug 2017 Ryan Swaine

---

Issues:87718

---

VIAMED Stock Meeting - Sale or Returns

**SOR - Gulf & World**

Issue ID #98567

Date Created 11/Jul/2017

Issued To: Derek Lamb

Completed Status: 12 Jul 2017

---

Notes:

Added by Ryan Swaine sent to Derek Lamb

After chasing the SOR's for Gulf & World, they came back with the following information:

1224 - 0014751 - EWM02160 - 252003 - SpO2 Silicone Finger Sensor - Paediatric  
sent  
02/05/12  
Still with GWT.

1225 - 0014835 - EXC00065 - 269316 - Pulse Oximetry Wrap Sensor  
sent  
02/05/12  
Still with GWT.

1363 - 2510000 - H0000921 - 111936 - Microstim DB3 - Supramaximal Nerve  
sent  
06/03/13  
Returned

1364 - 2520000 - Tracking243 - 325898 - Microstim DB3 - Patient Lead.  
sent  
06/03/13  
Returned.

They believe that they have returned the Microstim and cable at a previous exhibition by a rep who no longer works there. Then there is two sensors that they have had a while, but are still using. Due to the age of all these products and considering that some of the items were demo equipment to start with, I would suggest that we write them off as a good will gesture. They continue to do business with us and have placed a \$6k order this week.

---

11 Jul 2017 Derek Lamb  
sorted

---

12 Jul 2017 Ryan Swaine

---

Issues:98567

---

VIAMED Stock Meeting - Sale or Returns

**SOR778 - James Paget Hospital sold to AGA Equipment**

Issue ID #92571

Date Created 10/Apr/2017

Issued To: Jonathan Connor

Completed Status: 11 Apr 2017

---

**Notes:**

Added by Steve Hardaker sent to Jonathan Connor

We put a VM-2500-S in on SOR via James Paget Hospital, SOR778 in conjunction with a company called AGA Equipment. AGA Equipment want to purchase it for the patient so we need to return it from SOR and scan it to their order.

I've given the PO to the office with a copy of this issue.

SOR778 also has 1 x 0014750 Adult silicone finger sensor, but the patient is paediatric so they want to swap it for a wrap sensor 0014835.

I brought the sensor back with me (ID994844), it has not been on a patient and can go back to stock.

The purchase order will therefore be:

1x 4410520 VM-2500-S - invoice only

1x 0014835 Wrap sensor - to ship FOC

---

11 Apr 2017 Jonathan Connor

Thanks

---

11 Apr 2017 Steve Hardaker

IN150220 11/4/17

---

Issues:92571

---

## Intrastats - Cross Reference Database Tables Updates

### SpO2 Cross Reference

Issue ID #81486

Date Created 13/Oct/2016

Issued To: Derek Lamb

Completed Status: Still Outstanding

Linked Files

25 Oct 2016 SpO2 Cross Reference Xcel

---

#### Notes:

Added by Catrin Hird sent to Derek Lamb

I have updated the dropdowns for the ones you hvae included in the admin list.

I have also updated Opera to show the superseded and discontinued ones which had been misseed.

I have noticed though that part number 0018970 shows as direct compatible for 0014623 which I had not put in the spreadsheet as a direct compatible.

---

14 Oct 2016 Derek Lamb

might have been in the database before

is it not a direct compatable?? will need to be corrected if its not

---

17 Oct 2016 Catrin Hird

It is not a direct compatible it is an alternative as it is a shorter cable length. I don't know what is says in the cross reference as I don't know how to update this (or dare to!)

---

17 Oct 2016 Derek Lamb

Next Action Changed From Derek Lamb To Steve Nixon

---

25 Oct 2016 Steve Nixon

Next Action Changed From Steve Nixon To Derek Lamb Don't have a copy of the spreadsheet. Does it say `compatibile` or `direct compatible`?

---

25 Oct 2016 Catrin Hird

Next Action Changed From Derek Lamb To Steve Nixon Yellow are direct compatibles the rest of compatibles as in the same connector but different length or different sampling type i.e. silicone or disposable.

---

25 Oct 2016 Steve Nixon

OK, may cause problems as many are slightly different lengths. In this case the lengths are larger at: 3.65m and 2.90m

---

25 Oct 2016 Catrin Hird

Next Action Changed From Steve Nixon To Derek Lamb Derek, when you do the cross reference please can you make it so that it will show what the difference is, i.e. 0018520 is 3m, compatible 0014614 is 2.95m. This information should have been filled in in the cross reference.

---

24 Oct 2017 Catrin HirdHas this been dealt with?

---

Issues:81486

---

Non Conformance Issues - Non Conformance Issues

**Stock Picked out of s/n order**

Issue ID #86641

Date Created 06/Jan/2017

Issued To: Helen Lamb

Completed Status: 20 Feb 2017

---

Notes:

Added by Robert Connor sent to Helen Lamb

I've found a couple of instances of stock in the warehouse being picked out of order. The standard viamed silicone SPO2 probe 0014750s have been picked out of serial number order. The MAX-250E and MAX-250A oxygen sensors are being picked out of serial number order in the duckets, and also out of order in that newer duckets are being picked before older ones.

---

16 Feb 2017 Helen Lamb

Priority Changed From 5 To 10

---

16 Feb 2017 Helen Lamb

This will be covered in the next Non Conformance review that will come out on the 20th Feb 17.

---

16 Feb 2017 Helen Lamb

This will be covered in the next Non Conformance review that will come out on the 20th Feb 17.

---

20 Feb 2017 Robert Connor

Ok, thanks.

---

Issues:86641

---

VIAMED Stock Meeting - Stock Queries

**Stock request**

Issue ID #78584

Date Created 07/Sep/2016

Issued To: Jonathan Connor

Completed Status: 24 Nov 2016

---

Notes:

Added by Katie Evans sent to Jonathan Connor

Medline requires 10 x 0012160 with 0014750 and 10 x 0014825 for the 17th of October.

They also require 15 x 0012160 with 0014750 and 15 x 0014825 for the 28th of October.

They have also asked for a lead time for 40 x 0012160 with 0014750 and 40 x 0014825.

Please could you ensure we have the stock and advise me on the lead time?

---

07 Sep 2016 Katie Evans

Dates moved forward by 2 working days.

The sets of 10 are now required to dispatch on the 13th of October and the sets of 15 are required to dispatch on the 26th of October

---

09 Sep 2016 Jonathan Connor

We are due stock of 00014825 any day really, so should be ok for october. I'll save them when they arrive

I'll place a top up order for 0014825 and see what dates they come back with for the 40, assume 5/6 weeks to get the rest of them in, 2160 and 0014750 we have stock

---

15 Sep 2016 Katie Evans

They have requested a quote for an additional 40 sets, please could we ensure we have enough stock incase this goes forward? Advised of a 50-60 day manufacture time as per previous POR's but as it is for a military tender, they would be needed in good time

---

04 Oct 2016 Jonathan Connor

Have this covered, 25 saved for the orders and 75 on PO should be here end of october

---

05 Oct 2016 Katie Evans

thank you

---

24 Nov 2016 Jonathan Connor

Got this sorted

---

24 Nov 2016 Katie Evans

Thank you

---

Issues:78584

---

VIAMED Sales And Marketing - Sales Viamed Medical Export

**Stock required for a meeting, will be returned as new - Deadline 19th January**

Issue ID #86810

Date Created 10/Jan/2017

Issued To: Jonathan Connor

Completed Status: 19 Jan 2017

---

Notes:

Added by Ryan Swaine sent to Jonathan Connor  
Stock required for a meeting, will be returned as new

Deadline: 19/01/17 for meeting on 20th

0014609

0014709

0014769

1 of each plaster disposable (Pref BCI, but any will do)

---

16 Jan 2017 Ryan Swaine There will be an order for Cypromedica that is also to take with me

---

19 Jan 2017 Jonathan Connor  
scanned. Not seen the order, I will check with Cathy

---

19 Jan 2017 Ryan Swaine

---

Issues:86810

---

VIAMED Stock Meeting - Stock Processing

**Stock to move, count etc**

Issue ID #71139

Date Created 10/May/2016

Issued To: Ethan Sanderson

Completed Status: 01 Jun 2016

---

**Notes:**

Added by Jonathan Connor sent to Ethan Sanderson

Please move 00149XX series to labelled shelves, start with moving 0014901 to 0014911. Once these are moved and below completed I will check where the others need to go and update issue

They need to be in use by date order and stacked in the same way as the 0014904 on the front shelf

One box of each type needs to be counted to make sure there are 24 in the box, the rest of that type need to be weighed, they should be within +/-10 grams of each other, if they are outside of that they should be counted to make sure there are 24 in a box

Also need to log the use by date of each type, there may be multiple use by dates for each type. I will need number of boxes of each of the use by dates. Some of the older ones will not be entered on the system so I will need to update them, you can either hand write these or return on the issue and I will sort it out

You will need to use either stock take or move stock in the stock book menu. Either will do the job. Both will show them as counted and moved to the correct shelf. The 0014904 that have already been moved to the front rack have been completed

---

11 May 2016 Ethan Sanderson  
All sorted.

---

01 Jun 2016 Jonathan Connor  
Thanks

---

Issues:71139

---

VIAMED Board Directors Meeting - ISO Issues

**Symbols on the 4000 Series Packaging**

Issue ID #82072

Date Created 21/Oct/2016

Issued To: Catrin Hird

Completed Status: 02 Dec 2016

---

Notes:

Added by Ryan Swaine sent to Derek Lamb

Our distributor in the Czech Republic have just undergone an audit from their governing body, SUKL.

The SUKL inspectors picked out a Viamed sensor from their stock and they have some issues with the box label:

0014605, Datex compatible SpO2 Finger Sensor, 4000 Series, SN EAE00227

1) They believe that the manufacturers name should be listed next to the manufacturer symbol, not just the address.

2) They say that the water droplet is not the harmonised symbol for humidity, so we should either change the symbol or add the humidity information in text.

Our customer, ASQA, has not received their official protocol yet, but they have said that we may hear from SUKL directly.

---

21 Oct 2016 Derek Lamb

Created Related Issue #82085

Added by Derek Lamb sent to Steve Nixon

see main issue

---

24 Oct 2016 Derek Lamb

Next Action Changed From Derek Lamb To Catrin Hird

---

24 Oct 2016 Catrin Hird

Hi Ryan, I have had a similar discussion with SN a few months ago (issue #70354) which Ralf has responded to over the weekend (see below)..I shall go back to him with this query and see what he says :)

Sorry for my late reply.

I will care about the SPO2 sensor labeling issues now.

There are 2 issues.

#1 Labels for running SPO2 sensors made by Envisen

#2 Labels for SMARTsat sensors made by LHI

Your query is regarding #1. The water drop symbol is used regarding IPx1 (Ingress Protection per IEC 60529). That was accepted by our test lab. It is different issue than humidity limitation.

We have not planned to update the #1 labels because we are focused on new SMARTsat sensors.

SMARTsat sensor labeling details are just in discussion with the sensor manufacturer.

I will send you label layout drafts for review regarding all new sensor products soon.

Thanks, best regards,

Ralf

---

02 Dec 2016 Ryan Swaine

---

Issues:82072

---

VIAMED Stock Meeting - Stock Processing

**things to do**

Issue ID #78965

Date Created 13/Sep/2016

Issued To: Callum Watson

Completed Status: 15 Sep 2016

---

Notes:

Added by Jonathan Connor sent to Callum Watson

Callum, please can you do your auto issues first. then book in the stock. there should be 40 x 0014905, 25 x 0014990, 47 x 0015000 to do first. remember lot numbers date codes etc. Ifor you aren't sure check with cathy. make sure to do the expiry date and manufacture date. the printer is a little odd now. it will pdf the labelsame and download them. click on the download at the bottom and print using zm400 as normal. any problems with printing see Derek.any problems you can give me a call on mobile 07834563165 . decorating at home so not really busy busy.

---

15 Sep 2016 Callum Watson  
Done

---

15 Sep 2016 Jonathan Connor  
thanks

---

Issues:78965

---

Marketing Job Logger - Marketing Job Logger

**VM-2160 Leaflet - Remove Microfoam**

Issue ID #84462

Date Created 29/Nov/2016

Issued To: 38

Completed Status: 12 Dec 2016

---

Notes:

0014985-0014988 remove from leaflet as superseded

---

Issues:84462

---

Marketing Job Logger - Marketing Job Logger

**VM-2160 on SOR to SMC Ltd**

Issue ID #74837

Date Created 12/Jul/2016

Issued To: 0

Completed Status: Still Outstanding

---

Notes:

Please can you send the following VM-2160 and sensor to Richard Quigley at SMC Ltd., Northern Ireland:

0012160 - ID842231 (currently with JC with issue 74834 requesting it be scanned to SOR)

0014750 - ID711777 (already scanned to pending SOR stock and in box of the above)

Please provide a quote at distributor pricing.

---

Issues:74837

---

VIAMED Sales And Marketing - Sales Viamed Medical UK

**VM-2500-S sales lead - Suffolk CH**

Issue ID #101024

Date Created 18/Aug/2017

Issued To: Steve Hardaker

Completed Status: 06 Oct 2017

---

**Notes:**

Added by Steve Nixon sent to Steve Hardaker  
julie.stone@suffolkch.nhs.uk

FYI, copy of email sent to Julie Stone at the Children`s Community Nursing Team. I have tried to follow up by phone and have left messages. Will address again next week.

Dear Julie

Please accept my apologies for the delay in getting back to you. The following is with reference to the discussion yesterday morning with Jo Rackham:

We can supply a VM-2500-S sidestream capnograph for you to evaluate on a 'sale or return' basis, this unit monitors both CO2 and SpO2. I suggest that we also supply some VersaStream long-term one-piece airway adapter sampling lines, these will be more convenient and easier for the parents to use and change. How old is the patient, we can either supply adult/paediatric (4420823) or the Infant (4420824) version? The sampling lines have built in airway adapter, bacterial filter and moisture removal. So a big benefit is that there is no water trap to maintain or replace.

Also, which is the preferred site for the pulse oximetry monitoring, the finger or toes? Please see the attached leaflets showing sensor types: finger clip, silicone finger sensors, wrap sensor or disposable sensors.

Would you require an extension cable for the pulse oximetry sensor, we have two lengths: 1.2m or 2.4m.

How would the unit be used in the home care environment; would you need a pole mount or perhaps should we supply in a rubber boot (orange) with built in stand? See 4420615 protective rubber boot, which is shown on the rear page of the leaflet for the VM-2500 Series.

To provide ease of use and flexibility, the unit can be powered by three different ways and we will supply all three options for you to evaluate:

a)Standard AA batteries.

b)Rechargeable battery.

c)Mains power pack.

Can you please let me know the delivery address to which we should send the unit, should we mark it for your attention?

---

22 Aug 2017 Steve Nixon

Julie returned my call this morning. They said that they have already processed an order and we should have received it.

Order processed via MedEquip (Ipswich/Suffolk).

Discussed sensors, I suggested a wrap sensor. Sounds like they have already ordered a pack of

VersaStream lines with airway adapter.

If any queries the, clinical contact is Sara Carnell - Complex Care Lead - Children's Community Nursing Team

sara.carnell@suffolkch.nhs.uk

Tel. 01473 275320

Suggest we include future contact details for serving, she asked if she should send back to us after say 2 years.

Also suggest we discuss the accessories to supply.

There is a quote on Opera  
A/C 00000835  
West Suffolk Hospital  
QUO04907

This looks like it is related. Need to change the SpO2 sensor.

---

22 Aug 2017 Steve Hardaker

This is related to West Suffolk Hospital, I spoke to Julie Stone from Suffolk Community Healthcare in January (see Goldmine) and raised a quotation for West Suffolk Hospital.

MedEquip4Kids are a charity that purchases equipment for paediatric use (we have previously supplied items to them) so it is clear that Julie has successfully obtained charitable funding.

As SN suggested, the wrap sensor is the most appropriate so we need to advise them accordingly when the order comes in. At the time of raising the quotation, no mention was made of the patient age but I did advise that they can swap for any reusable sensor.

SN also suggested an orange rubber boot 4420615 FOC.

---

23 Aug 2017 Steve Hardaker

Emailed Julie Stone and Sara Carnell with a copy of the Operators Manual and some training information and to advise that the item is offered at this price as supply-only and does not include on-site training.

As they are only purchasing 1 unit and are a 400 mile round trip away, I advised that on-site visits would incur travelling expenses of £765.00 for 7.5h travelling and 1h on-site and that they should consider this before purchase.

---

25 Sep 2017 Steve Hardaker

Urgent Flag Changed To Off Purchase order received 25/9/17. Agreed to supply with 0014835 wrap sensor, also offered the rubber boot FOC as agreed previously by SN, along with a box of 0014890 Viamed foam wraps.

I have deleted QUO04907 for West Suffolk Hospital Community Nursing as this is for them.

Note: this is not MedEquip4Kids as I previously guessed, I made contact and determined that its company also called Medequip, full name Medequip Assistive Technology Ltd.

I have set them up as a UK end-user for now, will discuss distribution at a later time if they want to go down that route. First order is to be by proforma.

As far as the system shows, no Sale Or Return unit was ever sent. So this can be completed off unless you have any other info or loose ends to tie up.

---

Issues:101024

---

VIAMED Stock Meeting - Stock Processing

**WYA Consignment Stock (348)**

Issue ID #102931

Date Created 15/Sep/2017

Issued To: Katie Evans

Completed Status: 16 Sep 2017

---

Notes:

System Generated

Please check all is well with the Yorkshire ambulance consignment Stock. If the target doesnt match the total, forward this issue to SJH or DL if SJH unavailable, and explain what is incorrect

---

15 Sep 2017 Katie Evans

0012105, 0014602 and 0014762 still showing. 0014497 present and correct

---

Issues:102931

---

VIAMED Stock Meeting - Stock Processing

**WYA Consignment Stock (348)**

Issue ID #68832

Date Created 18/Mar/2016

Issued To: Catherine Spence

Completed Status: 21 Apr 2016

---

Notes:

System Generated

Please check all is well with the Yorkshire ambulance consignment Stock.

---

18 Mar 2016 Katie Evans

Next Action Changed From Katie Evans To Derek Lamb 0014602 out of sync again

---

18 Mar 2016 Derek Lamb

Next Action Changed From Derek Lamb To Catherine Spence its not out of sync as such, but looks like we need to send them 3

---

07 Apr 2016 Catherine Spence

can u check if this has corrected its self

---

14 Apr 2016 Catherine Spence

---

Issues:68832

---

VIAMED Stock Meeting - Stock Processing

**WYA Consignment Stock (349)**

Issue ID #87450

Date Created 20/Jan/2017

Issued To: Steve Hardaker

Completed Status: 02 Feb 2017

---

Notes:

System Generated

Please check all is well with the Yorkshire ambulance consignment Stock.

---

01 Feb 2017 Steve Hardaker

Correctly showing the updated stock levels following discontinuation of a number of Masimo products:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

---

Issues:87450

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85438

Date Created 14/Dec/2016

Issued To: Steve Nixon

Completed Status: 14 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85438

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85437

Date Created 14/Dec/2016

Issued To: Sarah Walton

Completed Status: 30 Jan 2017

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85437

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85436

Date Created 14/Dec/2016

Issued To: Ryan Swaine

Completed Status: 14 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85436

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85435

Date Created 14/Dec/2016

Issued To: Robert Connor

Completed Status: 15 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85435

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85434

Date Created 14/Dec/2016

Issued To: Philip Crossley

Completed Status: 15 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85434

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85433

Date Created 14/Dec/2016

Issued To: Michael Green

Completed Status: 15 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85433

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85432

Date Created 14/Dec/2016

Issued To: Jonathan Connor

Completed Status: 15 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85432

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85431

Date Created 14/Dec/2016

Issued To: Helen Lamb

Completed Status: 14 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85431

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85430

Date Created 14/Dec/2016

Issued To: Emma Clark

Completed Status: 14 Feb 2017

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85430

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85429

Date Created 14/Dec/2016

Issued To: Emily Hanson

Completed Status: 24 Jan 2017

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85429

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85428

Date Created 14/Dec/2016

Issued To: Derek Lamb

Completed Status: 15 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85428

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85427

Date Created 14/Dec/2016

Issued To: Catrin Hird

Completed Status: 14 Dec 2016

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85427

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85426

Date Created 14/Dec/2016

Issued To: Catherine Spence

Completed Status: 15 Feb 2017

---

**Notes:**

Copy Issue Main Issue #85425

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85426

---

VIAMED Sales And Marketing - Sales Viamed Medical UK  
**Yorkshire Ambulance - Major changes to consignment stock**

Issue ID #85425

Date Created 14/Dec/2016

Issued To: Katie Evans

Completed Status: 13 Feb 2017

---

**Notes:**

Added by Steve Hardaker sent to Katie Evans

INFORMATION ONLY ISSUE DO NOT ADD NOTES!

As you will all be aware, we can no longer obtain OEM Masimo products, which has a massive impact on the Yorkshire Ambulance account and the way that we process their consignment stock orders.

00144954 Masimo RED DCI-dc12 - We have now run out of stock on our own shelves. YAS have 60 on their shelves and, as of today, any orders for this item are to be invoice-only: \*please do not raise a consignment stock order\*.

0014520 Masimo PC04 - I retrieved the 20 units from their shelves and returned them to our stock. Their consignment stock is now 0 and they will not be ordering any more as the defibs that these were used on are no longer in service.

0014500 Masimo LNOP DCI - Again no longer in service. I retrieved 18 of the 20 that were on consignment; there are 2 that are unaccounted for. If they find them, they will return them; if they cant, they will order them. If they do order them: \*please do not raise a consignment stock order\*.

We will maintain the consignment stock at the following levels these items:

0012105 VM-2105 - 3

0014497 Masimo RED LNC-10 - 20

0014602 Nellcor SpO2 finger sensor - 5

0014702 Nellcor Silicone finger sensor - 2

Please continue to add orders to top up consignment stock for the above 4 lines until further notice.

Repairs should still continue to come in for now but this could change, depending upon whether we can determine the age of sensors supplied to YAS by Masimo.

Obviously, this is a bitter pill to swallow: YAS made up around 13% of UK Turnover in 2016, with 0014495 alone making up around 9% of UK Turnover.

As such, we will have to make a considerable increase in sales in 2017 to recover this, but as 2016 is about 14% up on 2015, if we continue this increasing sales trend we should be able to achieve this.

---

Issues:85425

---

VIAMED Stock Meeting - Missing Stock or Adjustments  
**Yorkshire Ambulance consignment stock update**

Issue ID #102025

Date Created 01/Sep/2017

Issued To: Derek Lamb

Completed Status: 03 Oct 2017

---

**Notes:**

Copy Issue Main Issue #102024

Added by Steve Hardaker sent to Jonathan Connor

Please can the items returned on SRS66164 be removed from YAS consignment stock pages to accurately reflect their stock holding.

3x 0012105 - Returned

5x 0014602 - Returned

2x 0014762 - Ordered on ORD88104

They only have 0110047 on consignment now.

---

01 Oct 2017 Derek Lamb  
done

---

03 Oct 2017 Steve Hardaker

---

Issues:102025

---

VIAMED Stock Meeting - Missing Stock or Adjustments  
**Yorkshire Ambulance consignment stock update**

Issue ID #102024

Date Created 01/Sep/2017

Issued To: Jonathan Connor

Completed Status: 08 Sep 2017

---

**Notes:**

Added by Steve Hardaker sent to Jonathan Connor

Please can the items returned on SRS66164 be removed from YAS consignment stock pages to accurately reflect their stock holding.

3x 0012105 - Returned

5x 0014602 - Returned

2x 0014762 - Ordered on ORD88104

They only have 0110047 on consignment now.

---

06 Sep 2017 Jonathan Connor  
returned

---

08 Sep 2017 Steve Hardaker

---

Issues:102024

# Clinical / FDA Incidents online search

## Clinical Investigation online review

Do any of the Results indicate a Risk / Problem : No

Do any of the Results indicate outdated Technology : No

Comments on Clinical Search :

Done in the 2160 review earlier today

## Review of online FDA Incident reports

Do any of the Results indicate a Risk / Problem : No

Do any of the Results indicate outdated Technology : No

Comments on Clinical Search :

Done in the 2160 review earlier today