

QA
0310034 Tom Thumb Infant Resuscitator (Oxygen).
Barcode ID: 365494
Serial Number 440121

Carried out by: Philip Crossley
06/Nov/2017 11:46 am

Performance evaluation: Pass

Signed: _____

Name: _____

Position: _____

Date: _____

QA 0310034 Serial Number 440121 Barcode ID: 365494
Carried out by: Philip Crossley 06/Nov/2017 11:46 am

[1]3038 Serial Number
440121

[2]818 Requested Value Limit
42

[3]819 No Leaks Pass
Yes

[4]820 Adjustable Valve @ 10 lpm Minimum < 8 cmH2O Pass
Yes

[5]2836 Adjustable Value @ 10lpm 42.1 +/- 1 mbar
42.0

[6]822 Precision Value @ 10lpm Maximum 44.1 mbar +/- 1
44.2

[7]823 Pressure Gauge @ 10 lpm Pressure 30cmH2O +/- 1 (28.4 - 30.4 mbar)
30.1

[8]824 Visual Check Flowmeter reamins steady after slight shake
Yes

[9]825 Visual Check Gauge needs tools to be removed
Yes

[10]826 Visual Check appears straight
Yes

[11]827 All settings set to a minimum
Yes

[12]828 All Labels in Place
Yes

[13]829 Built By
Philip Crossley

[14]872 Calibrated Manometer ID
674700

[15]2013 Customers Asset Number (service only)