

Probability of an Incident

Product: Tom Thumb;

Probability is based on the estimated use of the Tom Thumb in 4 Hospital Labour Wards/Delivery Suites and a description of there use in Neonatal Units

EN14971 D.2.3 Risks whose probability cannot be estimated.

The probability of risk depends on the number of units in use at any one time.

In practice this device is used in about 10% of births see docID 23809

Supporting documentation

DOC ID 23809

J.S.Lamb

11/09/17