

Certificate of Calibration

Certificate number: 1511862891

Tom Thumb Infant Resuscitator.
Part Number: 0310032
Serial number: 240070

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with Tom Thumb service manual.

Equipment used: Digitron Digital Manometer P200 (CE176)
Digitron Digital Manometer P200 (CE091)
Flow Meter (CE130)
Snoop liquid

Test Equipment	Test	Specification	Reading	Pass/Fail
Snoop liquid	Check all ports / connections for leaks	No bubbling	N/A	PASS
Digital Manometer (CE***)	Adjustable Valve: @ 10 lpm	Minimum: <8 cmH ₂ O	3.1	PASS
Digital Manometer (CE***)	Adjustable Valve: @ 10 lpm	Maximum: 42.1 ±1 mbar	41.8	PASS
Digital Manometer (CE***)	Precision Valve: @ 10 lpm	Maximum: 44.1 ±1 mbar	44.0	PASS
Digital Manometer (CE***)	Pressure Gauge Test @30cmH ₂ O	Pressure: 30cmH ₂ O ±1 (28.4 – 30.4mbar)	29.3	PASS

Name: Phil Crossley

Date: 28th November 2017

Signed: _____