





ALL RIGHTS STRICTLY RESERVED. REPRODUCTION OR ISSUE TO
THIRD PARTIES IN ANY FORM WHATEVER IS NOT PERMITTED
WITHOUT THE WRITTEN PERMISSION OF VIAMED Ltd.

Tolerance: = 1 Decimal Place ± 0.5
2 Decimal Place ± 0.1

REMOVE ALL SHARP EDGES



Surface Finish: 0.8 microns

Process Finish: Flash Nickel to a Maximum

Thickness of 0.005 microns

					Title				VIAMED Ltd. 15 Station Rd Cross Hills, Keighley West Yorkshire BD20 7DT	
					Valve Body		Dim in mm			
							Tol ± 0.2			
					Material : BRASS		Scale		Not To Scale	
					BS2874 CZ 121		Drawn			
REV	Date	N°	Drawn	Approv	Part No. 330229		Date	21/01/04	Dwg No.	TT0100.10