

DVM Calibration record

1	DVM Calibration Index Number	CE169
2	DVM Bar code	102153
3	Equipment Used	MICROCAL 1030
4	Serial number Test Equipment	6839 K6
5	Calibration Certificate Number	1512404873
6	Test Equipment Calibration Index Number	CE076
Micro Cal @		DVM Reading
7	0mV +/- 0.1	0.071 mV
8	5mV +/- 0.1	5.062 mV
9	10mV +/- 0.1	10.059 mV
10	15mV +/- 0.1	15.095 mV
11	20mV +/- 0.1	20.122 mV
12	25mV +/- 0.1	25.155 mV
13	30mV +/- 0.1	30.111 mV
14	35mV +/- 0.1	35.161 mV
15	40mv +/- 0.1	40.130 mV
16	Pass / Fail	PASS
17	Tested By	M. GREEN
18	Tested On	4-12-17
19	Signed	M. Green

1. Enter the details of the unit being tested in boxes 1 – 2

2. Enter the details of the calibrated Micro Cal unit being used to test against the equipment. In to boxes 3 -6

Ensure the calibrated unit is still inside its calibration period before you begin.

3. set the micro cal to the required settings (7 – 15) and enter the readings on the DVM in the Boxes

4. Assuming the units passes the tests, date and sign the form (17 – 19). Scan the form and upload to the issue requesting calibration and complete the Issue.

If the unit fails the test return the Unit to RS for calibration