l	r j	Design Changes
	Project	931111 Tom Thumb
	Date	Description of Change
	Dec 1994	Outlets changed to 15mm
i i	July 1995	Added the subtracted restrictor in O2 tubing inlet (used to prevent EBME; destroying during acceptance tests)
	June 1996	Suppliers changed
	Oct 1996	Altered safety valve into a precision valve by removing the ball
	Nov 1996	Grease changed (supplier reasons)
	December 96	Discussions with Drager on modification
		extended block by 32mm
		Increas the bore to accomodate two flow
		meters
(4	
•	<i>-</i>	
ĺ		
ļ		
· '	· .	
	/	