

Foreseeable Misuse of Product

PRODUCT : Foetal Simulator	
User Misuse	
Mechanical misuse	Dropping: could damage the enclosure
Electrical misuse	Not applicable
Other equipment	Not applicable ;
User misuse	The battery could be discharged.
Other	

Table 1: