

Foreseeable Misuse of Product

PRODUCT : Microstim	
User Misuse	
Mechanical misuse	Dropping: could damage the or the enclosure or the knob
Electrical misuse	Not applicable
Other equipment	Not applicable ; except pacemakers
User misuse	The user could disconnect the Microstim ; audio would not soundTom Thumb and by-pass the circuit
Other	

Table 1: