

# Therapy Equipment Ltd

Suppliers of  
Health Care  
Equipment

UNIT 1, CRANBOURNE AVENUE, CRANBOURNE INDUSTRIAL ESTATE  
POTTERS BAR, HERTS, EN6 3JN TEL: 0707 652270 FAX: 0707 652622



R 0211/MB

**FAX**

**FROM:** Chris Munn

**No. of Pages**

**TO:** John Lamb

**Date** 27th May 1994

John

Your Fax 23rd May 1994

The 0 - 100PSI Gauge is as purchased from the manufacturers. We can obtain a separate certificate of conformity if you wish, but I would assume that the manufacturers warranty of accuracy would suffice.

The Part Number for the Plastic Gauge Faces is 9901-151.

Your enquiry regarding the length of the main block. The size has been increased from 3 1/2" to 3 3/4", and the larger size has been supplied with the latest delivery.

Regards