

Certificate of Calibration

Certificate number: 1522250122

Tom Thumb Infant Resuscitator.
Part Number: 0310030
Serial number: VMD121

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with Tom Thumb service manual.

Equipment used: Digitron Digital Manometer P200 (CE176)
Digitron Digital Manometer P200 (CE091)
Flow Meter (CE130)
Snoop liquid

Test Equipment	Test	Specification	Reading	Pass/Fail
Snoop liquid	Check all ports / connections for leaks	No bubbling	N/A	PASS
Digital Manometer (CE176)	Adjustable Valve: @ 10 lpm	Minimum: <8 cmH ₂ O	1.0	PASS
Digital Manometer (CE176)	Adjustable Valve: @ 10 lpm	Maximum: 42.1 ±1 mbar	41.1	PASS
Digital Manometer (CE176)	Precision Valve: @ 10 lpm	Maximum: 44.1 ±1 mbar	43.7	PASS
Digital Manometer (CE176)	Pressure Gauge Test @30cmH ₂ O	Pressure: 30cmH ₂ O ±1 (28.4 – 30.4mbar)	28.5	PASS

Name: Phil Crossley

Date: 28th March 2018

Signed: _____