

Certificate of Calibration

Certificate number: 1523460695

Tom Thumb Infant Resuscitator.
Part Number: 0310030
Serial number: VMDH06

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with Tom Thumb service manual.

Equipment used: Digitron Digital Manometer P200 (CE176)
Digitron Digital Manometer P200 (CE091)
Flow Meter (CE130)
Snoop liquid

Test Equipment	Test	Specification	Reading	Pass/Fail
Snoop liquid	Check all ports / connections for leaks	No bubbling	N/A	PASS
Digital Manometer (CE176)	Adjustable Valve: @ 10 lpm	Minimum: <8 cmH ₂ O	4.9	PASS
Digital Manometer (CE176)	Adjustable Valve: @ 10 lpm	Maximum: 42.1 ±1 mbar	42.7	PASS
Digital Manometer (CE176)	Precision Valve: @ 10 lpm	Maximum: 44.1 ±1 mbar	44.3	PASS
Digital Manometer (CE176)	Pressure Gauge Test @30cmH ₂ O	Pressure: 30cmH ₂ O ±1 (28.4 – 30.4mbar)	29.0	PASS

Name: Phil Crossley

Date: 11th April 2018

Signed: _____