



CERTIFICATE OF CALIBRATION AND CONFORMANCE

This document certifies that the products identified below have been manufactured and tested in accordance with company standards and in compliance with the requirements of the purchase order/contract and reference documentation.

Product: IRMA AX+

Serial no: 912016

Date of Calibration: 23-Feb-2017

Final test

Gas Concentration	Measured Values (%)	Low Limit (%)	High Limit (%)	Status Pass/Fail
10 % CO ₂ BAL. N ₂	10,06	9,82	10,18	Pass
65 % N ₂ O BAL. N ₂	64,80	63,00	67,00	Pass
3 % HALOTHANE BAL. N ₂	3,06	2,76	3,24	Pass
3 % ENFLURANE BAL. N ₂	2,99	2,76	3,24	Pass
3 % ISOFLURANE BAL. N ₂	3,01	2,76	3,24	Pass
3 % SEVOFLURANE BAL. N ₂	3,03	2,76	3,24	Pass
5 % DESFLURANE BAL. N ₂	4,99	4,70	5,30	Pass

Traceability

Gas Concentration	Produced By	Accuracy
10 % CO ₂ BAL. N ₂	AGA Gas AB	0,03 % abs
65 % N ₂ O	AGA Gas AB	0,03 % abs
3 % HALOTHANE BAL. N ₂	Scott Medical	0,03 % abs
3 % ENFLURANE BAL. N ₂	Scott Medical	0,03 % abs
3 % ISOFLURANE BAL. N ₂	Scott Medical	0,03 % abs
3 % SEVOFLURANE BAL. N ₂	Scott Medical	0,03 % abs
5 % DESFLURANE BAL. N ₂	Scott Medical	0,03 % abs

Date of issue: 19-06-2018

Issued by:

Artur Kierbalaj

OEM Product Manager



CERTIFICATE OF CALIBRATION AND CONFORMANCE

This document certifies that the products identified below have been manufactured and tested in accordance with company standards and in compliance with the requirements of the purchase order/contract and reference documentation.

Product: IRMA AX+

Serial no: 912018

Date of Calibration: 23-Feb-2017

Final test

Gas Concentration	Measured Values (%)	Low Limit (%)	High Limit (%)	Status Pass/Fail
10 % CO ₂ BAL. N ₂	10,07	9,82	10,18	Pass
65 % N ₂ O BAL. N ₂	65,16	63,00	67,00	Pass
3 % HALOTHANE BAL. N ₂	3,07	2,76	3,24	Pass
3 % ENFLURANE BAL. N ₂	2,98	2,76	3,24	Pass
3 % ISOFLURANE BAL. N ₂	3,02	2,76	3,24	Pass
3 % SEVOFLURANE BAL. N ₂	3,04	2,76	3,24	Pass
5 % DESFLURANE BAL. N ₂	4,99	4,70	5,30	Pass

Traceability

Gas Concentration	Produced By	Accuracy
10 % CO ₂ BAL. N ₂	AGA Gas AB	0,03 % abs
65 % N ₂ O	AGA Gas AB	0,03 % abs
3 % HALOTHANE BAL. N ₂	Scott Medical	0,03 % abs
3 % ENFLURANE BAL. N ₂	Scott Medical	0,03 % abs
3 % ISOFLURANE BAL. N ₂	Scott Medical	0,03 % abs
3 % SEVOFLURANE BAL. N ₂	Scott Medical	0,03 % abs
5 % DESFLURANE BAL. N ₂	Scott Medical	0,03 % abs

Date of issue: 19-06-2018

Issued by:

Artur Kierbalaj
OEM Product Manager



CERTIFICATE OF CALIBRATION AND CONFORMANCE

This document certifies that the products identified below have been manufactured and tested in accordance with company standards and in compliance with the requirements of the purchase order/contract and reference documentation.

Product: IRMA AX+

Serial no: 912015

Date of Calibration: 23-Feb-2017

Final test

Gas Concentration	Measured Values (%)	Low Limit (%)	High Limit (%)	Status Pass/Fail
10 % CO ₂ BAL. N ₂	10,01	9,82	10,18	Pass
65 % N ₂ O BAL. N ₂	64,75	63,00	67,00	Pass
3 % HALOTHANE BAL. N ₂	3,05	2,76	3,24	Pass
3 % ENFLURANE BAL. N ₂	2,99	2,76	3,24	Pass
3 % ISOFLURANE BAL. N ₂	3,01	2,76	3,24	Pass
3 % SEVOFLURANE BAL. N ₂	3,02	2,76	3,24	Pass
5 % DESFLURANE BAL. N ₂	5,03	4,70	5,30	Pass

Traceability

Gas Concentration	Produced By	Accuracy
10 % CO ₂ BAL. N ₂	AGA Gas AB	0,03 % abs
65 % N ₂ O	AGA Gas AB	0,03 % abs
3 % HALOTHANE BAL. N ₂	Scott Medical	0,03 % abs
3 % ENFLURANE BAL. N ₂	Scott Medical	0,03 % abs
3 % ISOFLURANE BAL. N ₂	Scott Medical	0,03 % abs
3 % SEVOFLURANE BAL. N ₂	Scott Medical	0,03 % abs
5 % DESFLURANE BAL. N ₂	Scott Medical	0,03 % abs

Date of issue: 19-06-2018

Issued by:

Artur Kierbalaj
OEM Product Manager