

Monometer Calibration record Date 23-7-18

Mono meter Calibration Index Number	CE 078
Mono meter Bar code	102062
Equipment Used	Manometer
Serial number Test Equipment	102075
Calibration Certificate Number	1532426583
Test Equipment Calibration Index Number	CE091
Micro Cal @	DVM Reading

CE091 Mbar	Results Mbar	Error
10.0	09.8	0.2
20.0	19.9	0.1
30.0	29.9	0.1
40.0	39.9	0.1
50.0	49.9	0.1
60.0	59.8	0.2
70.0	69.6	0.4
80.2	79.0	1.2
90.0	88.2	1.8
Pass / Fail	PASS	
Tested By	D. CROSSEY	
Tested On	23-7-18	
Signed	M. Cull	