

Monometer Calibration record Date 23.7.18

Mono meter Calibration Index Number	CE149
Mono meter Bar code	102133
Equipment Used	Manometer
Serial number Test Equipment	102075
Calibration Certificate Number	1532426703
Test Equipment Calibration Index Number	CE091
Micro Cal @	DVM Reading

CE091 Mbar	Results Mbar	Error
10.0	9.9	0.1
20.0	19.9	0.1
30.0	29.9	0.1
40.0	39.7	0.3
50.0	49.6	0.4
60.0	59.3	0.7
70.0	69.2	0.8
80.2	78.7	1.3
90.0	88.1	1.9
Pass / Fail	PASS	
Tested By	P. CROSSLEY	
Tested On	23.7.18	
Signed	Mr. [Signature]	