

EC Declaration of Conformity

We hereby declare under our sole responsibility
that the generic product group

Disposable SMARTsat SpO₂ Sensors (VM)

**Sensors for continuous and spot check measurement of functional arterial
oxygen saturation (SPO₂) and pulse rate**

REF No.	Type	Product
6020131204	0015010	10-AP-VM, Adult Plaster Disposable SpO ₂ Sensor
6020131207	0015011	10-PP-VM, Paediatric Plaster Disposable SpO ₂ Sensor
6020131209	0015012	10-IP-VM, Infant Plaster Disposable SpO ₂ Sensor
6020131211	0015013	10-NP-VM, Neonatal Plaster Disposable SpO ₂ Sensor

Conforms with the essential requirements of Annex I and Annex II of the Council Directive 93/42/EEC as amended by 2007/47/EC concerning medical devices.

In accordance with Annex IX of the Council Directive 93/42/EEC the product has been classified as Class IIb.

Application of the CE-marking:



MEDCERT GmbH, Pilatuspool 2, 20355 Hamburg

Validity:

Date of expiry: 31 December 2019

Issuer:

bluepoint medical GmbH & Co. KG
An der Trave 15
23923 Selmsdorf
Germany

Place, Date:

Selmsdorf, 01 February 2018

Legally binding signature:

Bernd Lindner
General Manager