

Certificate of Compliance

Certificate: 1390739

Master Contract: 163296 (LR 68348)


Edition: 1


Date Issued: December 13, 2002

Issued to: Teledyne Instruments
Analytical Instruments
16830 Chestnut Street
City Of Industry, CA 91749-1020
USA

The products listed below are eligible to bear the CSA Mark shown




Issued by: Qrest Ewanchyna, P. Eng.


Approved by: Terry Nagy
Operations Manager

PRODUCTS

CLASS 8750-01 - MEDICAL ELECTRICAL EQUIPMENT

Oxygen Monitors, Models MX300 and AX300, battery operated, no patient applied parts.

APPLICABLE REQUIREMENTS

CAN/CSA C22.2 No 601.1-M90 - Safety of Medical Electrical Equipment, Part I, General
CSA 601.1 Supplement 1:1994 Requirements for Safety
CSA 601.1 Amendment 2:1998



Supplement to Certificate of Compliance

Certificate: 1390739

Master Contract: 163296 (LR 68348)

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Edition	Date	Description
1	December 13, 2002	Original Certification.