

CERTIFICATE OF CALIBRATION



Issued by: RS Components Ltd

Date Issued: 17 Jan 2008

Certificate No.

1082471

RS Calibration

Calibration and Repair Service

Page 1 of 2 Pages

DPN 175, Lammas Rd,
Weldon Industrial Est
Corby, Northants, NN17 9RS

Tel: 01536 405545

Fax: 01536 401590

Simon Perkins

Client	VIAMED KEIGHLEY WEST YORKSHIRE BD20 7DT
Instrument	Digitron P200L i.s. 0 to 500 mbar Digital Manometer
Serial No.	101989
Client Reference	CE091
Procedure ID.	188.1175 Rev. P2
Date of Calibration	17 Jan 2008
Performance Status	Pass

Equipment Used to Carry Out Calibration

Equipment ID.

Ruska Deadweight Tester
Weights

Cal 711
Cal 713, 714

The measurements reported in this certificate were carried out using equipment whose values are traceable to national standards.

All procedures employed and results reported are in compliance with the requirements of the International Standard ISO/IEC 17025:2005.

The management controls of the RS Calibration Laboratory are registered under the British Standard BS EN ISO 9001 : 2000 No. RS 00362.

Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

This certificate reports recorded values after adjustment to the instrument was carried out.

CERTIFICATE OF CALIBRATION

Certificate No.

1082471



Calibration and Repair Service

Page 2 of 2 Pages

The ambient temperature and relative humidity throughout the calibration were $(20 \pm 2) ^\circ\text{C}$ and $(50 \pm 10) \% \text{RH}$ respectively.

The instrument was calibrated using a dead weight pressure standard, applying a set mass to achieve nominal values of gauge pressure and recording the instrument's displayed values.

Prior to calibration the instrument was exercised to full scale pressure and then returned to zero.

The pressure medium used was nitrogen.

Range	Applied Value	Instrument Reading	Deviation as % of F.S.	L.S.D. Stability
500 mbar	0 mbar	0.0 mbar	0.00	0
	100 mbar	100.0 mbar	0.00	0
	190 mbar	190.2 mbar	0.04	0
Hi	300 mbar	300 mbar	0.00	0
	400 mbar	400 mbar	0.00	0
	500 mbar	500 mbar	0.00	0

Measurement uncertainty of the above recorded values: $\pm (0.01\% + 1 \text{ L.S.D.})$

END OF CALIBRATION

CALIBRATED BY:- SPP

Compliance to Specification

The specification published by the manufacturer and found in the instrument's handbook has been used to determine performance at the measured points.

Reported values

The uncertainties quoted refer to the applied values, which include any identified contribution of the instrument under test and not to the ability of the instrument to maintain its calibration.