

Viamed R-75V Oxygen Sensor Packaging & Sensor Labelling

0110075 R-75V

Medical Oxygen Sensor R- 75V


REF 0110075
SN V123456
2019-03-01



(01)04036616000307(11)190301(21)V123456


VIAMED


CE 0123
 EnviteC-Wismar GmbH
 Alter Holzhafen 18
 D-23966 Wismar
 www.envitec.com

Tel: +44 (0)1535 634542
 Email: info@viamed.co.uk Website: www.viamed.co.uk

Depending on the ambient conditions the sensor may require up to 30 minutes to reach signal stability after installation. The sensor contains caustic liquid. In case of leakage avoid contact with eyes and skin! Recommended storage temperature: 5°C - 15°C!

Packaging label: 63.5 x 46.6 mm


VIAMED

Medical Oxygen Sensor R- 75V

REF 0110075
SN: V123456




CE 0123
 EnviteC-Wismar GmbH



2019-03

Tel: +44 (0)1535 634542
 Email: info@viamed.co.uk
 Website: www.viamed.co.uk

Sensor label: 90 x 15 mm