

WIKA INSTRUMENTS LIMITED



STATION APPROACH, COULSDON, SURREY CR5 2UD
Telephone: 0181 - 668 0924 Fax: 0181 - 660 5956

Customers Ref.: DL/EG/8889
WIKA Ref.: 203633
Customer: VIAMED LTD

Certificate of Calibration

WIKA Model No.: 611.10.63

Size and Style: 63mm back

Calibrated Range: -5/+60 cm wg

Serial Number: 203633

Tag Number: N/A

Accuracy of Determination: $\pm 1.6\%$ FSD

Test Equipment: RUSKA AIR BENCH

DKD Number: 4733

Accuracy of Test Equipment: $\pm 0.009\%$ FS 800mbar

Testing Temp.: 20 ° C

Method of Test: Direct Comparison

Test Medium: Air

True Units cm wg	Pressure	Gauge Readings			
		Rising Pressure		Falling Pressure	
		Reading	Error [%]	Reading	Error [%]
-5,000		-5,650	-1,00	-5,740	-1,14
0,000		0,000	0,00	0,000	0,00
10,000		10,120	0,18	10,060	0,09
30,000		30,320	0,49	30,070	0,11
60,000		60,850	1,31	60,810	1,25

All equipment utilised in the preparation of this Certificate is directly traceable to PTB primary standards under licence DKD-K-3701

Tested by:

D.P. Winkler

Signed

P. Weller

Date:

07.06.1997

Quality Assurance

WIKA INSTRUMENTS LIMITED



STATION APPROACH, COULSDON, SURREY CR5 2UD
Telephone: 0181 - 668 0924 Fax: 0181 - 660 5956

Customers Ref.: DL/EG/8889
WIKA Ref.: 203633
Customer: VIAMED LTD

Certificate of Calibration

WIKA Model No.: 611.10.63
Size and Style: 63mm back
Calibrated Range: -5/+60 cm wg
Serial Number: 203633
Tag Number: N/A
Accuracy of Determination: $\pm 1.6\%$ FSD
Test Equipment: RUSKA AIR BENCH DKD Number: 4733
Accuracy of Test Equipment: $\pm 0.009\%$ FS 800mbar Testing Temp.: 20 °C
Method of Test: Direct Comparison Test Medium: Air

True Units cm wg	Pressure	Gauge Readings			
		Rising Pressure		Falling Pressure	
		Reading	Error [%]	Reading	Error [%]
-5,000		-5,650	-1,00	-5,740	-1,14
0,000		0,000	0,00	0,000	0,00
10,000		10,120	0,18	10,060	0,09
30,000		30,320	0,49	30,070	0,11
60,000		60,850	1,31	60,810	1,25

All equipment utilised in the preparation of this Certificate is directly traceable to PTB primary standards under licence DKD-K-3701

Tested by :

D.P. Wicks

Signed

P. Weller

Date:

07.06.1997

Quality Assurance

ce DL

THERAPY EQUIPMENT LTD
Unit 1 Cranbourne Avenue
Cranbourne Industrial Estate
Potters Bar, Hertfordshire EN6 3JN
ENGLAND

Fax Cover Sheet

DATE: June 4, 1997 **TIME:** 13:26
TO: Derek Lamb **PHONE:** 01535 634542
Viamed Ltd **FAX:** 01535 635582
FROM: STEVE MUNN **PHONE:** 01707 652270
Therapy Equipment **FAX:** 01707 652622
RE: Copper Sealing Washer

Number of pages including cover sheet: 1**Message**

Dear Derek,

Thank you for your fax addressed to Chris, who is currently away from the office on business.

We continue to use the gauge seal in question on another of our products and we would therefore be willing to sell these to you for your use on the Tom Thumb units.


The details are as follows:

Part No. 6650-05 - 1/8" Copper Washer - Price £0.50.

I trust the above meets with your approval.

Kind regards.

Yours sincerely,



Steve Munn

09/06/97

09:48

VIAMED → 01707652622

NO. 009 001



VIAMED



FAX

Fax No. DL / 154 / 8107
 Company Therapy Equipment
 To Chris Munn
 Co Fax No. 01707 652622
 Co Tel No. 01707 652270

From Derek Lamb
 VIAMED Ltd
 Date 09/06/97
 Subject: O Rings.
 No. Pages

Dear Chris,

I am going through Our file on the Tom Thumb Unit so as to get the CE Mark.
 The tom Thumb uses 3 Different 'O' Rings, Although I have 4 P/Ns which were / are
 4701 - 13 O ring seal 1/8 fixing NO INFORMATION AVAILABLE (OBSOLETE)
 4450 - 13 Main valve O Ring OS BS 1 X 1/6 N.70
 4701 - 18 Adj valve O ring. OSA BS 110 N.70
 4701 - 64 Flowmeter Bolt O ring BS012 WHITE 31.

Do you have a specification of these O Rings ?

Thankyou.

Leana

Chris Munn

RS.

4450 - 13

128-827

4701 - 18

689-647

4701 - 64

689-631

Yours,

O'ring

£1.90

O'ring

£1.89

O'ring

£1.89.

these are correct!

Derek Lamb



Viamed Limited, 15 Station Road, Cross Hills,
 Kethley, West Yorkshire BD20 7DT
 Tel: +44 (0) 1535 634542/636757 Fax: +44 (0) 1535 635582
 Registration No. 1291765 in England

TT 101