



Ryan Swaine <viamed.ryan.swaine@gmail.com>

TEK-OX analyser emc

1 message

Jens Vilöhr <viloehr@cecert.de>

23 March 2018 at 14:52

To: "viamed.ryan.swaine@gmail.com" <viamed.ryan.swaine@gmail.com>

Cc: Anja Hartig <hartig@cecert.de>

Hello Mr Swaine,

the TEK-OX analyser can you tested in accordance to EN 61326-1 as laboratory equipment (emission class B, immunity industrial environment).

For the FCC (USA) I would tested it as a class B digital device (only emission).

Mit freundlichen Grüßen / Best Regards

Jens Vilöhr

CEcert GmbH
Alter Holzhafen 19a
D-23966 Wismar

Telefon: [+49\(0\)3841/30305-29](tel:+49038413030529)

Telefax: [+49\(0\)3841/30305-18](tel:+49038413030518)

Homepage: www.cecert.de

Mail: viloehr@cecert.de

Amtsgericht / local court: Schwerin

Handelsregister / commercial register: HRB 6695

Geschäftsführer / managing director: Bernd Schmidt

Internet-E-Mail-Vertraulichkeit Haftungsausschluss:

Diese E-Mail enthält vertrauliche und/oder rechtlich geschützte Informationen. Wenn Sie nicht der richtige Adressat sind oder diese E-Mail irrtümlich erhalten haben, informieren Sie bitte sofort den Absender und vernichten Sie diese Mail. Das unerlaubte Kopieren sowie die unbefugte Weitergabe dieser Mail ist nicht gestattet.

Internet E-mail Confidentiality Disclaimer:

This e-mail may contain confidential and/or privileged information. If you are not the intended recipient (or have received this e-mail in error) please notify the sender immediately and destroy this e-mail. Any unauthorized copying, disclosure or distribution of the material in this e-mail is strictly forbidden.

Bitte denken Sie an Ihre Verantwortung für unsere Umwelt, bevor Sie diese E-Mail ausdrucken.

Please consider your environmental responsibilities before printing this email.