

## **Work Instructions & Test Methods.**

The correct information should be supplied with the Customer Order as a reference for the manufacture of the Microstim. If the correct information has not been supplied, the Customer should be contacted, and the information should then be recorded and a copy filed with the original order.

- Using the Information, parts are collected, and booked out, from stock.
- The parts are assembled as per the required information,
- The modified unit should now be tested in line with the calibrated equipment (see the relevant operating procedure).
- The unit should now have a visual Quality Assurance / Control check performed, to
  ensure that all labels are in position and that the unit is to the same design as that
  specified by the Customer.
- The completed unit, with a clean copy of the test sheet should now be presented to the Q.A. department for assessment.

Complete Microstims are received HTI having been set up for correct pulse widths and spaces.

The units only require a simple go / no-go test as per the Microstim Test & Q.A. procedure VM3COP36.01

Testing & QA is carried out by suitably familiarised staff.