

Service Repair Sheet SRS67463

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Lucy Montague

Vale Vets Animal Hospital

Nursing Manager

01453 542092

00012761

lucym@valevets.co.uk

01/Apr/2020

Catherine Spence

Viamed

Quote

Notes 20/Mar/2020 Catrin Hird

20/Mar/2020 Catrin Hird

20/Mar/2020 Catrin Hird

"We have two of your Microstim DB3 supramaximal nerve stimulator, one of them is not working properly- it has power but we cannot get it to work on TOF."

01/Apr/2020 Catherine Spence

1 unit received S/N M0004954

Ready For quote

Repair Complete Signed

CGNEEN 21.4.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32452	Microstim DB3	2510000	M0004954	N

2540000.

x1 @ £0.

S/N, SRS, SRN

NO fault found.

UPS x 1 @ £0.

Sp
20
S/Z

valevets

The Animal Hospital
Stinchcombe
Dursley
Glos
GL11 6AJ

T: 01453 542092

Accounts: 01453 544668

E: reception@valevets.co.uk

www.valevets.co.uk

We cannot get the TOF to work on this unit. But it seems to have power

Kind Regards

Lucy Montague (Nursing Manager)



.....with compliments