

Service Repair Sheet SRS67477

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Anita Birkett

Company/ Hospital Name

Lincoln County Hospital

Department

Clinical Engineering

Position

Resp Tech

Direct Phone

01522 707434

General Phone

0152 251 2512

Opera Account

00002610

Email

Order Number

Date Received

27/May/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Warranty

Notes 27/Apr/2020 Sophie Lines

27/Apr/2020 Sophie Lines

27/Apr/2020 Sophie Lines

2x R-43v used in Giraffe Omnibed since March 2020 - already failed. Advised to send back and if faulty we will replace - have asked for serial numbers

27/Apr/2020 Sophie Lines

27/Apr/2020 Sophie Lines

27/Apr/2020 Sophie Lines

27/Apr/2020 Sophie Lines

2x R-43v used in Giraffe Omnibed since March 2020 - already failed. Advised to send back and if faulty we will replace - have asked for serial numbers

update:

The serial numbers are V105957 and V105995

27/May/2020 Catherine Spence

received in

Ready For quote

CC

Repair Complete Signed

CCNOON 4.6.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32463	Oxygen Sensor	0110043	V105957/V105995	Yes

0140013 x 1 FOC  
S/N old SRS, SRN

0110043 x 1 FOC  
New S/N, SRS, SRN

UPS x 1 FOC

UPS X 1 FOL

hce 3372



United Lincolnshire Hospitals

NHS Trust

Lincoln County Hospital  
Greetwell Road  
Lincoln  
LN2 5QY

Enclosed are two  
R-43V sensors which  
have failed in use, s/n's  
V105995 and V105957.  
Your reference number is  
SRS67477.

Yours sincerely  
*Anita Birkett*

Mrs Anita Birkett  
Specialist Clinical Engineering Technician  
Clinical Engineering Dept  
Direct Dial: +44 (0) 1522 707434  
Email: [Anita.Birkett@ulh.nhs.uk](mailto:Anita.Birkett@ulh.nhs.uk)

Return: 01 Jan 1970 -3600