

Service Repair Sheet SRS67504

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

A Graham

Sunderland Royal Hospital

Clinical Engineering

0191 5699933

00004992

02/Jun/2020

Catherine Spence

Viamed

Quote

Notes 28/May/2020 Catherine Spence

28/May/2020 Catherine Spence

28/May/2020 Catherine Spence

Tom Thumb with pole clamp

stuck at 40

02/Jun/2020 Catherine Spence

S/N VMC9072

Ready For quote

Repair Complete Signed

CC Noon 4-6-2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32488	Tom Thumb	0310030	VMC9072	N

0340000 @ £22.50.
S/N, SRS, SRN

0330203 x 1 @ £125
SRS, SRN

UPS x 1 @ £10.

MEDICAL DEVICE RETURNS FORM

CLINICAL ENGINEERING
SUNDERLAND ROYAL HOSPITAL
KAYLL ROAD
SUNDERLAND
SR4 7TP

Tel: 0191 5699933
Fax: 0191 5699274

Date:

20/05/20

The enclosed equipment is in need of repair.

Description: Tom Thumb

C.H.S. Asset No: 088326

C.H.S. Job No: 937276

Make: Vicamed

TNT Cons. No: 532250595

Model: Tom Thumb

Serial No: TT480

Order No: For Repair E/

On Contract: ☒

Estimate Requested. YES/NO

Warranty: ☒

Sent For Attention Of:

Vicamed Ltd.

15 Station Road

Cross Hills

Keighley

West Yorkshire

Post Code BD20 7DT

Fault Description/Action to Take:

Pressure stuck at 40+

Accessories Sent:

1x Tom Thumb

Sent By: A. GRAHAM

Return To:

Clinical Engineering, Sunderland Royal Hospital, A Floor Theatre Block, Kayll Road,
Sunderland, SR4 7TP

Note: It is the responsibility of the receiver to confirm the contents that are listed above are correct.
Any damaged or missing items must be reported to the Electronics Department immediately.
DECLARATION OF CONTAMINATION STATUS OVERLEAF