

Service Repair Sheet SRS67476

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Tony Parsons

Company/ Hospital Name

Princess Alexandra Hospital

Department

EBME

Position

Technician

Direct Phone

01279 82 75 33

General Phone

01279 827951

Opera Account

00002079

Email

anthony.parsons@pah.nhs.uk

Order Number

Date Received

27/May/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Warranty

Notes 27/Apr/2020 Zoey Teal

27/Apr/2020 Zoey Teal

Sending in a faulty R-17MED sensor purchased in Feb 2020.

Serial number: 389989.

Advised them that there will be a delay in repairs.

27/May/2020 Catherine Spence

received in

Ready For quote

Repair Complete Signed

CCNEEN 4-6-2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32464	Oxygen Sensor	0110017	389989	Yes

0140011 x 1 FOC

SIN, SRS, SRN

No fault found.

UPS x 1 FOC

Repairs, Planned preventative maintenance, Acceptance testing, Asset management, and Contract management.

The Princess Alexandra Hospital

NHS Trust



ISO 9001:2008

EBME DEPARTMENT

HAMSTEL ROAD, HARLOW, ESSEX, CM20 1QX

Phone 01279 827533

Fax 01279 827924

SLS 67476

28/04/20

Oxygen cell return - serial number 389989

Please find our returned oxygen cell - p/n R-17MED – 0110017

This cell was purchased recently and has just been opened for use – but found to be faulty.

Please could you consider sending us a replacement?

Regards,

Tony Parsons

EBME Dept.