

Service Repair Sheet SRS67513

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Misba Suhail

Company/ Hospital Name

Luton & Dunstable Hospital

Department

Position

Clinical Engineer

Direct Phone

01582 491166 In Ext:2480

General Phone

Opera Account

00003490

Email

Misba.Suhail@ldh.nhs.uk

Order Number

Date Received

16/Jun/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Quote

Notes 10/Jun/2020 Kate Griffiths

10/Jun/2020 Kate Griffiths

Customer has

16/Jun/2020 Kate Griffiths

10/Jun/2020 Kate Griffiths

10/Jun/2020 Kate Griffiths

Customer has ``VM-2500-S with broken battery terminal and a battery case to send in for repair``. Serial number is 60940147.

Email from customer - I would appreciate if this device could be dealt as a priority as it belongs to our A&E department.
16.6.2020

16/Jun/2020 Catherine Spence

no parts with it, unit only

Ready For quote

C Green 17-6-2020

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32791	Capnograph	4410520	60940147	N

4440000 x 1/2 hr @ £90 = £45.
S/N, SRS, S/N

UPS x 1 @ £10-00

The Luton and Dunstable University Hospital NHS Trust
Lewsey Road
Luton
Bedfordshire
LU4 0DZ

Telephone : 08451270127 ext. 2480
Facsimile : 01582570606

CLINICAL ENGINEERING FAULT REPORT

TO : VIAMED

From : Misba Hu
Suhail
Clinical Engineer

Misba.Suhail@ldh.nhs.uk

The following equipment has been taken out of service.

If you have any questions or concerns please contact Misba Hu Suhail on ext. 2480

Date

: 10/06/2020

Asset no

: 26233

WR no

: 97082

Make / Description of Equipment

: VIAMED / Capnograph and
Oximeter

Model

: VM-2500-S

Serial No

: 60940147

Fault : *Broken battery terminal
and casing. *(battery casing)*