

Service Repair Sheet SRS67481

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Magnus Persson

Company/ Hospital Name

Swevet - AB

Department

Position

Direct Phone

General Phone

46 771215500

Opera Account

00006308

Email

magnus.persson@swevet.se

Order Number

Date Received

28/May/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Quote

Notes 05/May/2020 Zoey Teal

05/May/2020 Zoey Teal

Sending in a VM2160. Serial number: T91100162.

Unit comes up with no sensor even when a new sensor is tried. I have advised that there will be a delay with repairs.

28/May/2020 Catherine Spence

Unit only received

Ready For quote

Repair Complete Signed

CC Neen 23.6.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32468	Pulse Oximeter	0012176	T91100162	N

0042160 x 1 FOC

S/N, SRS, SRN

NO fault found

UPS x 1 FOC