

Service Repair Sheet SRS67522

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Laura B

Company/ Hospital Name

J H M Butt

Department

Position

Direct Phone

General Phone

Opera Account

00020030

Email

Order Number

Date Received

18/Jun/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Warranty

Notes 18/Jun/2020 Catherine Spence

18/Jun/2020 Catherine Spence

18/Jun/2020 Catherine Spence

2 X R22a s/n 405701,373677

18/Jun/2020 Catherine Spence

405701: took a long time to pass , kept faulting in machine every couple of days, then every day

373677: would not pass.

Please see customer letter for full details

Ready For quote

Repair Complete Signed

C. Spence 23.6.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32794	Oxygen Sensor	0110122	405701	Yes
SRN32795	Oxygen Sensor	0110122	373677	Yes

0140013 x 2
old SIN, SRS, SRN

0110122 x 2
new SIN, SRS, SRN

UPS x 1

DATE: 12/6/20

To:

TO:
Vianed Ltd.
15 station Road
Cross Hills
Keighley.
BD20 7DT.

TO:
Vianed Ltd.
15 station Road
Cross Hills
Keighley.
BD20 7DT.

[illegible]

Method of return:

Supplier Collecting / Butts Transport or Own Carrier

Date Returned: 12/6/20

PLEASE ENSURE THAT A CREDIT NOTE IS RAISED TO COVER THE GOODS LISTED ABOVE IF IT IS NOT A WARRANTY ISSUE