Service Repair Sheet SRS67520

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Misba Suhail

Luton and Dunstable Hospital

Clinical Engineer

Tel: 01582 491166 In Ext:2480

00003490

Misba.Suhail@ldh.nhs.uk

25/Jun/2020

Catherine Spence

Viamed

Quote

Notes 16/Jun/2020 Kate Griffiths

16/Jun/2020 Kate Griffiths

Customer has a VM-2500-S, Serial number 80640163, that they need to send to your workshop for repair.

16/Jun/2020 Kate Griffiths

Customer has a VM-2500-S, Serial number 80640163, that they need to send to your workshop for repair. Update -

"The device does not start up and also has a broken display screen."

25/Jun/2020 Catherine Spence

received in unit only with mounting bracket attached

Ready For quote

Green 4-7.2020

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
	Capnograph			2

4440000 × 1/2 = £45 S/N, SEN, SRS SES SEN

4430120 × 1 0 E72 5RS, SEN

UPSX 1 DE10





The Luton and Dunstable University Hospital NHS Trust Lewsey Road Luton Bedfordshire LU4 0DZ

Telephone: 08451270127 ext. 2480

Facsimile: 01582570606

CLINICAL ENGINEERING FAULT REPORT

TO: VIAMED

From: Misba Hu

Suhail

Clinical Engineer

The following equipment has been taken out of service.

If you have any questions or concerns please contract Misba Hu Suhail on ext. 2480

Date

: 16/06/2020

Asset no

: 25308

WR no

: 97245

Make / Description of Equipment

: VIAMED / Capnograph and

Oximeter

Model

: VM-2500-S

Serial No

: 80640163

Fault: *The device has got a broken screen and device does not turn on.

3