

Service Repair Sheet SRS67537

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Maria Hewitt

Company/ Hospital Name

Eurotec Equipment Services

Department

Position

Ms Hewitt

Direct Phone

General Phone

01948 780546

Opera Account

00020096

Email

maria-hewitt@btconnect.com

Order Number

Date Received

17/Jul/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Warranty

Notes 10/Jul/2020 Sophie Lines

10/Jul/2020 Sophie Lines

10/Jul/2020 Sophie Lines

returning faulty sensor - s/n 376937

17/Jul/2020 Catherine Spence

received in

Ready For quote

Repair Complete Signed

CC Neen 23-7-2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32941	Oxygen Sensor	0110132	376937	Yes

0140013 x 1 @ £0.00
old S/N, SRS, SRN

0110132 x 1 @ £0.
New S/N, SRS, SRN

UPS x 1 @ £0

10th July 2020

Hi Sophie.

Please find enclosed faulty O2 sensor
showing 14 mv. S/no. 376937

Ref. SRS 67537.

Many thanks

Maria.

www.eurotec-ge.co.uk

