

Caution

Do not apply excessive force to the treatment platform.

Maximum loading: 25Kg

Ensure that the door is free of obstructions before closing.

Do not attempt to close the clasp if force is required to close the door.