

Service Repair Sheet SRS67542

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600  
Sumera Ahmed  
East Surrey Hospital

01737 768511 ext 6602  
01737 76 85 11  
00004300  
sumera.ahmed@nhs.net

03/Aug/2020  
Catherine Spence  
Viamed  
Quote

Notes 16/Jul/2020 Zoey Teal

16/Jul/2020 Zoey Teal

Sending in a V1000. It is not producing a heartbeat. They have changed the batteries and also tried multiple dopplers which work on a different unit.

Serial number: PR02094A26

03/Aug/2020 Catherine Spence  
received in

Ready For quote

CAneen 4-8-2020

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32961	Foetal Simulator	1410000	PR02094A26	N

1480000 x 1 £45

S/N, SRS, S/N

1430309 x 1 @ £0.

SRS, S/N

UPS x 1 £10