



EME Services Equipment Work Ticket

Surrey and Sussex
Healthcare NHS Trust 

Email Address: sash.eme.helpdesk@nhs.net

16 July 2020

Equipment: TEST EQUIPMENT

Asset Number: ES24539

Serial Number: PR02094A26

Reported Date: 16/07/2020

Model: V1000

Batteries last changed: Bat1 :

Class and Type: 3 B

Bat2 :

Manufacturer: VIAMED LTD

Next PPM: 04/05/2021

Equipment Location: EME SERVICES - CH

Service Book EBME - ESH DEPT - CH
Location:

EME Contact Name: SUMERA AHMED

Return Address: EME SERVICES - EAST SURREY HOSPITAL

Email Address: sash.eme.helpdesk@nhs.net

Fault: WORKING DOPPLER IS NOT PICKING UP HEART BEAT FROM SIMULATOR

Notes: JOB REFERENCE NUMBER: SRS67542

East Surrey Hospital, Canada Avenue, Redhill, Surrey, RH1 5RH. Tel: - 01737 768511 ext 6600.

Crawley Hospital, West Green Drive, Crawley, Sussex, RH11 7DH. Tel: - 01293 600300 ext 3171.