



Ing. Nufer AG
Schürmattstrasse 6
Postfach 125
CH-3073 Gümligen

Tel. ++41 31.958 66 66 Fax ++41 31 951 46 73
E-Mail: info@nufer-medical.ch
Internet: <http://www.nufer-medical.ch>
UBS Bern
MWSt-Nr. 121 519



FAX MESSAGE

Gümligen, 11 February 2000	Pages: 1 incl. cover
To: Viamed Limited, Keighley	
Fax No.: 0044 1535 635 582	
Attention: Mr S Watmouth, cc Mr Peter Lamb	

IMPORTANT**URGENT**IMPORTANT**URGENT**IMPORTANT

Electrical hazard with "Wandarm" Type 79

Dear Mr Watmouth,

Further to your fax of 28 October 1999 we would like to inform you that Ondal has contacted us by phone to confirm that you were correct and the hazard you mentioned exists (which is probably why it took them so long to get back to us).

The letter of confirmation which they promised has not been received, yet. We, therefore, do not know their next steps. However, it seems obvious that they have to implement some modifications or to replace this type of arm.

For the time being we hold a stock of about 50 arms and will have to continue selling these until we either receive a modification kit or an official re-call from Ondal.

As soon as we know their decision about how to go on, we will contact you again.

Best regards,

ING. NUFER AG

Dieter Indermühle
Product Manager

IMPORTANT**URGENT**IMPORTANT**URGENT**IMPORTANT