

## **Biocompatibility & Toxicity Evaluation**

The materials Perspex: Acrylic: has been used continuously in the Neonatal Intensive care areas since the mid 1960's without adverse effects being reported.