

Service Repair Sheet SRS67544

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Dieter Eckert

Company/ Hospital Name

DEM Medizintechnik

Department

Position

Direct Phone

General Phone

49 519397590

Opera Account

00007016

Email

info@dem-med.de

Order Number

Date Received

18/Jul/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Quote

Notes 18/Jul/2020 Catherine Spence

18/Jul/2020 Catherine Spence

18/Jul/2020 Catherine Spence

Microstim and leads

S/N M0004318

Divice has loose contact, works for a few seconds after inserting new batteries and then stops working

Ready For quote

CC, Neon 18-8-2020

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32946	Microstim DB3	2510000	M0004318	N

2540010 x 1 @ € ~~49.42~~
S/N, SRS, SRN

\$ 57.29

UPS x 1 @ £ \$
32x26x12cm